

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051822\
 Data File : BG053617.D
 Acq On : 19 May 2022 00:00
 Operator : CG/JU
 Sample : N2879-11
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-4-0516

Quant Time: May 19 05:13:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 11 06:00:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.082	152	21225	20.000	ng	0.00
21) Naphthalene-d8	10.889	136	88021	20.000	ng	0.00
39) Acenaphthene-d10	14.702	164	66390	20.000	ng	-0.01
64) Phenanthrene-d10	17.451	188	166360	20.000	ng	0.00
76) Chrysene-d12	21.734	240	183728	20.000	ng	0.00
86) Perylene-d12	25.006	264	190208	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.650	112	184827	147.786	ng	0.00
7) Phenol-d6	7.247	99	247936	130.628	ng	0.00
23) Nitrobenzene-d5	9.239	82	209932	101.527	ng	-0.01
42) 2,4,6-Tribromophenol	16.194	330	151442	154.663	ng	0.00
45) 2-Fluorobiphenyl	13.327	172	453105	99.731	ng	0.00
79) Terphenyl-d14	20.054	244	965101	97.833	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051822\
Data File : BG053617.D
Acq On : 19 May 2022 00:00
Operator : CG/JU
Sample : N2879-11
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-4-0516

Quant Time: May 19 05:13:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 11 06:00:46 2022
Response via : Initial Calibration

