




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051822\
Data File : BG053604.D
Acq On : 18 May 2022 14:48
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 19 05:11:47 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 11 06:00:46 2022
Response via : Initial Calibration



TIC: BG053604.D\data.ms

(70) Pentachlorophenol (C)




17.110min (-0.009) 23904.58 ng

response 100851

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	70.20	63.98
264.00	66.20	60.92
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051822\
Data File : BG053604.D
Acq On : 18 May 2022 14:48
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 19 05:11:47 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 11 06:00:46 2022
Response via : Initial Calibration



TIC: BG053604.D\data.ms

(77) Benzidine

19.677min (-0.003) 0.00 ng

response 510009

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.58
183.00	11.10	11.84
0.00	0.00	0.00