



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051822\
 Data File : BG053596.D
 Acq On : 18 May 2022 8:57
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 11 06:00:46 2022






AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	30013	PASS
68	69	0.00	2	0.5	154	PASS
69	198	0.00	100	36.5	30352	PASS
70	69	0.00	2	0.8	244	PASS
127	198	10	80	42.8	35640	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	83269	PASS
199	198	5	9	6.5	5420	PASS
275	198	10	60	33.0	27445	PASS
365	198	1	100	5.6	4687	PASS
441	198	0.01	100	16.1	13418	PASS
442	442	50	100	100.0	84176	PASS
443	442	15	24	18.8	15830	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051822\
Data File : BG053596.D
Acq On : 18 May 2022 8:57
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: May 18 14:28:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 11 06:00:46 2022
Response via : Initial Calibration



TIC: BG053596.D\data.ms

(70) Pentachlorophenol (C)

17.112min (-0.006) 30554.00 ng




response 111260

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	70.20	61.53
264.00	66.20	62.16
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051822\
Data File : BG053596.D
Acq On : 18 May 2022 8:57
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: May 18 14:28:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 11 06:00:46 2022
Response via : Initial Calibration



TIC: BG053596.D\data.ms

(77) Benzidine

19.679min (-0.000) 0.00 ng

response 333671

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.66
183.00	11.10	11.79
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/18/2022	BNA_G	<u>BG053596.D</u>
Compound Name	Response	Retention Time
DDT	293147	20.91
DDD	8958	20.47
DDE	517	19.97
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9475	302622	3.13

Instrument :
BNA_G
ClientSampleId :
DFTPP