

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052924\
 Data File : BG061321.D
 Acq On : 29 May 2024 9:50
 Operator : MA/JU
 Sample : PB161128BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK128

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/30/2024
 Supervised By :mohammad ahmed 06/01/2024

Quant Time: May 30 00:33:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG052124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 21 15:23:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.882	152	24773	20.000	ng/u1	0.00
20) Naphthalene-d8	10.678	136	106416	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.515	164	85372	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.265	188	192543	20.000	ng/u1	0.00
79) Chrysene-d12	21.513	240	205656	20.000	ng/u1	0.00
88) Perylene-d12	24.603	264	235311	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	2615	5.907	ng/uL	0.00
4) Pyridine-d5	3.745	84	37704	23.942	ng/u1	0.00
7) Phenol-d5	7.065	99	54199	26.642	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.212	67	34579	27.042	ng/u1	0.00
11) 2-Chlorophenol-d4	7.418	132	41948	27.861	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	48178	27.773	ng/u1	0.00
21) Nitrobenzene-d5	9.039	128	23410	28.572	ng/u1	0.00
24) 2-Nitrophenol-d4	9.762	143	26363	27.493	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.308	165	52361	28.010	ng/u1	0.00
31) 4-Chloroaniline-d4	10.814	131	68381	27.960	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	176823	26.705	ng/u1	0.00
49) Acenaphthylene-d8	14.210	160	193500	28.110	ng/u1	0.00
54) 4-Nitrophenol-d4	14.744	143	19007m	23.219	ng/u1	0.00
60) Fluorene-d10	15.508	176	155564	27.213	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.643	200	23389	20.084	ng/u1	0.00
73) Anthracene-d10	17.365	188	252681	29.767	ng/u1	0.00
81) Pyrene-d10	19.656	212	300039	28.414	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.392	264	333352	29.467	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052924\
Data File : BG061321.D
Acq On : 29 May 2024 9:50
Operator : MA/JU
Sample : PB161128BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK128

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/30/2024
Supervised By :mohammad ahmed 06/01/2024

Quant Time: May 30 00:33:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG052124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 21 15:23:39 2024
Response via : Initial Calibration

