



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052924\
Data File : BG061319.D
Acq On : 29 May 2024 8:29
Operator : MA/JU
Sample : DFTPP573
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP573

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG052124.MA.M
Title : SVOA CALIBRATION
Last Update : Tue May 21 15:23:39 2024



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.7	34171	PASS
68	69	0.00	2	0.7	169	PASS
69	198	0.00	100	31.7	25373	PASS
70	69	0.00	2	0.4	89	PASS
127	198	10	80	38.2	30595	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	79995	PASS
199	198	5	9	6.8	5472	PASS
275	198	10	60	32.4	25893	PASS
365	198	1	100	5.0	4021	PASS
441	198	0.01	100	14.5	11578	PASS
442	198	50	100	89.6	71699	PASS
443	442	15	24	18.7	13400	PASS

m/z	Abundance
44.00	913.0

Instrument :

BNA_G

ClientSampleId :

DFTPP573

m/z	Abundance
37.00	352.0
38.10	362.0
39.00	3050.0
40.00	275.0
43.90	1064.0
48.90	293.0
50.00	9220.0
51.00	36360.0
52.00	1774.0
55.00	222.0
56.00	1116.0
57.00	2628.0
61.00	416.0
61.90	394.0
63.10	926.0
64.00	183.0
65.00	529.0
69.00	24712.0
69.90	267.0
73.10	342.0
74.00	2884.0
75.00	4027.0
76.10	1869.0
77.00	28664.0
78.00	2540.0
79.00	2110.0
80.00	1664.0
80.90	1702.0
82.00	342.0
82.90	649.0
84.90	378.0
86.00	861.0
87.10	308.0
87.90	160.0
91.10	518.0
92.00	552.0
93.00	3055.0
96.00	186.0
98.00	3122.0
99.00	2269.0
99.90	191.0
100.90	1416.0
102.90	222.0
104.00	683.0
105.00	988.0
107.00	9425.0
108.00	1362.0
110.00	15620.0
111.00	2786.0
112.00	277.0
116.00	549.0
117.00	8105.0
118.00	457.0
122.00	591.0
123.00	1026.0
123.80	498.0
125.00	447.0
127.00	30352.0
128.10	2698.0
129.00	12620.0
130.00	1347.0
130.90	217.0
132.00	265.0
134.00	361.0
135.00	881.0
136.00	492.0
137.00	704.0
138.10	157.0
139.90	270.0
140.90	1473.0
141.90	466.0
143.00	431.0
146.10	375.0
147.00	974.0
147.90	2724.0
149.00	290.0
150.90	158.0
153.00	540.0
154.00	483.0
155.00	917.0
156.00	1396.0
156.90	355.0
158.00	566.0
159.10	280.0
160.10	858.0
161.00	942.0
162.00	234.0
165.00	713.0
166.00	604.0
167.00	3255.0
168.00	1713.0
168.90	229.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

169.10	229.0
172.10	333.0
173.00	393.0
174.00	1098.0
175.00	1664.0
176.10	461.0
176.90	1048.0
177.90	367.0
179.00	3050.0
180.00	1943.0
181.00	965.0
182.00	182.0
183.80	281.0
185.00	1637.0
186.00	11759.0
187.00	2938.0
187.90	314.0
188.90	783.0
190.10	267.0
191.00	320.0
192.00	862.0
193.00	1281.0
193.90	230.0
194.90	203.0
195.90	1783.0
198.00	74360.0
199.00	5511.0
199.80	416.0
201.60	646.0
202.20	405.0
203.00	732.0
204.00	3101.0
205.00	5271.0
206.00	22152.0
207.10	2982.0
208.00	1051.0
209.00	200.0
210.30	450.0
210.90	826.0
212.10	273.0
215.10	218.0
216.00	704.0
217.00	6525.0
218.10	854.0
219.40	157.0
221.00	3613.0
221.90	701.0
223.00	1610.0
224.00	13201.0
225.00	3339.0
225.90	309.0
227.00	6797.0
228.00	605.0
229.00	1371.0
230.00	275.0
231.00	427.0
233.80	283.0
235.10	479.0
237.00	433.0
239.00	219.0
240.00	265.0
241.10	287.0
242.00	737.0
242.90	717.0
244.10	10354.0
245.10	1506.0
246.00	2241.0
247.00	373.0
248.80	259.0
251.90	249.0
252.10	247.0
253.00	308.0
255.00	56568.0
256.00	8307.0
257.10	624.0
258.10	2976.0
259.10	516.0
265.10	1035.0
271.00	191.0
272.10	255.0
273.00	2150.0
274.00	4295.0
275.00	22760.0
276.00	3139.0
277.00	2151.0
277.80	386.0
282.10	151.0
282.90	198.0
284.10	194.0
285.10	438.0
288.90	301.0
290.90	150.0
293.00	640.0
296.00	6982.0
297.00	976.0
298.00	177.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

303.10	705.0
304.20	270.0
313.60	213.0
314.10	269.0
315.00	733.0
316.00	344.0
323.00	2286.0
323.80	269.0
324.20	309.0
326.90	444.0
328.00	154.0
332.10	222.0
333.00	307.0
334.10	1280.0
335.00	553.0
345.80	335.0
352.10	538.0
353.10	462.0
354.10	871.0
355.30	195.0
359.00	152.0
364.90	3062.0
366.10	289.0
370.90	213.0
372.10	901.0
373.00	275.0
382.80	289.0
402.00	342.0
402.90	629.0
404.00	322.0
421.00	776.0
423.00	3077.0
424.00	699.0
441.00	8099.0
442.00	54304.0
443.00	10777.0
444.00	1023.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

m/z	Abundance
36.90	249.0
38.10	349.0
39.00	3060.0
39.80	186.0
41.00	253.0
44.00	812.0
44.90	193.0
50.00	9971.0
51.00	40704.0
52.00	1961.0
55.10	330.0
56.00	1388.0
57.00	2600.0
61.00	368.0
61.90	405.0
63.00	1498.0
65.10	495.0
69.00	30152.0
70.10	244.0
72.90	369.0
74.00	2876.0
75.00	4457.0
77.00	32528.0
78.00	2075.0
79.00	2406.0
79.90	1650.0
81.00	2324.0
82.00	840.0
83.00	352.0
85.00	414.0
86.00	656.0
86.90	202.0
91.00	641.0
92.00	530.0
93.00	3754.0
93.90	170.0
96.00	188.0
98.00	3100.0
99.00	2101.0
100.00	185.0
101.00	1533.0
102.90	365.0
104.00	1078.0
105.00	998.0
106.00	218.0
107.00	11439.0
108.00	1522.0
110.00	18424.0
111.00	2587.0
111.90	219.0
116.10	800.0
117.00	8727.0
117.90	791.0
120.10	272.0
122.00	761.0
123.00	1395.0
124.00	680.0
125.00	456.0
127.00	35688.0
128.00	2876.0
129.00	16185.0
130.00	1173.0
130.80	173.0
131.00	173.0
134.00	696.0
135.00	1465.0
136.00	557.0
137.10	662.0
138.90	190.0
141.00	2160.0
142.00	883.0
143.00	396.0
145.00	182.0
145.90	403.0
147.00	1207.0
148.00	2195.0
148.90	588.0
151.20	363.0
153.00	785.0
154.00	723.0
155.00	1104.0
156.00	2048.0
157.10	370.0
157.90	495.0
158.90	227.0
160.00	592.0
161.00	945.0
162.10	497.0
163.90	266.0
164.90	1015.0
166.00	875.0
167.00	4116.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

168.10	1957.0
169.10	519.0
171.00	252.0
172.10	534.0
173.00	632.0
174.00	1046.0
175.00	2131.0
175.90	333.0
176.90	843.0
178.00	637.0
179.00	3787.0
180.00	2583.0
181.00	1150.0
182.20	291.0
182.90	213.0
184.00	306.0
185.00	2225.0
186.00	14033.0
187.10	3777.0
188.00	517.0
188.90	1028.0
190.00	230.0
191.10	632.0
191.90	1037.0
193.10	1557.0
193.90	316.0
196.00	2704.0
198.00	94520.0
199.00	6056.0
199.90	523.0
201.60	446.0
203.00	803.0
204.00	4175.0
205.00	7854.0
206.00	28336.0
207.00	4126.0
208.00	1206.0
209.00	193.0
210.20	560.0
210.90	1215.0
212.20	183.0
214.90	401.0
216.10	852.0
217.00	8525.0
218.00	1126.0
221.00	4037.0
221.70	1008.0
221.90	994.0
223.00	1745.0
224.00	17192.0
225.10	4511.0
226.20	565.0
227.00	7775.0
227.90	895.0
229.00	1258.0
229.90	221.0
231.10	345.0
231.90	167.0
234.00	511.0
235.00	500.0
236.00	394.0
236.90	568.0
238.90	358.0
240.00	337.0
241.00	777.0
242.00	788.0
243.10	1023.0
244.00	11251.0
245.10	1969.0
246.00	3396.0
247.00	576.0
248.90	445.0
250.80	220.0
251.70	166.0
253.00	487.0
253.80	726.0
255.00	74256.0
256.00	11133.0
257.00	884.0
258.00	5041.0
259.00	719.0
265.10	1986.0
266.10	321.0
270.80	166.0
272.00	269.0
273.00	2427.0
274.10	5472.0
275.00	30968.0
276.00	4299.0
277.00	3461.0
278.10	568.0
279.10	171.0
283.00	342.0
283.90	375.0
284.90	411.0
289.10	199.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

292.00	233.0
293.00	803.0
296.00	9505.0
297.00	1430.0
301.90	232.0
303.10	1386.0
304.10	437.0
309.00	213.0
310.00	164.0
314.10	540.0
314.90	1237.0
316.00	562.0
321.10	578.0
323.10	3308.0
324.10	696.0
327.00	419.0
327.90	364.0
332.10	227.0
333.00	301.0
334.10	2029.0
335.10	399.0
340.20	198.0
340.90	254.0
346.00	1065.0
347.00	166.0
352.10	817.0
353.00	593.0
354.00	1310.0
355.10	302.0
365.00	4741.0
366.00	557.0
370.90	201.0
372.00	1452.0
373.00	336.0
383.00	543.0
383.80	230.0
390.10	322.0
402.00	786.0
403.00	781.0
403.90	362.0
421.00	910.0
422.10	752.0
423.00	4587.0
423.90	847.0
441.00	13566.0
442.00	79912.0
443.00	14773.0
444.00	1628.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

m/z	Abundance
38.00	262.0
39.10	1976.0
40.90	208.0
44.00	994.0
49.00	254.0
50.00	6639.0
51.00	25448.0
52.00	1676.0
56.00	708.0
57.00	2253.0
60.90	280.0
62.00	275.0
63.00	906.0
65.00	363.0
67.90	508.0
69.00	21256.0
73.00	196.0
74.00	1771.0
75.00	3401.0
76.10	1123.0
77.00	20880.0
78.00	1246.0
79.00	1227.0
80.00	1389.0
81.00	1284.0
82.00	431.0
82.90	368.0
85.10	287.0
86.00	583.0
86.90	213.0
91.00	468.0
92.10	547.0
93.00	2594.0
94.00	170.0
98.00	2200.0
99.00	1748.0
100.00	241.0
101.00	1201.0
103.00	324.0
103.90	667.0
105.00	537.0
107.00	7744.0
108.00	1085.0
110.00	12598.0
111.00	2063.0
112.00	334.0
115.90	258.0
117.00	6096.0
118.10	351.0
121.80	393.0
123.00	715.0
124.00	495.0
124.90	576.0
127.00	25744.0
128.00	1902.0
129.00	11333.0
130.00	784.0
130.90	214.0
134.00	304.0
135.00	1261.0
136.00	492.0
137.00	473.0
139.80	161.0
141.00	1617.0
142.10	568.0
143.00	484.0
143.90	205.0
146.00	294.0
147.00	1230.0
148.00	1521.0
149.00	215.0
151.10	224.0
153.00	608.0
154.00	463.0
155.00	795.0
156.00	1120.0
157.10	207.0
158.00	267.0
159.00	206.0
160.00	592.0
161.00	630.0
164.90	865.0
166.10	635.0
167.00	2815.0
167.90	1455.0
169.10	269.0
170.00	169.0
171.00	363.0
171.90	362.0
173.00	356.0
173.90	748.0
175.00	1283.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

176.00	415.0
176.90	755.0
178.10	229.0
178.90	2676.0
180.00	1499.0
181.00	676.0
184.00	354.0
185.00	1433.0
186.00	11049.0
187.00	2881.0
188.00	267.0
189.10	940.0
191.00	322.0
191.90	1014.0
193.00	949.0
194.10	248.0
195.00	283.0
196.00	2049.0
198.00	71104.0
199.00	4848.0
200.10	340.0
201.50	193.0
201.90	182.0
203.00	667.0
204.00	3410.0
205.10	5156.0
206.00	21464.0
207.00	2978.0
208.00	779.0
209.10	250.0
210.10	316.0
211.00	1113.0
211.90	250.0
214.90	308.0
216.10	576.0
217.00	5979.0
218.00	934.0
221.10	2855.0
221.90	832.0
223.00	1553.0
224.00	11780.0
225.00	2963.0
226.00	315.0
227.00	6071.0
228.00	968.0
229.00	1583.0
230.10	211.0
231.10	409.0
233.00	172.0
234.10	446.0
235.00	572.0
236.00	354.0
237.00	540.0
238.90	227.0
240.00	319.0
241.00	282.0
242.10	869.0
243.10	823.0
244.00	10329.0
245.00	1319.0
246.00	2157.0
247.00	440.0
249.10	345.0
250.00	196.0
252.10	298.0
253.20	351.0
255.00	59472.0
256.00	8430.0
257.00	599.0
258.00	3607.0
259.10	657.0
265.00	1582.0
265.90	192.0
271.10	202.0
271.90	197.0
273.00	2204.0
274.00	4129.0
275.00	23952.0
276.00	2993.0
277.00	1908.0
277.90	358.0
283.00	327.0
284.10	247.0
285.10	252.0
288.90	209.0
292.10	190.0
293.00	613.0
294.10	438.0
296.00	8056.0
296.90	1095.0
302.00	181.0
303.10	932.0
304.10	260.0
315.10	714.0
316.00	607.0
320.80	298.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573

323.00	2383.0
324.00	460.0
327.00	561.0
327.80	253.0
332.00	363.0
332.90	232.0
333.20	232.0
334.10	1532.0
334.90	549.0
341.10	529.0
345.80	426.0
346.10	453.0
351.90	857.0
353.00	459.0
354.00	898.0
355.10	265.0
365.10	4260.0
366.00	606.0
372.10	1562.0
372.90	274.0
383.00	342.0
390.20	209.0
401.90	594.0
403.00	783.0
404.10	347.0
420.90	627.0
422.00	516.0
423.00	4639.0
424.00	848.0
425.20	163.0
441.00	13069.0
442.00	80880.0
443.00	14651.0
443.90	1208.0

Instrument :
BNA_G
ClientSampleId :
DFTPP573