

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG053119\  
 Data File : BG040975.D  
 Acq On : 31 May 2019 12:47  
 Operator : HP/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: May 31 19:29:41 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed May 29 18:39:57 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                     | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------------|--------|--------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4       | 20.000 | 20.000 | 0.0  | 98    | -0.01    |
| 2    | 1,4-Dioxane                  | 40.000 | 40.779 | -1.9 | 100   | 0.00     |
| 3    | Pyridine                     | 40.000 | 38.654 | 3.4  | 94    | 0.00     |
| 4    | n-Nitrosodimethylamine       | 40.000 | 41.225 | -3.1 | 100   | 0.00     |
| 5 S  | 2-Fluorophenol               | 80.000 | 77.101 | 3.6  | 93    | 0.00     |
| 6    | Aniline                      | 40.000 | 38.452 | 3.9  | 94    | -0.01    |
| 7 S  | Phenol-d6                    | 80.000 | 73.951 | 7.6  | 91    | 0.00     |
| 8    | 2-Chlorophenol               | 40.000 | 36.928 | 7.7  | 91    | 0.00     |
| 9    | Benzaldehyde                 | 40.000 | 38.688 | 3.3  | 94    | 0.00     |
| 10 C | Phenol                       | 40.000 | 37.007 | 7.5  | 91    | 0.00     |
| 11   | bis(2-Chloroethyl)ether      | 40.000 | 38.872 | 2.8  | 94    | 0.00     |
| 12   | 1,3-Dichlorobenzene          | 40.000 | 38.444 | 3.9  | 93    | 0.00     |
| 13 C | 1,4-Dichlorobenzene          | 40.000 | 38.609 | 3.5  | 93    | 0.00     |
| 14   | 1,2-Dichlorobenzene          | 40.000 | 38.149 | 4.6  | 93    | 0.00     |
| 15   | Benzyl Alcohol               | 40.000 | 37.766 | 5.6  | 93    | 0.00     |
| 16   | 2,2'-oxybis(1-Chloropropane) | 40.000 | 37.761 | 5.6  | 93    | 0.00     |
| 17   | 2-Methylphenol               | 40.000 | 36.604 | 8.5  | 89    | 0.00     |
| 18   | Hexachloroethane             | 40.000 | 38.946 | 2.6  | 96    | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine   | 40.000 | 37.386 | 6.5  | 92    | -0.01    |
| 20   | 3+4-Methylphenols            | 40.000 | 35.857 | 10.4 | 89    | 0.00     |
| 21 I | Naphthalene-d8               | 20.000 | 20.000 | 0.0  | 92    | 0.00     |
| 22   | Acetophenone                 | 40.000 | 40.804 | -2.0 | 92    | 0.00     |
| 23 S | Nitrobenzene-d5              | 80.000 | 81.699 | -2.1 | 93    | 0.00     |
| 24   | Nitrobenzene                 | 40.000 | 41.559 | -3.9 | 94    | 0.00     |
| 25   | Isophorone                   | 40.000 | 41.401 | -3.5 | 93    | 0.00     |
| 26 C | 2-Nitrophenol                | 40.000 | 42.276 | -5.7 | 95    | 0.00     |
| 27   | 2,4-Dimethylphenol           | 40.000 | 39.043 | 2.4  | 88    | 0.00     |
| 28   | bis(2-Chloroethoxy)methane   | 40.000 | 40.772 | -1.9 | 92    | 0.00     |
| 29 C | 2,4-Dichlorophenol           | 40.000 | 37.607 | 6.0  | 84    | 0.00     |
| 30   | 1,2,4-Trichlorobenzene       | 40.000 | 39.776 | 0.6  | 90    | 0.00     |
| 31   | Naphthalene                  | 40.000 | 39.379 | 1.6  | 89    | 0.00     |
| 32   | Benzoic acid                 | 40.000 | 33.219 | 17.0 | 74    | -0.02    |
| 33   | 4-Chloroaniline              | 40.000 | 39.309 | 1.7  | 89    | 0.00     |
| 34 C | Hexachlorobutadiene          | 40.000 | 41.450 | -3.6 | 97    | 0.00     |
| 35   | Caprolactam                  | 40.000 | 40.550 | -1.4 | 91    | -0.01    |
| 36 C | 4-Chloro-3-methylphenol      | 40.000 | 37.827 | 5.4  | 86    | 0.00     |
| 37   | 2-Methylnaphthalene          | 40.000 | 37.696 | 5.8  | 86    | 0.00     |
| 38   | 1-Methylnaphthalene          | 40.000 | 40.003 | -0.0 | 90    | -0.23    |
| 39 I | Acenaphthene-d10             | 20.000 | 20.000 | 0.0  | 91    | 0.00     |
| 40   | 1,2,4,5-Tetrachlorobenzene   | 40.000 | 38.831 | 2.9  | 86    | 0.00     |
| 41 P | Hexachlorocyclopentadiene    | 40.000 | 43.589 | -9.0 | 107   | 0.00     |
| 42 S | 2,4,6-Tribromophenol         | 80.000 | 78.876 | 1.4  | 88    | 0.00     |
| 43 C | 2,4,6-Trichlorophenol        | 40.000 | 39.174 | 2.1  | 86    | 0.00     |
| 44   | 2,4,5-Trichlorophenol        | 40.000 | 38.396 | 4.0  | 86    | 0.00     |

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG053119\  
 Data File : BG040975.D  
 Acq On : 31 May 2019 12:47  
 Operator : HP/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: May 31 19:29:41 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed May 29 18:39:57 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | Amount | Calc.  | %Dev  | Area% | Dev(min) |
|------|----------------------------|--------|--------|-------|-------|----------|
| 45 S | 2-Fluorobiphenyl           | 80.000 | 79.912 | 0.1   | 88    | 0.00     |
| 46   | 1,1'-Biphenyl              | 40.000 | 39.273 | 1.8   | 86    | 0.00     |
| 47   | 2-Chloronaphthalene        | 40.000 | 39.027 | 2.4   | 87    | 0.00     |
| 48   | 2-Nitroaniline             | 40.000 | 41.707 | -4.3  | 93    | 0.00     |
| 49   | Acenaphthylene             | 40.000 | 39.363 | 1.6   | 86    | 0.00     |
| 50   | Dimethylphthalate          | 40.000 | 39.799 | 0.5   | 87    | 0.00     |
| 51   | 2,6-Dinitrotoluene         | 40.000 | 38.567 | 3.6   | 86    | 0.00     |
| 52 C | Acenaphthene               | 40.000 | 38.435 | 3.9   | 85    | 0.00     |
| 53   | 3-Nitroaniline             | 40.000 | 39.613 | 1.0   | 86    | 0.00     |
| 54 P | 2,4-Dinitrophenol          | 40.000 | 45.234 | -13.1 | 112   | 0.00     |
| 55   | Dibenzofuran               | 40.000 | 39.274 | 1.8   | 86    | 0.00     |
| 56 P | 4-Nitrophenol              | 40.000 | 42.387 | -6.0  | 92    | 0.00     |
| 57   | 2,4-Dinitrotoluene         | 40.000 | 40.886 | -2.2  | 89    | 0.00     |
| 58   | Fluorene                   | 40.000 | 38.320 | 4.2   | 85    | 0.00     |
| 59   | 2,3,4,6-Tetrachlorophenol  | 40.000 | 39.650 | 0.9   | 88    | 0.00     |
| 60   | Diethylphthalate           | 40.000 | 39.141 | 2.1   | 86    | 0.00     |
| 61   | 4-Chlorophenyl-phenylether | 40.000 | 38.108 | 4.7   | 85    | 0.00     |
| 62   | 4-Nitroaniline             | 40.000 | 41.712 | -4.3  | 93    | 0.00     |
| 63   | Azobenzene                 | 40.000 | 38.924 | 2.7   | 84    | 0.00     |
| 64 I | Phenanthrene-d10           | 20.000 | 20.000 | 0.0   | 90    | 0.00     |
| 65   | 4,6-Dinitro-2-methylphenol | 40.000 | 37.653 | 5.9   | 88    | 0.00     |
| 66 c | n-Nitrosodiphenylamine     | 40.000 | 39.069 | 2.3   | 86    | 0.00     |
| 67   | 4-Bromophenyl-phenylether  | 40.000 | 38.764 | 3.1   | 84    | 0.00     |
| 68   | Hexachlorobenzene          | 40.000 | 38.325 | 4.2   | 84    | 0.00     |
| 69   | Atrazine                   | 40.000 | 43.458 | -8.6  | 88    | 0.00     |
| 70 C | Pentachlorophenol          | 40.000 | 39.267 | 1.8   | 86    | 0.00     |
| 71   | Phenanthrene               | 40.000 | 40.514 | -1.3  | 88    | 0.00     |
| 72   | Anthracene                 | 40.000 | 41.283 | -3.2  | 90    | 0.00     |
| 73   | Carbazole                  | 40.000 | 42.033 | -5.1  | 91    | 0.00     |
| 74   | Di-n-butylphthalate        | 40.000 | 41.645 | -4.1  | 88    | 0.00     |
| 75 C | Fluoranthene               | 40.000 | 42.428 | -6.1  | 91    | 0.00     |
| 76 I | Chrysene-d12               | 20.000 | 20.000 | 0.0   | 92    | 0.00     |
| 77   | Benzidine                  | 40.000 | 42.479 | -6.2  | 92    | 0.00     |
| 78   | Pyrene                     | 40.000 | 41.723 | -4.3  | 92    | 0.00     |
| 79 S | Terphenyl-d14              | 80.000 | 84.353 | -5.4  | 91    | 0.00     |
| 80   | Butylbenzylphthalate       | 40.000 | 41.098 | -2.7  | 91    | 0.00     |
| 81   | Benzo(a)anthracene         | 40.000 | 40.634 | -1.6  | 91    | 0.00     |
| 82   | 3,3'-Dichlorobenzidine     | 40.000 | 40.746 | -1.9  | 93    | 0.00     |
| 83   | Chrysene                   | 40.000 | 40.900 | -2.2  | 90    | 0.00     |
| 84   | Bis(2-ethylhexyl)phthalate | 40.000 | 39.844 | 0.4   | 89    | 0.00     |
| 85 c | Di-n-octyl phthalate       | 40.000 | 39.919 | 0.2   | 88    | 0.00     |
| 86   | Indeno(1,2,3-cd)pyrene     | 40.000 | 39.725 | 0.7   | 91    | -0.02    |
| 87 I | Perylene-d12               | 20.000 | 20.000 | 0.0   | 92    | -0.01    |

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG053119\  
Data File : BG040975.D  
Acq On : 31 May 2019 12:47  
Operator : HP/JU  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_G  
LabSampleId :  
SSTDCCC040

Quant Time: May 31 19:29:41 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed May 29 18:39:57 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 88   | Benzo(b)fluoranthene   | 40.000 | 39.805 | 0.5  | 90    | 0.00     |
| 89   | Benzo(k)fluoranthene   | 40.000 | 39.458 | 1.4  | 88    | 0.00     |
| 90 C | Benzo(a)pyrene         | 40.000 | 39.552 | 1.1  | 89    | -0.01    |
| 91   | Dibenzo(a,h)anthracene | 40.000 | 39.200 | 2.0  | 88    | -0.01    |
| 92   | Benzo(q,h,i)perylene   | 40.000 | 39.738 | 0.7  | 90    | -0.02    |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0