

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG053119\
 Data File : BG040989.D
 Acq On : 31 May 2019 23:03
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 01 06:17:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 29 18:39:57 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.503	0.495	1.6	99	0.00
3	Pyridine	1.489	1.480	0.6	99	0.00
4	n-Nitrosodimethylamine	0.691	0.720	-4.2	103	0.00
5 S	2-Fluorophenol	1.109	1.084	2.3	97	0.00
6	Aniline	2.286	2.250	1.6	99	-0.01
7 S	Phenol-d6	1.713	1.601	6.5	94	0.00
8	2-Chlorophenol	1.322	1.246	5.7	95	0.00
9	Benzaldehyde	1.022	1.025	-0.3	100	0.00
10 C	Phenol	1.837	1.757	4.4	96	0.00
11	bis(2-Chloroethyl)ether	1.332	1.314	1.4	97	0.00
12	1,3-Dichlorobenzene	1.579	1.585	-0.4	100	0.00
13 C	1,4-Dichlorobenzene	1.599	1.599	0.0	98	0.00
14	1,2-Dichlorobenzene	1.552	1.520	2.1	97	0.00
15	Benzyl Alcohol	1.585	1.551	2.1	98	0.00
16	2,2'-oxybis(1-Chloropropane	2.177	2.165	0.6	100	0.00
17	2-Methylphenol	1.330	1.217	8.5	91	0.00
18	Hexachloroethane	0.616	0.613	0.5	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.318	1.283	2.7	97	0.00
20	3+4-Methylphenols	1.842	1.704	7.5	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
22	Acetophenone	0.561	0.576	-2.7	97	0.00
23 S	Nitrobenzene-d5	0.451	0.466	-3.3	99	0.00
24	Nitrobenzene	0.459	0.476	-3.7	98	0.00
25	Isophorone	0.857	0.870	-1.5	95	0.00
26 C	2-Nitrophenol	0.189	0.189	0.0	93	0.00
27	2,4-Dimethylphenol	0.313	0.307	1.9	93	0.00
28	bis(2-Chloroethoxy)methane	0.463	0.472	-1.9	96	0.00
29 C	2,4-Dichlorophenol	0.397	0.382	3.8	90	0.00
30	1,2,4-Trichlorobenzene	0.452	0.458	-1.3	95	0.00
31	Naphthalene	1.074	1.070	0.4	94	0.00
32	Benzoic acid	0.223	0.201	9.9	84	-0.01
33	4-Chloroaniline	0.498	0.493	1.0	94	0.00
34 C	Hexachlorobutadiene	0.346	0.362	-4.6	102	0.00
35	Caprolactam	0.131	0.124	5.3	89	-0.01
36 C	4-Chloro-3-methylphenol	0.453	0.431	4.9	90	0.00
37	2-Methylnaphthalene	0.827	0.803	2.9	92	0.00
38	1-Methylnaphthalene	0.779	0.803	-3.1	97	-0.23
39 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	0.806	0.821	-1.9	93	0.00
41 P	Hexachlorocyclopentadiene	0.361	0.462	-28.0#	118	0.00
42 S	2,4,6-Tribromophenol	0.297	0.295	0.7	91	0.00
43 C	2,4,6-Trichlorophenol	0.521	0.515	1.2	89	0.00
44	2,4,5-Trichlorophenol	0.550	0.526	4.4	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG053119\
 Data File : BG040989.D
 Acq On : 31 May 2019 23:03
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 01 06:17:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 29 18:39:57 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.389	1.420	-2.2	93	0.00
46	1,1'-Biphenyl	1.480	1.487	-0.5	90	0.00
47	2-Chloronaphthalene	1.271	1.278	-0.6	92	0.00
48	2-Nitroaniline	0.456	0.468	-2.6	94	0.00
49	Acenaphthylene	1.913	1.904	0.5	89	0.00
50	Dimethylphthalate	1.693	1.683	0.6	90	0.00
51	2,6-Dinitrotoluene	0.365	0.362	0.8	91	0.00
52 C	Acenaphthene	1.159	1.136	2.0	89	0.00
53	3-Nitroaniline	0.365	0.373	-2.2	91	0.00
54 P	2,4-Dinitrophenol	0.142	0.235	-65.5#	160#	0.00
55	Dibenzofuran	1.950	1.923	1.4	88	0.00
56 P	4-Nitrophenol	0.250	0.262	-4.8	93	0.00
57	2,4-Dinitrotoluene	0.516	0.520	-0.8	90	0.00
58	Fluorene	1.553	1.510	2.8	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.540	0.544	-0.7	91	0.00
60	Diethylphthalate	1.728	1.690	2.2	88	0.00
61	4-Chlorophenyl-phenylether	1.009	0.989	2.0	90	0.00
62	4-Nitroaniline	0.386	0.384	0.5	91	0.00
63	Azobenzene	1.570	1.564	0.4	89	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	0.096	0.110	-14.6	104	0.00
66 c	n-Nitrosodiphenylamine	0.562	0.559	0.5	89	0.00
67	4-Bromophenyl-phenylether	0.270	0.266	1.5	87	0.00
68	Hexachlorobenzene	0.287	0.281	2.1	88	0.00
69	Atrazine	0.217	0.233	-7.4	88	0.00
70 C	Pentachlorophenol	0.139	0.144	-3.6	89	0.00
71	Phenanthrene	1.042	1.052	-1.0	89	0.00
72	Anthracene	1.027	1.049	-2.1	90	0.00
73	Carbazole	0.882	0.917	-4.0	92	0.00
74	Di-n-butylphthalate	1.096	1.112	-1.5	88	0.00
75 C	Fluoranthene	1.287	1.361	-5.7	93	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
77	Benzidine	0.477	0.524	-9.9	97	0.00
78	Pyrene	1.300	1.332	-2.5	92	0.00
79 S	Terphenyl-d14	0.942	0.979	-3.9	92	0.00
80	Butylbenzylphthalate	0.506	0.514	-1.6	92	0.00
81	Benzo(a)anthracene	1.331	1.341	-0.8	92	0.00
82	3,3'-Dichlorobenzidine	0.468	0.480	-2.6	96	0.00
83	Chrysene	1.274	1.294	-1.6	92	0.00
84	Bis(2-ethylhexyl)phthalate	0.712	0.709	0.4	91	0.00
85 c	Di-n-octyl phthalate	1.186	1.182	0.3	90	0.00
86	Indeno(1,2,3-cd)pyrene	1.561	1.537	1.5	92	-0.03
87 I	Perylene-d12	1.000	1.000	0.0	92	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG053119\
Data File : BG040989.D
Acq On : 31 May 2019 23:03
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 01 06:17:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 29 18:39:57 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.213	1.236	-1.9	92	0.00
89	Benzo(k)fluoranthene	1.200	1.217	-1.4	91	-0.01
90 C	Benzo(a)pyrene	1.157	1.157	0.0	90	-0.01
91	Dibenzo(a,h)anthracene	1.156	1.139	1.5	88	-0.02
92	Benzo(q,h,i)perylene	1.124	1.123	0.1	91	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0