

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060322\
 Data File : BG053806.D
 Acq On : 3 Jun 2022 10:18
 Operator : CG/JU
 Sample : PB145213TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB145213TB

Quant Time: Jun 03 13:28:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 02 23:22:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.090	152	24571	20.000	ng	0.00
21) Naphthalene-d8	10.898	136	98174	20.000	ng	0.00
39) Acenaphthene-d10	14.711	164	67297	20.000	ng	0.00
64) Phenanthrene-d10	17.455	188	173971	20.000	ng	0.00
76) Chrysene-d12	21.732	240	212260	20.000	ng	# 0.00
86) Perylene-d12	24.999	264	203780	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.651	112	202902	151.642	ng	0.00
7) Phenol-d6	7.238	99	260219	143.569	ng	0.00
23) Nitrobenzene-d5	9.247	82	163221	109.178	ng	0.00
42) 2,4,6-Tribromophenol	16.198	330	127972	156.657	ng	0.00
45) 2-Fluorobiphenyl	13.342	172	436634	97.063	ng	0.00
79) Terphenyl-d14	20.064	244	1001699	93.884	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG060322\
Data File : BG053806.D
Acq On : 3 Jun 2022 10:18
Operator : CG/JU
Sample : PB145213TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB145213TB

Quant Time: Jun 03 13:28:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 02 23:22:27 2022
Response via : Initial Calibration

