

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG060619\
 Data File : BG041082.D
 Acq On : 6 Jun 2019 9:14
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 23:22:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 29 18:39:57 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	-0.02
2	1,4-Dioxane	0.503	0.530	-5.4	76	-0.02
3	Pyridine	1.489	1.492	-0.2	71	-0.01
4	n-Nitrosodimethylamine	0.691	0.750	-8.5	77	-0.02
5 S	2-Fluorophenol	1.109	1.162	-4.8	74	-0.01
6	Aniline	2.286	2.183	4.5	69	-0.02
7 S	Phenol-d6	1.713	1.589	7.2	66	-0.01
8	2-Chlorophenol	1.322	1.296	2.0	71	-0.02
9	Benzaldehyde	1.022	0.996	2.5	69	-0.02
10 C	Phenol	1.837	1.739	5.3	68	-0.02
11	bis(2-Chloroethyl)ether	1.332	1.297	2.6	69	-0.01
12	1,3-Dichlorobenzene	1.579	1.631	-3.3	73	-0.02
13 C	1,4-Dichlorobenzene	1.599	1.634	-2.2	72	-0.02
14	1,2-Dichlorobenzene	1.552	1.565	-0.8	72	-0.02
15	Benzyl Alcohol	1.585	1.486	6.2	67	-0.02
16	2,2'-oxybis(1-Chloropropane	2.177	2.078	4.5	68	-0.02
17	2-Methylphenol	1.330	1.195	10.2	64	-0.02
18	Hexachloroethane	0.616	0.657	-6.7	77	-0.02
19 P	n-Nitroso-di-n-propylamine	1.318	1.188	9.9	64	-0.02
20	3+4-Methylphenols	1.842	1.630	11.5	64	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	67	-0.02
22	Acetophenone	0.561	0.564	-0.5	66	-0.02
23 S	Nitrobenzene-d5	0.451	0.454	-0.7	67	-0.02
24	Nitrobenzene	0.459	0.463	-0.9	67	-0.01
25	Isophorone	0.857	0.804	6.2	62	-0.02
26 C	2-Nitrophenol	0.189	0.190	-0.5	66	-0.02
27	2,4-Dimethylphenol	0.313	0.299	4.5	63	-0.02
28	bis(2-Chloroethoxy)methane	0.463	0.449	3.0	64	-0.02
29 C	2,4-Dichlorophenol	0.397	0.361	9.1	60	-0.02
30	1,2,4-Trichlorobenzene	0.452	0.451	0.2	66	-0.02
31	Naphthalene	1.074	1.047	2.5	64	-0.02
32	Benzoic acid	0.223	0.185	17.0	54	-0.03
33	4-Chloroaniline	0.498	0.469	5.8	62	-0.02
34 C	Hexachlorobutadiene	0.346	0.353	-2.0	69	-0.02
35	Caprolactam	0.131	0.116	11.5	58	-0.02
36 C	4-Chloro-3-methylphenol	0.453	0.401	11.5	59	-0.02
37	2-Methylnaphthalene	0.827	0.764	7.6	61	-0.02
38	1-Methylnaphthalene	0.779	0.711	8.7	60	-0.02
39 I	Acenaphthene-d10	1.000	1.000	0.0	61	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.806	0.817	-1.4	60	-0.02
41 P	Hexachlorocyclopentadiene	0.361	0.407	-12.7	68	-0.02
42 S	2,4,6-Tribromophenol	0.297	0.297	0.0	59	-0.02
43 C	2,4,6-Trichlorophenol	0.521	0.522	-0.2	59	-0.02
44	2,4,5-Trichlorophenol	0.550	0.533	3.1	58	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG060619\
 Data File : BG041082.D
 Acq On : 6 Jun 2019 9:14
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 23:22:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 29 18:39:57 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.389	1.422	-2.4	60	-0.02
46	1,1'-Biphenyl	1.480	1.483	-0.2	58	-0.02
47	2-Chloronaphthalene	1.271	1.300	-2.3	60	-0.02
48	2-Nitroaniline	0.456	0.479	-5.0	62	-0.02
49	Acenaphthylene	1.913	1.902	0.6	58	-0.02
50	Dimethylphthalate	1.693	1.678	0.9	58	-0.02
51	2,6-Dinitrotoluene	0.365	0.361	1.1	59	-0.02
52 C	Acenaphthene	1.159	1.121	3.3	57	-0.02
53	3-Nitroaniline	0.365	0.366	-0.3	58	-0.02
54 P	2,4-Dinitrophenol	0.142	0.170	-19.7	75	-0.01
55	Dibenzofuran	1.950	1.951	-0.1	58	-0.01
56 P	4-Nitrophenol	0.250	0.275	-10.0	63	-0.01
57	2,4-Dinitrotoluene	0.516	0.522	-1.2	59	-0.02
58	Fluorene	1.553	1.545	0.5	59	-0.02
59	2,3,4,6-Tetrachlorophenol	0.540	0.547	-1.3	60	-0.02
60	Diethylphthalate	1.728	1.689	2.3	57	-0.02
61	4-Chlorophenyl-phenylether	1.009	0.987	2.2	58	-0.02
62	4-Nitroaniline	0.386	0.409	-6.0	63	-0.02
63	Azobenzene	1.570	1.530	2.5	56	-0.02
64 I	Phenanthrene-d10	1.000	1.000	0.0	62	-0.02
65	4,6-Dinitro-2-methylphenol	0.096	0.107	-11.5	68	-0.02
66 c	n-Nitrosodiphenylamine	0.562	0.533	5.2	58	-0.02
67	4-Bromophenyl-phenylether	0.270	0.250	7.4	56	-0.02
68	Hexachlorobenzene	0.287	0.271	5.6	57	-0.02
69	Atrazine	0.217	0.217	0.0	56	-0.02
70 C	Pentachlorophenol	0.139	0.149	-7.2	62	-0.02
71	Phenanthrene	1.042	1.043	-0.1	60	-0.02
72	Anthracene	1.027	1.050	-2.2	61	-0.02
73	Carbazole	0.882	0.938	-6.3	64	-0.02
74	Di-n-butylphthalate	1.096	1.114	-1.6	60	-0.02
75 C	Fluoranthene	1.287	1.382	-7.4	64	-0.02
76 I	Chrysene-d12	1.000	1.000	0.0	67	-0.02
77	Benzidine	0.477	0.572	-19.9	76	-0.02
78	Pyrene	1.300	1.315	-1.2	65	-0.02
79 S	Terphenyl-d14	0.942	0.984	-4.5	66	-0.02
80	Butylbenzylphthalate	0.506	0.503	0.6	64	-0.02
81	Benzo(a)anthracene	1.331	1.365	-2.6	67	-0.02
82	3,3'-Dichlorobenzidine	0.468	0.517	-10.5	73	-0.02
83	Chrysene	1.274	1.298	-1.9	66	-0.02
84	Bis(2-ethylhexyl)phthalate	0.712	0.705	1.0	64	-0.02
85 c	Di-n-octyl phthalate	1.186	1.225	-3.3	67	-0.02
86	Indeno(1,2,3-cd)pyrene	1.561	1.568	-0.4	67	-0.06
87 I	Perylene-d12	1.000	1.000	0.0	68	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG060619\
Data File : BG041082.D
Acq On : 6 Jun 2019 9:14
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 06 23:22:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 29 18:39:57 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.213	1.225	-1.0	67	-0.02
89	Benzo(k)fluoranthene	1.200	1.180	1.7	65	-0.03
90 C	Benzo(a)pyrene	1.157	1.139	1.6	65	-0.04
91	Dibenzo(a,h)anthracene	1.156	1.177	-1.8	67	-0.05
92	Benzo(q,h,i)perylene	1.124	1.148	-2.1	69	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0