

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG060719\
 Data File : BG041126.D
 Acq On : 8 Jun 2019 9:11
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 08 18:24:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 08 17:33:39 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.503	0.454	9.7	79	0.00
3	Pyridine	1.489	1.452	2.5	84	0.00
4	n-Nitrosodimethylamine	0.691	0.703	-1.7	87	0.00
5 S	2-Fluorophenol	1.109	1.066	3.9	83	0.00
6	Aniline	2.286	2.286	0.0	87	0.00
7 S	Phenol-d6	1.713	1.656	3.3	84	0.00
8	2-Chlorophenol	1.322	1.313	0.7	87	0.00
9	Benzaldehyde	1.022	1.039	-1.7	88	0.00
10 C	Phenol	1.837	1.764	4.0	84	0.00
11	bis(2-Chloroethyl)ether	1.332	1.320	0.9	85	0.00
12	1,3-Dichlorobenzene	1.579	1.537	2.7	84	0.00
13 C	1,4-Dichlorobenzene	1.599	1.567	2.0	84	0.00
14	1,2-Dichlorobenzene	1.552	1.513	2.5	84	0.00
15	Benzyl Alcohol	1.585	1.626	-2.6	90	0.00
16	2,2'-oxybis(1-Chloropropane	2.177	2.108	3.2	84	0.00
17	2-Methylphenol	1.330	1.254	5.7	82	0.00
18	Hexachloroethane	0.616	0.614	0.3	87	0.00
19 P	n-Nitroso-di-n-propylamine	1.318	1.377	-4.5	91	0.00
20	3+4-Methylphenols	1.842	1.751	4.9	84	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
22	Acetophenone	0.561	0.562	-0.2	89	0.00
23 S	Nitrobenzene-d5	0.451	0.446	1.1	88	0.00
24	Nitrobenzene	0.459	0.451	1.7	87	0.00
25	Isophorone	0.857	0.846	1.3	87	0.00
26 C	2-Nitrophenol	0.189	0.189	0.0	88	0.00
27	2,4-Dimethylphenol	0.313	0.297	5.1	84	0.00
28	bis(2-Chloroethoxy)methane	0.463	0.450	2.8	86	0.00
29 C	2,4-Dichlorophenol	0.397	0.366	7.8	81	0.00
30	1,2,4-Trichlorobenzene	0.452	0.437	3.3	85	0.00
31	Naphthalene	1.074	1.037	3.4	85	0.00
32	Benzoic acid	0.223	0.218	2.2	86	0.00
33	4-Chloroaniline	0.498	0.490	1.6	87	0.00
34 C	Hexachlorobutadiene	0.346	0.336	2.9	88	0.00
35	Caprolactam	0.131	0.125	4.6	84	0.00
36 C	4-Chloro-3-methylphenol	0.453	0.426	6.0	83	0.00
37	2-Methylnaphthalene	0.827	0.796	3.7	86	0.00
38	1-Methylnaphthalene	0.779	0.755	3.1	85	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	0.806	0.812	-0.7	85	0.00
41 P	Hexachlorocyclopentadiene	0.361	0.442	-22.4	104	0.00
42 S	2,4,6-Tribromophenol	0.297	0.291	2.0	83	0.00
43 C	2,4,6-Trichlorophenol	0.521	0.526	-1.0	84	0.00
44	2,4,5-Trichlorophenol	0.550	0.545	0.9	84	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG060719\
 Data File : BG041126.D
 Acq On : 8 Jun 2019 9:11
 Operator : HP/JU
 Sample : SSTDC0040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDC0040

Quant Time: Jun 08 18:24:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 08 17:33:39 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.389	1.431	-3.0	86	0.00
46	1,1'-Biphenyl	1.480	1.493	-0.9	83	0.00
47	2-Chloronaphthalene	1.271	1.286	-1.2	85	0.00
48	2-Nitroaniline	0.456	0.471	-3.3	87	0.00
49	Acenaphthylene	1.913	1.910	0.2	82	0.00
50	Dimethylphthalate	1.693	1.671	1.3	82	0.00
51	2,6-Dinitrotoluene	0.365	0.358	1.9	83	0.00
52 C	Acenaphthene	1.159	1.132	2.3	82	0.00
53	3-Nitroaniline	0.365	0.365	0.0	82	0.00
54 P	2,4-Dinitrophenol	0.142	0.195	-37.3#	123	0.00
55	Dibenzofuran	1.950	1.948	0.1	82	0.00
56 P	4-Nitrophenol	0.250	0.260	-4.0	86	0.00
57	2,4-Dinitrotoluene	0.516	0.521	-1.0	83	0.00
58	Fluorene	1.553	1.513	2.6	82	0.00
59	2,3,4,6-Tetrachlorophenol	0.540	0.527	2.4	81	0.00
60	Diethylphthalate	1.728	1.674	3.1	81	0.00
61	4-Chlorophenyl-phenylether	1.009	0.988	2.1	83	0.00
62	4-Nitroaniline	0.386	0.390	-1.0	85	0.00
63	Azobenzene	1.570	1.524	2.9	80	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
65	4,6-Dinitro-2-methylphenol	0.096	0.121	-26.0#	103	0.00
66 c	n-Nitrosodiphenylamine	0.562	0.564	-0.4	82	0.00
67	4-Bromophenyl-phenylether	0.270	0.264	2.2	79	0.00
68	Hexachlorobenzene	0.287	0.274	4.5	78	0.00
69	Atrazine	0.217	0.211	2.8	73	0.00
70 C	Pentachlorophenol	0.139	0.148	-6.5	83	0.00
71	Phenanthrene	1.042	1.049	-0.7	81	0.00
72	Anthracene	1.027	1.032	-0.5	81	0.00
73	Carbazole	0.882	0.897	-1.7	82	0.00
74	Di-n-butylphthalate	1.096	1.086	0.9	78	0.00
75 C	Fluoranthene	1.287	1.315	-2.2	81	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
77	Benzidine	0.477	0.536	-12.4	88	0.00
78	Pyrene	1.300	1.332	-2.5	82	0.00
79 S	Terphenyl-d14	0.942	0.977	-3.7	81	0.00
80	Butylbenzylphthalate	0.506	0.513	-1.4	81	0.00
81	Benzo(a)anthracene	1.331	1.356	-1.9	82	0.00
82	3,3'-Dichlorobenzidine	0.468	0.496	-6.0	87	0.00
83	Chrysene	1.274	1.298	-1.9	82	0.00
84	Bis(2-ethylhexyl)phthalate	0.712	0.722	-1.4	82	0.00
85 c	Di-n-octyl phthalate	1.186	1.211	-2.1	82	0.00
86	Indeno(1,2,3-cd)pyrene	1.561	1.543	1.2	82	0.00
87 I	Perylene-d12	1.000	1.000	0.0	83	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG060719\
Data File : BG041126.D
Acq On : 8 Jun 2019 9:11
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 08 18:24:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG052919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 08 17:33:39 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.213	1.209	0.3	81	0.00
89	Benzo(k)fluoranthene	1.200	1.202	-0.2	81	0.00
90 C	Benzo(a)pyrene	1.157	1.167	-0.9	81	0.00
91	Dibenzo(a,h)anthracene	1.156	1.158	-0.2	80	0.00
92	Benzo(q,h,i)perylene	1.124	1.141	-1.5	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0