




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061221\
 Data File : BG048930.D
 Acq On : 13 Jun 2021 5:41
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 14 05:56:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 16:03:10 2021
 Response via : Initial Calibration



TIC: BG048930.D\data.ms

(70) Pentachlorophenol (C)




17.154min (-0.005) 0.00 ng

response 107228

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	63.77
264.00	61.30	59.22
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061221\
Data File : BG048930.D
Acq On : 13 Jun 2021 5:41
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 14 05:56:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 11 16:03:10 2021
Response via : Initial Calibration



TIC: BG048930.D\data.ms

(77) Benzidine

19.739min (-0.005) 0.00 ng

response 570609

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	15.41
183.00	11.70	12.99
0.00	0.00	0.00