

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061221\
 Data File : BG048914.D
 Acq On : 12 Jun 2021 15:59
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 14 03:15:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 16:03:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	36.734	8.2	104	0.00
3	Pyridine	40.000	37.364	6.6	100	0.00
4	n-Nitrosodimethylamine	40.000	41.210	-3.0	107	0.00
5 S	2-Fluorophenol	80.000	72.940	8.8	101	0.00
6	Aniline	40.000	36.731	8.2	101	0.00
7 S	Phenol-d6	80.000	72.961	8.8	99	0.00
8	2-Chlorophenol	40.000	36.397	9.0	102	0.00
9	Benzaldehyde	40.000	35.059	12.4	101	0.00
10 C	Phenol	40.000	36.876	7.8	102	0.00
11	bis(2-Chloroethyl)ether	40.000	35.299	11.8	97	0.00
12	1,3-Dichlorobenzene	40.000	36.933	7.7	101	0.00
13 C	1,4-Dichlorobenzene	40.000	36.141	9.6	101	0.00
14	1,2-Dichlorobenzene	40.000	36.482	8.8	102	0.00
15	Benzyl Alcohol	40.000	35.930	10.2	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.890	7.8	102	0.00
17	2-Methylphenol	40.000	36.180	9.6	99	0.00
18	Hexachloroethane	40.000	37.733	5.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.587	11.0	100	0.00
20	3+4-Methylphenols	40.000	36.567	8.6	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	35.386	11.5	102	0.00
23 S	Nitrobenzene-d5	80.000	72.764	9.0	103	0.00
24	Nitrobenzene	40.000	36.237	9.4	104	0.00
25	Isophorone	40.000	33.911	15.2	98	0.00
26 C	2-Nitrophenol	40.000	37.763	5.6	107	0.00
27	2,4-Dimethylphenol	40.000	34.510	13.7	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	34.205	14.5	98	0.00
29 C	2,4-Dichlorophenol	40.000	33.888	15.3	96	0.00
30	1,2,4-Trichlorobenzene	40.000	34.327	14.2	100	0.00
31	Naphthalene	40.000	34.410	14.0	99	0.00
32	Benzoic acid	40.000	38.762	3.1	108	0.00
33	4-Chloroaniline	40.000	34.199	14.5	99	0.00
34 C	Hexachlorobutadiene	40.000	34.068	14.8	99	0.00
35	Caprolactam	40.000	35.914	10.2	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	34.512	13.7	99	0.00
37	2-Methylnaphthalene	40.000	33.967	15.1	99	0.00
38	1-Methylnaphthalene	40.000	34.684	13.3	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	33.101	17.2	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.968	15.1	100	0.00
42 S	2,4,6-Tribromophenol	80.000	56.171	29.8#	82	0.00
43 C	2,4,6-Trichlorophenol	40.000	35.467	11.3	103	0.00
44	2,4,5-Trichlorophenol	40.000	36.150	9.6	105	0.00
45 S	2-Fluorobiphenyl	80.000	67.742	15.3	99	0.00
46	1,1'-Biphenyl	40.000	33.896	15.3	99	0.00
47	2-Chloronaphthalene	40.000	33.792	15.5	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG0611221\
 Data File : BG048914.D
 Acq On : 12 Jun 2021 15:59
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 14 03:15:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 16:03:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.161	4.6	109	0.00
49	Acenaphthylene	40.000	33.923	15.2	100	0.00
50	Dimethylphthalate	40.000	33.989	15.0	100	0.00
51	2,6-Dinitrotoluene	40.000	37.160	7.1	107	0.00
52 C	Acenaphthene	40.000	34.139	14.7	99	0.00
53	3-Nitroaniline	40.000	35.821	10.4	103	0.00
54 P	2,4-Dinitrophenol	40.000	37.957	5.1	109	0.00
55	Dibenzofuran	40.000	33.307	16.7	98	0.00
56 P	4-Nitrophenol	40.000	37.073	7.3	103	0.00
57	2,4-Dinitrotoluene	40.000	41.566	-3.9	107	0.00
58	Fluorene	40.000	33.505	16.2	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	34.461	13.8	99	0.00
60	Diethylphthalate	40.000	33.723	15.7	101	0.00
61	4-Chlorophenyl-phenylether	40.000	34.232	14.4	103	0.00
62	4-Nitroaniline	40.000	36.140	9.6	103	0.00
63	Azobenzene	40.000	33.744	15.6	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.048	-2.6	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	34.008	15.0	99	0.00
67	4-Bromophenyl-phenylether	40.000	28.435	28.9#	82	0.00
68	Hexachlorobenzene	40.000	33.426	16.4	98	0.00
69	Atrazine	40.000	37.028	7.4	99	0.00
70 C	Pentachlorophenol	40.000	35.338	11.7	100	0.00
71	Phenanthrene	40.000	33.507	16.2	98	0.00
72	Anthracene	40.000	33.450	16.4	99	0.00
73	Carbazole	40.000	33.745	15.6	99	0.00
74	Di-n-butylphthalate	40.000	33.916	15.2	101	0.00
75 C	Fluoranthene	40.000	33.808	15.5	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	45.497	-13.7	117	0.00
78	Pyrene	40.000	33.401	16.5	98	0.00
79 S	Terphenyl-d14	80.000	67.553	15.6	100	0.00
80	Butylbenzylphthalate	40.000	34.758	13.1	102	0.00
81	Benzo(a)anthracene	40.000	33.819	15.5	99	0.00
82	3,3'-Dichlorobenzidine	40.000	35.956	10.1	104	0.00
83	Chrysene	40.000	34.493	13.8	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	34.409	14.0	102	0.00
85 c	Di-n-octyl phthalate	40.000	34.598	13.5	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.562	11.1	101	0.00
88	Benzo(b)fluoranthene	40.000	34.904	12.7	101	0.00
89	Benzo(k)fluoranthene	40.000	35.404	11.5	100	0.00
90 C	Benzo(a)pyrene	40.000	34.981	12.5	101	0.00
91	Dibenzo(a,h)anthracene	40.000	35.789	10.5	101	0.00
92	Benzo(g,h,i)perylene	40.000	35.740	10.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061221\
Data File : BG048914.D
Acq On : 12 Jun 2021 15:59
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 14 03:15:27 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 11 16:03:10 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0