

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061223\
 Data File : BG057679.D
 Acq On : 12 Jun 2023 13:18
 Operator : CG/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Quant Time: Jun 13 01:10:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 13 01:09:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	26231	20.000	ng	0.00
21) Naphthalene-d8	10.745	136	110460	20.000	ng	0.00
39) Acenaphthene-d10	14.576	164	76203	20.000	ng	0.00
64) Phenanthrene-d10	17.320	188	189620	20.000	ng	0.00
76) Chrysene-d12	21.580	240	174530	20.000	ng	0.00
86) Perylene-d12	24.747	264	201724	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.742	70	4178	2.618	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061223\
Data File : BG057679.D
Acq On : 12 Jun 2023 13:18
Operator : CG/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Quant Time: Jun 13 01:10:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 13 01:09:38 2023
Response via : Initial Calibration

