

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061423\
 Data File : BG057705.D
 Acq On : 14 Jun 2023 13:01
 Operator : CG/JU
 Sample : 03152-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 RE137-CARBON-20230609

Quant Time: Jun 15 01:20:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 03:52:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	34323	20.000	ng	0.00
21) Naphthalene-d8	10.742	136	143906	20.000	ng	0.00
39) Acenaphthene-d10	14.573	164	100405	20.000	ng	0.00
64) Phenanthrene-d10	17.311	188	249278	20.000	ng	0.00
76) Chrysene-d12	21.571	240	207364	20.000	ng	0.00
86) Perylene-d12	24.732	264	242109	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.507	112	64718	30.702	ng	0.00
7) Phenol-d6	7.088	99	95463	29.480	ng	0.00
23) Nitrobenzene-d5	9.097	82	49102	32.084	ng	0.00
42) 2,4,6-Tribromophenol	16.054	330	32613	29.114	ng	0.00
45) 2-Fluorobiphenyl	13.198	172	127224	18.888	ng	0.00
79) Terphenyl-d14	19.932	244	235674	21.013	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061423\
Data File : BG057705.D
Acq On : 14 Jun 2023 13:01
Operator : CG/JU
Sample : 03152-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
RE137-CARBON-20230609

Quant Time: Jun 15 01:20:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 14 03:52:08 2023
Response via : Initial Calibration

