

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061423\
 Data File : BG057712.D
 Acq On : 14 Jun 2023 17:57
 Operator : CG/JU
 Sample : 03167-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-4-GA

Quant Time: Jun 15 01:22:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 03:52:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	45631	20.000	ng	0.00
21) Naphthalene-d8	10.740	136	188874	20.000	ng	0.00
39) Acenaphthene-d10	14.571	164	117096	20.000	ng	0.00
64) Phenanthrene-d10	17.315	188	252120	20.000	ng	0.00
76) Chrysene-d12	21.568	240	204950	20.000	ng	0.00
86) Perylene-d12	24.735	264	248041	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.517	112	370391	132.170	ng	0.00
7) Phenol-d6	7.103	99	505032	117.309	ng	0.00
23) Nitrobenzene-d5	9.101	82	286495	92.602	ng	0.00
42) 2,4,6-Tribromophenol	16.057	330	171634	131.380	ng	0.00
45) 2-Fluorobiphenyl	13.202	172	631224	80.353	ng	0.00
79) Terphenyl-d14	19.935	244	872394	78.699	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061423\
Data File : BG057712.D
Acq On : 14 Jun 2023 17:57
Operator : CG/JU
Sample : 03167-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-4-GA

Quant Time: Jun 15 01:22:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 14 03:52:08 2023
Response via : Initial Calibration

