

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061423\
 Data File : BG057701.D
 Acq On : 14 Jun 2023 10:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 01:19:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 03:52:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	36.688	8.3	97	0.00
3	Pyridine	40.000	40.747	-1.9	99	0.00
4	n-Nitrosodimethylamine	40.000	38.113	4.7	93	0.00
5 S	2-Fluorophenol	80.000	79.647	0.4	98	0.00
6	Aniline	40.000	40.341	-0.9	102	0.00
7 S	Phenol-d6	80.000	81.952	-2.4	102	0.00
8	2-Chlorophenol	40.000	42.446	-6.1	107	0.00
9	Benzaldehyde	40.000	37.616	6.0	97	0.00
10 C	Phenol	40.000	41.098	-2.7	102	0.00
11	bis(2-Chloroethyl)ether	40.000	39.605	1.0	99	0.00
12	1,3-Dichlorobenzene	40.000	38.288	4.3	102	0.00
13 C	1,4-Dichlorobenzene	40.000	39.277	1.8	101	0.00
14	1,2-Dichlorobenzene	40.000	39.956	0.1	102	0.00
15	Benzyl Alcohol	40.000	40.402	-1.0	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.551	3.6	98	0.00
17	2-Methylphenol	40.000	41.103	-2.8	103	0.00
18	Hexachloroethane	40.000	42.299	-5.7	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.337	-0.8	100	0.00
20	3+4-Methylphenols	40.000	41.121	-2.8	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	37.855	5.4	100	0.00
23 S	Nitrobenzene-d5	80.000	77.558	3.1	110	0.00
24	Nitrobenzene	40.000	39.040	2.4	109	0.00
25	Isophorone	40.000	38.274	4.3	102	0.00
26 C	2-Nitrophenol	40.000	42.271	-5.7	117	0.00
27	2,4-Dimethylphenol	40.000	39.881	0.3	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.216	2.0	103	0.00
29 C	2,4-Dichlorophenol	40.000	43.690	-9.2	107	0.00
30	1,2,4-Trichlorobenzene	40.000	38.287	4.3	100	0.00
31	Naphthalene	40.000	38.389	4.0	102	0.00
32	Benzoic acid	40.000	48.546	-21.4	130	0.00
33	4-Chloroaniline	40.000	38.408	4.0	101	0.00
34 C	Hexachlorobutadiene	40.000	37.060	7.3	99	0.00
35	Caprolactam	40.000	43.587	-9.0	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.639	-4.1	104	0.00
37	2-Methylnaphthalene	40.000	38.327	4.2	102	0.00
38	1-Methylnaphthalene	40.000	38.206	4.5	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.578	3.6	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.811	-9.5	108	0.00
42 S	2,4,6-Tribromophenol	80.000	80.525	-0.7	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.521	-1.3	106	0.00
44	2,4,5-Trichlorophenol	40.000	41.908	-4.8	111	0.00
45 S	2-Fluorobiphenyl	80.000	77.752	2.8	101	0.00
46	1,1'-Biphenyl	40.000	39.576	1.1	102	0.00
47	2-Chloronaphthalene	40.000	39.437	1.4	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061423\
 Data File : BG057701.D
 Acq On : 14 Jun 2023 10:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 01:19:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 03:52:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.891	-7.2	118	0.00
49	Acenaphthylene	40.000	39.946	0.1	103	0.00
50	Dimethylphthalate	40.000	41.186	-3.0	104	0.00
51	2,6-Dinitrotoluene	40.000	44.223	-10.6	122	0.00
52 C	Acenaphthene	40.000	39.832	0.4	102	0.00
53	3-Nitroaniline	40.000	43.963	-9.9	118	0.00
54 P	2,4-Dinitrophenol	40.000	43.581	-9.0	123	0.00
55	Dibenzofuran	40.000	39.318	1.7	103	0.00
56 P	4-Nitrophenol	40.000	43.911	-9.8	120	0.00
57	2,4-Dinitrotoluene	40.000	43.771	-9.4	121	0.00
58	Fluorene	40.000	38.503	3.7	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.967	-2.4	108	0.00
60	Diethylphthalate	40.000	41.137	-2.8	103	0.00
61	4-Chlorophenyl-phenylether	40.000	38.537	3.7	101	0.00
62	4-Nitroaniline	40.000	41.793	-4.5	115	0.00
63	Azobenzene	40.000	39.477	1.3	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.968	-7.4	125	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.929	2.7	102	0.00
67	4-Bromophenyl-phenylether	40.000	35.997	10.0	95	0.00
68	Hexachlorobenzene	40.000	38.253	4.4	100	0.00
69	Atrazine	40.000	40.420	-1.1	102	0.00
70 C	Pentachlorophenol	40.000	42.957	-7.4	110	0.00
71	Phenanthrene	40.000	38.228	4.4	102	0.00
72	Anthracene	40.000	38.717	3.2	102	0.00
73	Carbazole	40.000	38.174	4.6	101	0.00
74	Di-n-butylphthalate	40.000	42.338	-5.8	103	0.00
75 C	Fluoranthene	40.000	37.826	5.4	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzydine	40.000	46.058	-15.1	104	0.00
78	Pyrene	40.000	39.520	1.2	100	0.00
79 S	Terphenyl-d14	80.000	78.845	1.4	100	0.00
80	Butylbenzylphthalate	40.000	40.318	-0.8	104	0.00
81	Benzo(a)anthracene	40.000	39.530	1.2	99	0.00
82	3,3'-Dichlorobenzidine	40.000	41.331	-3.3	97	0.00
83	Chrysene	40.000	39.839	0.4	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.428	-11.1	102	0.00
85 c	Di-n-octyl phthalate	40.000	44.362	-10.9	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.801	3.0	97	0.00
88	Benzo(b)fluoranthene	40.000	39.148	2.1	99	0.00
89	Benzo(k)fluoranthene	40.000	38.427	3.9	96	0.00
90 C	Benzo(a)pyrene	40.000	39.267	1.8	97	0.00
91	Dibenzo(a,h)anthracene	40.000	39.250	1.9	98	0.00
92	Benzo(g,h,i)perylene	40.000	39.185	2.0	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061423\
Data File : BG057701.D
Acq On : 14 Jun 2023 10:06
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 15 01:19:15 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 14 03:52:08 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0