

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061523\
 Data File : BG057719.D
 Acq On : 15 Jun 2023 11:04
 Operator : CG/JU
 Sample : PB153430BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB153430BL

Quant Time: Jun 15 13:00:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 03:52:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	25224	20.000	ng	0.00
21) Naphthalene-d8	10.743	136	108052	20.000	ng	0.00
39) Acenaphthene-d10	14.573	164	79296	20.000	ng	0.00
64) Phenanthrene-d10	17.311	188	217215	20.000	ng	0.00
76) Chrysene-d12	21.571	240	202819	20.000	ng	0.00
86) Perylene-d12	24.732	264	240130	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.508	112	316369	204.227	ng	0.00
7) Phenol-d6	7.094	99	437704	183.925	ng	0.00
23) Nitrobenzene-d5	9.097	82	237325	127.583	ng	0.00
42) 2,4,6-Tribromophenol	16.060	330	190934	215.825	ng	0.00
45) 2-Fluorobiphenyl	13.199	172	592858	111.445	ng	0.00
79) Terphenyl-d14	19.938	244	1233832	112.474	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061523\
Data File : BG057719.D
Acq On : 15 Jun 2023 11:04
Operator : CG/JU
Sample : PB153430BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB153430BL

Quant Time: Jun 15 13:00:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 14 03:52:08 2023
Response via : Initial Calibration

