

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG061819\
 Data File : BG041296.D
 Acq On : 18 Jun 2019 14:19
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 18 16:24:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG061719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 17 14:58:02 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.546	0.589	-7.9	126	0.00
3	Pyridine	1.448	1.533	-5.9	115	0.00
4	n-Nitrosodimethylamine	0.772	0.790	-2.3	111	0.00
5 S	2-Fluorophenol	1.088	1.198	-10.1	119	0.00
6	Aniline	2.193	2.265	-3.3	109	0.00
7 S	Phenol-d6	1.578	1.693	-7.3	114	0.00
8	2-Chlorophenol	1.279	1.362	-6.5	113	0.00
9	Benzaldehyde	1.009	1.039	-3.0	114	0.00
10 C	Phenol	1.697	1.777	-4.7	110	0.00
11	bis(2-Chloroethyl)ether	1.287	1.399	-8.7	114	0.00
12	1,3-Dichlorobenzene	1.601	1.708	-6.7	114	0.00
13 C	1,4-Dichlorobenzene	1.636	1.721	-5.2	114	0.00
14	1,2-Dichlorobenzene	1.562	1.641	-5.1	114	0.00
15	Benzyl Alcohol	1.512	1.546	-2.2	110	0.00
16	2,2'-oxybis(1-Chloropropane	2.107	2.235	-6.1	111	-0.01
17	2-Methylphenol	1.213	1.276	-5.2	112	0.00
18	Hexachloroethane	0.648	0.689	-6.3	117	0.00
19 P	n-Nitroso-di-n-propylamine	1.276	1.294	-1.4	106	0.00
20	3+4-Methylphenols	1.615	1.733	-7.3	111	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.582	0.606	-4.1	110	0.00
23 S	Nitrobenzene-d5	0.476	0.495	-4.0	109	0.00
24	Nitrobenzene	0.478	0.502	-5.0	109	0.00
25	Isophorone	0.872	0.890	-2.1	106	0.00
26 C	2-Nitrophenol	0.194	0.209	-7.7	114	0.00
27	2,4-Dimethylphenol	0.291	0.310	-6.5	114	0.00
28	bis(2-Chloroethoxy)methane	0.455	0.466	-2.4	105	0.00
29 C	2,4-Dichlorophenol	0.360	0.383	-6.4	109	0.00
30	1,2,4-Trichlorobenzene	0.462	0.488	-5.6	110	0.00
31	Naphthalene	1.087	1.130	-4.0	108	0.00
32	Benzoic acid	0.189	0.185	2.1	94	-0.02
33	4-Chloroaniline	0.462	0.479	-3.7	107	0.00
34 C	Hexachlorobutadiene	0.366	0.382	-4.4	113	0.00
35	Caprolactam	0.124	0.121	2.4	104	0.00
36 C	4-Chloro-3-methylphenol	0.416	0.430	-3.4	108	0.00
37	2-Methylnaphthalene	0.801	0.822	-2.6	107	0.00
38	1-Methylnaphthalene	0.766	0.778	-1.6	106	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	0.826	0.898	-8.7	107	0.00
41 P	Hexachlorocyclopentadiene	0.556	0.521	6.3	90	0.00
42 S	2,4,6-Tribromophenol	0.307	0.316	-2.9	101	0.00
43 C	2,4,6-Trichlorophenol	0.511	0.553	-8.2	107	0.00
44	2,4,5-Trichlorophenol	0.526	0.560	-6.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG061819\
 Data File : BG041296.D
 Acq On : 18 Jun 2019 14:19
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 18 16:24:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG061719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 17 14:58:02 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.472	1.581	-7.4	105	0.00
46	1,1'-Biphenyl	1.512	1.614	-6.7	105	0.00
47	2-Chloronaphthalene	1.302	1.403	-7.8	107	0.00
48	2-Nitroaniline	0.477	0.491	-2.9	100	0.00
49	Acenaphthylene	1.975	2.055	-4.1	103	0.00
50	Dimethylphthalate	1.780	1.845	-3.7	102	0.00
51	2,6-Dinitrotoluene	0.375	0.383	-2.1	102	0.00
52 C	Acenaphthene	1.160	1.201	-3.5	103	0.00
53	3-Nitroaniline	0.370	0.380	-2.7	101	0.00
54 P	2,4-Dinitrophenol	0.240	0.242	-0.8	98	0.00
55	Dibenzofuran	2.017	2.097	-4.0	102	0.00
56 P	4-Nitrophenol	0.257	0.268	-4.3	96	0.00
57	2,4-Dinitrotoluene	0.548	0.556	-1.5	102	0.00
58	Fluorene	1.586	1.640	-3.4	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.544	0.550	-1.1	97	0.00
60	Diethylphthalate	1.809	1.850	-2.3	101	0.00
61	4-Chlorophenyl-phenylether	1.032	1.048	-1.6	100	0.00
62	4-Nitroaniline	0.395	0.410	-3.8	103	0.00
63	Azobenzene	1.614	1.655	-2.5	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	0.130	0.114	12.3	80	0.00
66 c	n-Nitrosodiphenylamine	0.531	0.577	-8.7	101	0.00
67	4-Bromophenyl-phenylether	0.259	0.281	-8.5	103	0.00
68	Hexachlorobenzene	0.277	0.297	-7.2	100	0.00
69	Atrazine	0.173	0.209	-20.8	94	0.00
70 C	Pentachlorophenol	0.142	0.146	-2.8	92	0.00
71	Phenanthrene	1.067	1.131	-6.0	100	0.00
72	Anthracene	1.052	1.125	-6.9	100	0.00
73	Carbazole	0.920	0.983	-6.8	101	0.00
74	Di-n-butylphthalate	1.148	1.235	-7.6	101	0.00
75 C	Fluoranthene	1.417	1.508	-6.4	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzidine	0.484	0.504	-4.1	97	0.00
78	Pyrene	1.349	1.426	-5.7	99	0.00
79 S	Terphenyl-d14	1.009	1.066	-5.6	98	0.00
80	Butylbenzylphthalate	0.510	0.556	-9.0	105	0.00
81	Benzo(a)anthracene	1.354	1.429	-5.5	102	0.00
82	3,3'-Dichlorobenzidine	0.475	0.517	-8.8	104	0.00
83	Chrysene	1.302	1.371	-5.3	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.695	0.739	-6.3	101	0.00
85 c	Di-n-octyl phthalate	1.176	1.252	-6.5	103	0.00
86	Indeno(1,2,3-cd)pyrene	1.545	1.585	-2.6	100	0.00
87 I	Perylene-d12	1.000	1.000	0.0	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG061819\
Data File : BG041296.D
Acq On : 18 Jun 2019 14:19
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 18 16:24:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG061719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 17 14:58:02 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.239	1.314	-6.1	102	0.00
89	Benzo(k)fluoranthene	1.228	1.269	-3.3	97	0.00
90 C	Benzo(a)pyrene	1.171	1.234	-5.4	101	0.00
91	Dibenzo(a,h)anthracene	1.178	1.227	-4.2	100	0.01
92	Benzo(q,h,i)perylene	1.165	1.178	-1.1	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0