

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG061923\  
 Data File : BG057755.D  
 Acq On : 19 Jun 2023 14:49  
 Operator : CG/JU  
 Sample : SSTDICC010  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDICC010

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 06/20/2023  
 Supervised By :mohammad ahmed 06/20/2023

Quant Time: Jun 20 03:09:01 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG061923.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jun 20 03:06:18 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.931  | 152  | 35006    | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8            | 10.733 | 136  | 155601   | 20.000 | ng    | 0.00     |
| 39) Acenaphthene-d10          | 14.564 | 164  | 109916   | 20.000 | ng    | 0.00     |
| 64) Phenanthrene-d10          | 17.308 | 188  | 271594   | 20.000 | ng    | 0.00     |
| 76) Chrysene-d12              | 21.556 | 240  | 233787   | 20.000 | ng    | 0.00     |
| 86) Perylene-d12              | 24.711 | 264  | 282198   | 20.000 | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 5) 2-Fluorophenol             | 5.498  | 112  | 46895    | 20.656 | ng    | 0.00     |
| 7) Phenol-d6                  | 7.085  | 99   | 70714    | 20.413 | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 9.088  | 82   | 57922    | 19.105 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 16.051 | 330  | 29844    | 19.492 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 13.189 | 172  | 157163   | 21.687 | ng    | 0.00     |
| 79) Terphenyl-d14             | 19.922 | 244  | 273720   | 21.832 | ng    | 0.00     |
| Target Compounds              |        |      |          |        |       |          |
|                               |        |      |          |        |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.406  | 88   | 10881    | 10.529 | ng    | 99       |
| 3) Pyridine                   | 3.800  | 79   | 31821    | 10.107 | ng    | 97       |
| 4) n-Nitrosodimethylamine     | 3.718  | 42   | 14852    | 10.327 | ng    | 94       |
| 6) Aniline                    | 7.249  | 93   | 45383    | 10.232 | ng    | 95       |
| 8) 2-Chlorophenol             | 7.490  | 128  | 23525    | 10.072 | ng    | 95       |
| 9) Benzaldehyde               | 7.067  | 77   | 23977    | 11.008 | ng    | 97       |
| 10) Phenol                    | 7.108  | 94   | 35165    | 10.310 | ng    | 94       |
| 11) bis(2-Chloroethyl)ether   | 7.349  | 93   | 26635    | 10.157 | ng    | 95       |
| 12) 1,3-Dichlorobenzene       | 7.819  | 146  | 24967    | 10.504 | ng    | 94       |
| 13) 1,4-Dichlorobenzene       | 7.966  | 146  | 24687    | 10.250 | ng    | 97       |
| 14) 1,2-Dichlorobenzene       | 8.283  | 146  | 24735    | 10.614 | ng    | 94       |
| 15) Benzyl Alcohol            | 8.160  | 79   | 26008    | 9.819  | ng    | 95       |
| 16) 2,2'-oxybis(1-Chloropr... | 8.454  | 45   | 51687m   | 10.252 | ng    |          |
| 17) 2-Methylphenol            | 8.365  | 107  | 25389    | 10.121 | ng    | 92       |
| 18) Hexachloroethane          | 9.012  | 117  | 8860     | 10.433 | ng    | 92       |
| 19) n-Nitroso-di-n-propyla... | 8.730  | 70   | 24357    | 10.085 | ng    | 99       |
| 20) 3+4-Methylphenols         | 8.689  | 107  | 34157    | 9.857  | ng    | 94       |
| 22) Acetophenone              | 8.747  | 105  | 46077    | 10.560 | ng    | # 97     |
| 24) Nitrobenzene              | 9.129  | 77   | 31160    | 9.875  | ng    | 97       |
| 25) Isophorone                | 9.652  | 82   | 68735    | 9.923  | ng    | 98       |
| 26) 2-Nitrophenol             | 9.840  | 139  | 8874     | 8.840  | ng    | # 86     |
| 27) 2,4-Dimethylphenol        | 9.899  | 122  | 24648    | 9.917  | ng    | 97       |
| 28) bis(2-Chloroethoxy)met... | 10.134 | 93   | 38413    | 10.298 | ng    | 97       |
| 29) 2,4-Dichlorophenol        | 10.375 | 162  | 22480    | 9.525  | ng    | 97       |
| 30) 1,2,4-Trichlorobenzene    | 10.592 | 180  | 27787    | 10.562 | ng    | 99       |
| 31) Naphthalene               | 10.786 | 128  | 86891    | 10.492 | ng    | 100      |
| 32) Benzoic acid              | 9.969  | 122  | 14133    | 7.677  | ng    | 94       |
| 33) 4-Chloroaniline           | 10.886 | 127  | 38862    | 10.128 | ng    | 98       |
| 34) Hexachlorobutadiene       | 11.068 | 225  | 16138    | 10.055 | ng    | 97       |
| 35) Caprolactam               | 11.632 | 113  | 9203     | 9.609  | ng    | # 91     |
| 36) 4-Chloro-3-methylphenol   | 12.002 | 107  | 30232    | 9.808  | ng    | 94       |
| 37) 2-Methylnaphthalene       | 12.390 | 142  | 62198    | 10.304 | ng    | 98       |
| 38) 1-Methylnaphthalene       | 12.607 | 142  | 59206    | 10.390 | ng    | 97       |
| 40) 1,2,4,5-Tetrachloroben... | 12.754 | 216  | 31529    | 10.347 | ng    | 99       |
| 41) Hexachlorocyclopentadiene | 12.737 | 237  | 15275    | 9.442  | ng    | 98       |
| 43) 2,4,6-Trichlorophenol     | 12.995 | 196  | 21097    | 9.669  | ng    | 100      |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG061923\  
 Data File : BG057755.D  
 Acq On : 19 Jun 2023 14:49  
 Operator : CG/JU  
 Sample : SSTDICC010  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDICC010

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 06/20/2023  
 Supervised By :mohammad ahmed 06/20/2023

Quant Time: Jun 20 03:09:01 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG061923.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jun 20 03:06:18 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol     | 13.060 | 196  | 25075    | 9.859  | ng    | 99       |
| 46) 1,1'-Biphenyl             | 13.395 | 154  | 79802    | 10.574 | ng    | 99       |
| 47) 2-Chloronaphthalene       | 13.442 | 162  | 60617    | 10.455 | ng    | 99       |
| 48) 2-Nitroaniline            | 13.636 | 65   | 17138    | 8.318  | ng    | 100      |
| 49) Acenaphthylene            | 14.282 | 152  | 104851   | 10.536 | ng    | 97       |
| 50) Dimethylphthalate         | 14.018 | 163  | 90131    | 10.676 | ng    | 99       |
| 51) 2,6-Dinitrotoluene        | 14.135 | 165  | 13844    | 8.905  | ng    | 98       |
| 52) Acenaphthene              | 14.629 | 154  | 63703    | 10.340 | ng    | 97       |
| 53) 3-Nitroaniline            | 14.458 | 138  | 16556    | 9.450  | ng    | 92       |
| 54) 2,4-Dinitrophenol         | 14.658 | 184  | 5240     | 11.306 | ng    | 94       |
| 55) Dibenzofuran              | 14.958 | 168  | 106417   | 10.742 | ng    | 97       |
| 56) 4-Nitrophenol             | 14.758 | 139  | 12813    | 9.203  | ng    | 94       |
| 57) 2,4-Dinitrotoluene        | 14.917 | 165  | 17825    | 8.455  | ng    | # 83     |
| 58) Fluorene                  | 15.610 | 166  | 84454    | 10.770 | ng    | 98       |
| 59) 2,3,4,6-Tetrachlorophenol | 15.181 | 232  | 20842m   | 9.616  | ng    |          |
| 60) Diethylphthalate          | 15.381 | 149  | 91721    | 10.590 | ng    | 99       |
| 61) 4-Chlorophenyl-phenyle... | 15.604 | 204  | 43866    | 10.520 | ng    | 97       |
| 62) 4-Nitroaniline            | 15.616 | 138  | 17618    | 9.864  | ng    | 97       |
| 63) Azobenzene                | 15.892 | 77   | 93671    | 10.573 | ng    | 95       |
| 65) 4,6-Dinitro-2-methylph... | 15.680 | 198  | 8423     | 11.316 | ng    | 96       |
| 66) n-Nitrosodiphenylamine    | 15.815 | 169  | 74347    | 10.372 | ng    | 100      |
| 67) 4-Bromophenyl-phenylether | 16.497 | 248  | 29747    | 10.328 | ng    | 97       |
| 68) Hexachlorobenzene         | 16.609 | 284  | 29983    | 10.211 | ng    | 96       |
| 69) Atrazine                  | 16.761 | 200  | 26128    | 10.618 | ng    | 97       |
| 70) Pentachlorophenol         | 16.949 | 266  | 19678    | 9.071  | ng    | 96       |
| 71) Phenanthrene              | 17.349 | 178  | 139638   | 10.619 | ng    | 100      |
| 72) Anthracene                | 17.437 | 178  | 140594   | 10.470 | ng    | 99       |
| 73) Carbazole                 | 17.707 | 167  | 132117   | 10.671 | ng    | 99       |
| 74) Di-n-butylphthalate       | 18.271 | 149  | 150419   | 10.261 | ng    | 98       |
| 75) Fluoranthene              | 19.358 | 202  | 166853   | 10.651 | ng    | 98       |
| 77) Benzidine                 | 19.541 | 184  | 57761    | 10.540 | ng    | 98       |
| 78) Pyrene                    | 19.717 | 202  | 176093   | 10.559 | ng    | 98       |
| 80) Butylbenzylphthalate      | 20.610 | 149  | 59604    | 9.395  | ng    | 99       |
| 81) Benzo(a)anthracene        | 21.538 | 228  | 165981   | 10.348 | ng    | 99       |
| 82) 3,3'-Dichlorobenzidine    | 21.456 | 252  | 55479    | 10.169 | ng    | 98       |
| 83) Chrysene                  | 21.603 | 228  | 156696   | 10.179 | ng    | 99       |
| 84) Bis(2-ethylhexyl)phtha... | 21.456 | 149  | 94188    | 10.226 | ng    | 99       |
| 85) Di-n-octyl phthalate      | 22.649 | 149  | 156763   | 9.703  | ng    | 98       |
| 87) Indeno(1,2,3-cd)pyrene    | 28.283 | 276  | 183623   | 9.952  | ng    | 99       |
| 88) Benzo(b)fluoranthene      | 23.706 | 252  | 157734   | 10.200 | ng    | 99       |
| 89) Benzo(k)fluoranthene      | 23.771 | 252  | 162812   | 10.337 | ng    | 99       |
| 90) Benzo(a)pyrene            | 24.558 | 252  | 153263   | 10.092 | ng    | 99       |
| 91) Dibenzo(a,h)anthracene    | 28.360 | 278  | 152980   | 9.993  | ng    | 100      |
| 92) Benzo(g,h,i)perylene      | 29.411 | 276  | 152712   | 9.963  | ng    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG061923\  
 Data File : BG057755.D  
 Acq On : 19 Jun 2023 14:49  
 Operator : CG/JU  
 Sample : SSTDICC010  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1


Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDICC010

### Manual Integrations APPROVED

Reviewed By :Yogesh Patel 06/20/2023  
 Supervised By :mohammad ahmed 06/20/2023


Quant Time: Jun 20 03:09:01 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG061923.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jun 20 03:06:18 2023  
 Response via : Initial Calibration





#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.931 min Scan# 81  
Delta R.T. 0.002 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010



Tgt Ion:152 Resp: 35000  
Ion Ratio Lower Upper  
152 100  
150 156.4 126.6 189.8  
115 58.6 48.9 73.3




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#2  
1,4-Dioxane  
Concen: 10.529 ng  
RT: 3.406 min Scan# 54  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion: 88 Resp: 10881  
Ion Ratio Lower Upper  
88 100  
58 96.3 76.2 114.4  
43 34.2 26.3 39.5





#3  
Pyridine  
Concen: 10.107 ng  
RT: 3.800 min Scan# 11  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G





ClientSampleId :

SSTDICC010

Tgt Ion: 79 Resp: 3182  
Ion Ratio Lower Upper  
79 100  
52 85.5 71.4 107.0  
51 43.4 34.9 52.3




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#4  
n-Nitrosodimethylamine  
Concen: 10.327 ng  
RT: 3.718 min Scan# 107  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion: 42 Resp: 14852  
Ion Ratio Lower Upper  
42 100  
74 117.9 99.9 149.9  
44 7.9 6.0 9.0





#5  
2-Fluorophenol  
Concen: 20.656 ng  
RT: 5.498 min Scan# 411  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:112 Resp: 4689  
Ion Ratio Lower Upper  
112 100  
64 70.1 62.1 93.1  
63 35.5 31.5 47.3

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#6  
Aniline  
Concen: 10.232 ng  
RT: 7.249 min Scan# 708  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion: 93 Resp: 45383  
Ion Ratio Lower Upper  
93 100  
66 40.3 34.5 51.7  
65 18.3 17.4 26.2









#7  
Phenol-d6  
Concen: 20.413 ng  
RT: 7.085 min Scan# 681  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010

Tgt Ion: 99 Resp: 70714  
Ion Ratio Lower Upper  
99 100  
42 19.5 16.4 24.6  
71 31.0 25.8 38.6




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#8  
2-Chlorophenol  
Concen: 10.072 ng  
RT: 7.490 min Scan# 749  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49


Tgt Ion:128 Resp: 23525  
Ion Ratio Lower Upper  
128 100  
130 29.4 10.6 50.6  
64 63.7 39.1 79.1





#9  
Benzaldehyde  
Concen: 11.008 ng  
RT: 7.067 min Scan# 61  
Delta R.T. 0.002 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010



Tgt Ion: 77 Resp: 2397  
Ion Ratio Lower Upper  
77 100  
105 97.7 73.7 113.7  
106 89.8 69.0 109.0




Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#10  
Phenol  
Concen: 10.310 ng  
RT: 7.108 min Scan# 684  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion: 94 Resp: 35165  
Ion Ratio Lower Upper  
94 100  
65 27.4 12.2 52.2  
66 40.3 18.3 58.3











#11  
bis(2-Chloroethyl)ether  
Concen: 10.157 ng  
RT: 7.349 min Scan# 71  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010

Tgt Ion: 93 Resp: 2663  
Ion Ratio Lower Upper  
93 100  
63 86.9 71.2 111.2  
95 29.7 14.2 54.2




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#12  
1,3-Dichlorobenzene  
Concen: 10.504 ng  
RT: 7.819 min Scan# 805  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:146 Resp: 24967  
Ion Ratio Lower Upper  
146 100  
148 64.0 57.0 85.6  
75 34.0 27.8 41.8





#13

1,4-Dichlorobenzene

Concen: 10.250 ng

RT: 7.966 min Scan# 81

Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010

Tgt Ion:146 Resp: 2468





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 148 | 70.3  | 57.0  | 85.4  |
| 111 | 40.4  | 35.1  | 52.7  |

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023

Supervised By :mohammad ahmed 06/20/2023



#14

1,2-Dichlorobenzene

Concen: 10.614 ng

RT: 8.283 min Scan# 884




Delta R.T. 0.001 min


Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Tgt Ion:146 Resp: 24735

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 148 | 70.6  | 51.0  | 76.6  |
| 111 | 45.6  | 34.9  | 52.3  |





#15  
Benzyl Alcohol  
Concen: 9.819 ng  
RT: 8.160 min Scan# 80  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G





ClientSampleId :

SSTDICC010

Tgt Ion: 79 Resp: 26008  
Ion Ratio Lower Upper  
79 100  
108 79.4 60.5 90.7  
77 65.3 49.3 73.9




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 10.252 ng m  
RT: 8.454 min Scan# 913  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion: 45 Resp: 51687  
Ion Ratio Lower Upper  
45 100  
77 11.3 0.0 30.6  
79 7.4 0.0 28.3





#17  
2-Methylphenol  
Concen: 10.121 ng  
RT: 8.365 min Scan# 898  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:107 Resp: 25389  
Ion Ratio Lower Upper  
107 100  
108 108.8 93.2 139.8  
77 40.1 37.8 56.8  
79 41.9 36.8 55.2

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#18  
Hexachloroethane  
Concen: 10.433 ng  
RT: 9.012 min Scan# 1008  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:117 Resp: 8860  
Ion Ratio Lower Upper  
117 100  
119 92.5 72.3 108.5  
201 95.7 87.8 131.8






#19  
n-Nitroso-di-n-propylamine  
Concen: 10.085 ng  
RT: 8.730 min Scan# 961  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion: 70 Resp: 2435  
Ion Ratio Lower Upper  
70 100  
42 66.4 54.2 81.2  
101 8.9 7.4 11.0  
130 16.4 13.4 20.0


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#20  
3+4-Methylphenols  
Concen: 9.857 ng  
RT: 8.689 min Scan# 953  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:107 Resp: 34157  
Ion Ratio Lower Upper  
107 100  
108 83.4 67.7 107.7  
77 27.5 13.3 53.3  
79 25.5 7.5 47.5






#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.733 min Scan# 111  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:136 Resp: 15560  
Ion Ratio Lower Upper  
136 100  
137 11.6 8.5 12.7  
54 11.9 9.5 14.3  
68 6.7 5.7 8.5

Manual Integrations

APPROVED




Reviewed By :Yogesh Patel 06/20/2023


Supervised By :mohammad ahmed 06/20/2023



#22  
Acetophenone  
Concen: 10.560 ng  
RT: 8.747 min Scan# 963  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:105 Resp: 46077  
Ion Ratio Lower Upper  
105 100  
71 2.1 2.4 3.6#  
51 30.8 25.9 38.9  
120 19.5 16.7 25.1





#23

Nitrobenzene-d5

Concen: 19.105 ng

RT: 9.088 min Scan# 1021

Delta R.T. -0.004 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010

Tgt Ion: 82 Resp: 5792





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 82  | 100   |       |       |
| 128 | 38.1  | 30.7  | 46.1  |
| 54  | 65.6  | 49.6  | 74.4  |

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023

Supervised By :mohammad ahmed 06/20/2023



#24

Nitrobenzene

Concen: 9.875 ng

RT: 9.129 min Scan# 1028




Delta R.T. -0.005 min


Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Tgt Ion: 77 Resp: 31160

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 77  | 100   |       |       |
| 123 | 36.5  | 31.4  | 47.2  |
| 65  | 14.1  | 11.2  | 16.8  |









#25  
Isophorone  
Concen: 9.923 ng  
RT: 9.652 min Scan# 1117  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010

Tgt Ion: 82 Resp: 6873  
Ion Ratio Lower Upper  
82 100  
95 6.9 5.2 7.8  
138 14.9 11.1 16.7




Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#26  
2-Nitrophenol  
Concen: 8.840 ng  
RT: 9.840 min Scan# 1149  
Delta R.T. -0.005 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:139 Resp: 8874  
Ion Ratio Lower Upper  
139 100  
109 32.4 22.5 33.7  
65 68.9 45.5 68.3#







#27

2,4-Dimethylphenol

Concen: 9.917 ng

RT: 9.899 min Scan# 1159

Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010





Tgt Ion:122 Resp: 24648  
 Ion Ratio Lower Upper  
 122 100  
 107 126.0 98.9 148.3  
 121 60.7 46.4 69.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023

Supervised By :mohammad ahmed 06/20/2023



#28

bis(2-Chloroethoxy)methane

Concen: 10.298 ng




RT: 10.134 min Scan# 1199


Delta R.T. -0.004 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Tgt Ion: 93 Resp: 38413  
 Ion Ratio Lower Upper  
 93 100  
 95 30.2 25.6 38.4  
 123 9.3 8.2 12.2





#29

2,4-Dichlorophenol

Concen: 9.525 ng

RT: 10.375 min Scan# 1240

Delta R.T. 0.001 min

Lab File: BG057755.D


Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 162   | Resp: | 22480 |
| Ion Ratio | Lower | Upper |       |
| 162       | 100   |       |       |
| 164       | 64.9  | 48.2  | 88.2  |
| 98        | 42.3  | 20.7  | 60.7  |

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023

Supervised By :mohammad ahmed 06/20/2023



#30

1,2,4-Trichlorobenzene

Concen: 10.562 ng




RT: 10.592 min Scan# 1277


Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 180   | Resp: | 27787 |
| Ion Ratio | Lower | Upper |       |
| 180       | 100   |       |       |
| 182       | 96.2  | 77.5  | 116.3 |
| 145       | 27.4  | 22.6  | 34.0  |






#31  
Naphthalene  
Concen: 10.492 ng  
RT: 10.786 min Scan# 1170  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:128 Resp: 8689  
Ion Ratio Lower Upper  
128 100  
129 10.7 8.5 12.7  
127 13.1 10.7 16.1


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#32  
Benzoic acid  
Concen: 7.677 ng  
RT: 9.969 min Scan# 1171  
Delta R.T. -0.064 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:122 Resp: 14133  
Ion Ratio Lower Upper  
122 100  
105 124.2 110.1 150.1  
77 91.5 78.2 118.2





#33  
4-Chloroaniline  
Concen: 10.128 ng  
RT: 10.886 min Scan# 1327  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:127 Resp: 3886  
Ion Ratio Lower Upper  
127 100  
129 34.4 26.6 40.0  
65 39.0 29.8 44.8  
92 20.6 16.7 25.1

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#34  
Hexachlorobutadiene  
Concen: 10.055 ng  
RT: 11.068 min Scan# 1358  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:225 Resp: 16138  
Ion Ratio Lower Upper  
225 100  
223 67.7 51.4 77.0  
227 62.6 49.7 74.5





#35  
Caprolactam  
Concen: 9.609 ng  
RT: 11.632 min Scan# 1460  
Delta R.T. -0.034 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:113 Resp: 920  
Ion Ratio Lower Upper  
113 100  
55 247.6 255.7 295.7  
56 193.3 175.2 215.2

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#36  
4-Chloro-3-methylphenol  
Concen: 9.808 ng  
RT: 12.002 min Scan# 1517  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49


Tgt Ion:107 Resp: 30232  
Ion Ratio Lower Upper  
107 100  
144 22.2 19.5 29.3  
142 67.3 57.9 86.9





#37  
2-Methylnaphthalene  
Concen: 10.304 ng  
RT: 12.390 min Scan# 11  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010




Tgt Ion:142 Resp: 62198  
Ion Ratio Lower Upper  
142 100  
141 89.0 72.1 108.1  
115 34.8 29.7 44.5

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#38  
1-Methylnaphthalene  
Concen: 10.390 ng  
RT: 12.607 min Scan# 1620  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:142 Resp: 59206  
Ion Ratio Lower Upper  
142 100  
141 95.4 74.2 111.4  
116 3.8 3.1 4.7






#39  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.564 min Scan# 1953  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:164 Resp: 109910  
Ion Ratio Lower Upper  
164 100  
162 96.9 78.3 117.5  
160 45.4 35.4 53.0


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 10.347 ng  
RT: 12.754 min Scan# 1645  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Tgt Ion:216 Resp: 31529  
Ion Ratio Lower Upper  
216 100  
214 79.8 63.0 94.4  
179 20.2 16.2 24.2  
108 18.5 14.2 21.2





#41  
Hexachlorocyclopentadiene  
Concen: 9.442 ng  
RT: 12.737 min Scan# 1642  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010




Tgt Ion: 237 Resp: 15279  
Ion Ratio Lower Upper  
237 100  
235 64.7 43.5 83.5  
272 13.7 0.0 32.8

Manual Integrations  
APPROVED



Reviewed By : Yogesh Patel 06/20/2023  
Supervised By : mohammad ahmed 06/20/2023




#42  
2,4,6-Tribromophenol  
Concen: 19.492 ng  
RT: 16.051 min Scan# 2206  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion: 330 Resp: 29844  
Ion Ratio Lower Upper  
330 100  
332 95.3 76.6 115.0  
141 30.9 24.7 37.1







#43  
2,4,6-Trichlorophenol  
Concen: 9.669 ng  
RT: 12.995 min Scan# 1686  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:196 Resp: 2109  
Ion Ratio Lower Upper  
196 100  
198 94.7 75.5 113.3  
200 30.9 24.6 36.8

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#44  
2,4,5-Trichlorophenol  
Concen: 9.859 ng  
RT: 13.060 min Scan# 1697  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49


Tgt Ion:196 Resp: 25075  
Ion Ratio Lower Upper  
196 100  
198 99.4 78.9 118.3  
97 47.5 38.3 57.5  
132 24.6 21.7 32.5





#45  
2-Fluorobiphenyl  
Concen: 21.687 ng  
RT: 13.189 min Scan# 1719  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010




Tgt Ion:172 Resp: 15716  
Ion Ratio Lower Upper  
172 100  
171 33.9 28.1 42.1  
170 22.5 18.6 27.8

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#46  
1,1'-Biphenyl  
Concen: 10.574 ng  
RT: 13.395 min Scan# 1754  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:154 Resp: 79802  
Ion Ratio Lower Upper  
154 100  
153 39.0 19.1 59.1  
76 15.5 0.0 34.3






#47  
2-Chloronaphthalene  
Concen: 10.455 ng  
RT: 13.442 min Scan# 11  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:162 Resp: 6061  
Ion Ratio Lower Upper  
162 100  
127 39.4 32.2 48.2  
164 34.0 28.0 42.0

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#48  
2-Nitroaniline  
Concen: 8.318 ng  
RT: 13.636 min Scan# 1795  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion: 65 Resp: 17138  
Ion Ratio Lower Upper  
65 100  
92 63.7 51.2 76.8  
138 84.1 67.2 100.8






#49  
Acenaphthylene  
Concen: 10.536 ng  
RT: 14.282 min Scan# 1906  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:152 Resp: 10485  
Ion Ratio Lower Upper  
152 100  
151 19.2 17.0 25.4  
153 12.6 10.5 15.7


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#50  
Dimethylphthalate  
Concen: 10.676 ng  
RT: 14.018 min Scan# 1860  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:163 Resp: 90131  
Ion Ratio Lower Upper  
163 100  
194 3.2 2.8 4.2  
164 10.0 8.2 12.4





#51  
2,6-Dinitrotoluene  
Concen: 8.905 ng  
RT: 14.135 min Scan# 1880  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G





ClientSampleId :

SSTDICC010

Tgt Ion:165 Resp: 1384  
Ion Ratio Lower Upper  
165 100  
63 67.9 52.8 79.2  
89 62.3 50.7 76.1




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#52  
Acenaphthene  
Concen: 10.340 ng  
RT: 14.629 min Scan# 1964  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:154 Resp: 63703  
Ion Ratio Lower Upper  
154 100  
153 113.8 93.6 140.4  
152 54.4 44.7 67.1





#53  
3-Nitroaniline  
Concen: 9.450 ng  
RT: 14.458 min Scan# 1935  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G





ClientSampleId :

SSTDICC010

Tgt Ion:138 Resp: 16550  
Ion Ratio Lower Upper  
138 100  
108 12.6 9.1 13.7  
92 142.4 106.4 159.6




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#54  
2,4-Dinitrophenol  
Concen: 11.306 ng  
RT: 14.658 min Scan# 1969  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:184 Resp: 5240  
Ion Ratio Lower Upper  
184 100  
63 73.2 55.2 82.8  
154 63.9 47.4 71.0






#55  
Dibenzofuran  
Concen: 10.742 ng  
RT: 14.958 min Scan# 20  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:168 Resp: 10641  
Ion Ratio Lower Upper  
168 100  
139 33.3 28.1 42.1  
169 12.5 10.7 16.1


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#56  
4-Nitrophenol  
Concen: 9.203 ng  
RT: 14.758 min Scan# 1986  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:139 Resp: 12813  
Ion Ratio Lower Upper  
139 100  
109 80.8 57.3 97.3  
65 127.2 100.0 140.0






#57  
2,4-Dinitrotoluene  
Concen: 8.455 ng  
RT: 14.917 min Scan# 2014  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:165 Resp: 17829  
Ion Ratio Lower Upper  
165 100  
63 59.1 38.6 58.0  
89 96.0 63.7 95.5  
182 1.5 1.8 2.6#


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#58  
Fluorene  
Concen: 10.770 ng  
RT: 15.610 min Scan# 2131  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:166 Resp: 84454  
Ion Ratio Lower Upper  
166 100  
165 97.8 79.5 119.3  
167 14.1 10.6 16.0







#59  
2,3,4,6-Tetrachlorophenol  
Concen: 9.616 ng m  
RT: 15.181 min Scan# 2058  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:232 Resp: 2084  
Ion Ratio Lower Upper  
232 100  
131 39.1 31.9 47.9  
130 2.3 1.6 2.4  
166 24.5 23.7 35.5

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#60  
Diethylphthalate  
Concen: 10.590 ng  
RT: 15.381 min Scan# 2092  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:149 Resp: 91721  
Ion Ratio Lower Upper  
149 100  
177 22.0 17.0 25.6  
150 12.8 10.0 15.0





#61  
4-Chlorophenyl-phenylether  
Concen: 10.520 ng  
RT: 15.604 min Scan# 2130  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDIC010

Tgt Ion:204 Resp: 43860  
Ion Ratio Lower Upper  
204 100  
206 33.0 26.9 40.3  
141 52.6 44.3 66.5

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#62  
4-Nitroaniline  
Concen: 9.864 ng  
RT: 15.616 min Scan# 2132  
Delta R.T. -0.016 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:138 Resp: 17618  
Ion Ratio Lower Upper  
138 100  
92 52.9 35.9 75.9  
108 80.3 62.1 102.1





#63

Azobenzene

Concen: 10.573 ng

RT: 15.892 min Scan# 2180

Delta R.T. -0.004 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010





|             |             |
|-------------|-------------|
| Tgt Ion: 77 | Resp: 9367  |
| Ion Ratio   | Lower Upper |
| 77 100      |             |
| 182 21.0    | 4.1 44.1    |
| 105 22.1    | 3.3 43.3    |
| 51 30.5     | 13.4 53.4   |

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023

Supervised By :mohammad ahmed 06/20/2023



#64

Phenanthrene-d10

Concen: 20.000 ng




RT: 17.308 min Scan# 2420


Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

|             |              |
|-------------|--------------|
| Tgt Ion:188 | Resp: 271594 |
| Ion Ratio   | Lower Upper  |
| 188 100     |              |
| 94 8.8      | 6.6 9.8      |
| 80 10.4     | 7.6 11.4     |





#65  
4,6-Dinitro-2-methylphenol  
Concen: 11.316 ng  
RT: 15.680 min Scan# 2143  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:198 Resp: 842  
Ion Ratio Lower Upper  
198 100  
51 59.7 36.6 76.6  
105 41.6 24.3 64.3

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#66  
n-Nitrosodiphenylamine  
Concen: 10.372 ng  
RT: 15.815 min Scan# 2166  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:169 Resp: 74347  
Ion Ratio Lower Upper  
169 100  
168 66.7 53.4 80.2  
167 39.2 31.2 46.8





#67  
4-Bromophenyl-phenylether  
Concen: 10.328 ng  
RT: 16.497 min Scan# 21  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:248 Resp: 2974  
Ion Ratio Lower Upper  
248 100  
250 97.0 81.1 121.7  
141 67.2 54.6 81.8

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#68  
Hexachlorobenzene  
Concen: 10.211 ng  
RT: 16.609 min Scan# 2301  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:284 Resp: 29983  
Ion Ratio Lower Upper  
284 100  
142 25.8 24.2 36.4  
249 32.1 25.9 38.9





#69

Atrazine

Concen: 10.618 ng

RT: 16.761 min Scan# 21

Delta R.T. -0.004 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:200 Resp: 26128  
 Ion Ratio Lower Upper  
 200 100  
 173 26.3 5.6 45.6  
 215 47.4 24.6 64.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023  
 Supervised By :mohammad ahmed 06/20/2023



#70

Pentachlorophenol

Concen: 9.071 ng




RT: 16.949 min Scan# 2359


Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Tgt Ion:266 Resp: 19678  
 Ion Ratio Lower Upper  
 266 100  
 268 68.2 50.5 75.7  
 264 63.6 51.8 77.6





#71  
Phenanthrene  
Concen: 10.619 ng  
RT: 17.349 min Scan# 2427  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :





SSTDICC010

Tgt Ion:178 Resp: 139638  
Ion Ratio Lower Upper  
178 100  
176 20.4 16.3 24.5  
179 15.9 13.0 19.6

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#72  
Anthracene  
Concen: 10.470 ng  
RT: 17.437 min Scan# 2442  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49


Tgt Ion:178 Resp: 140594  
Ion Ratio Lower Upper  
178 100  
176 19.0 15.4 23.0  
179 15.5 13.3 19.9





#73  
 Carbazole  
 Concen: 10.671 ng  
 RT: 17.707 min Scan# 2488  
 Delta R.T. 0.001 min  
 Lab File: BG057755.D  
 Acq: 19 Jun 2023 14:49




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDICC010




Tgt Ion:167 Resp: 132117  
 Ion Ratio Lower Upper  
 167 100  
 166 22.5 18.4 27.6  
 139 11.8 9.9 14.9

Manual Integrations  
 APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
 Supervised By :mohammad ahmed 06/20/2023




#74  
 Di-n-butylphthalate  
 Concen: 10.261 ng  
 RT: 18.271 min Scan# 2584  
 Delta R.T. 0.001 min  
 Lab File: BG057755.D  
 Acq: 19 Jun 2023 14:49



Tgt Ion:149 Resp: 150419  
 Ion Ratio Lower Upper  
 149 100  
 150 10.0 7.5 11.3  
 104 5.5 4.1 6.1








#75  
Fluoranthene  
Concen: 10.651 ng  
RT: 19.358 min Scan# 21  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010




Tgt Ion:202 Resp: 16685  
Ion Ratio Lower Upper  
202 100  
101 11.2 0.0 31.3  
203 17.3 0.0 38.7

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.556 min Scan# 3143  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:240 Resp: 233787  
Ion Ratio Lower Upper  
240 100  
120 9.7 8.1 12.1  
236 26.0 20.6 30.8





#77

Benzidine

Concen: 10.540 ng

RT: 19.541 min Scan# 2800

Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010





|             |             |
|-------------|-------------|
| Tgt Ion:184 | Resp: 5776  |
| Ion Ratio   | Lower Upper |
| 184 100     |             |
| 185 13.9    | 12.0 18.0   |
| 183 12.4    | 9.4 14.0    |

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023

Supervised By :mohammad ahmed 06/20/2023



#78

Pyrene

Concen: 10.559 ng




RT: 19.717 min Scan# 2830


Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

|             |              |
|-------------|--------------|
| Tgt Ion:202 | Resp: 176093 |
| Ion Ratio   | Lower Upper  |
| 202 100     |              |
| 200 20.4    | 17.4 26.0    |
| 203 18.1    | 14.6 22.0    |






#79  
 Terphenyl-d14  
 Concen: 21.832 ng  
 RT: 19.922 min Scan# 2865  
 Delta R.T. 0.001 min  
 Lab File: BG057755.D  
 Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:244 Resp: 273720  
 Ion Ratio Lower Upper  
 244 100  
 212 6.9 6.1 9.1  
 122 8.9 7.6 11.4


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
 Supervised By :mohammad ahmed 06/20/2023




#80  
 Butylbenzylphthalate  
 Concen: 9.395 ng  
 RT: 20.610 min Scan# 2982  
 Delta R.T. 0.001 min  
 Lab File: BG057755.D  
 Acq: 19 Jun 2023 14:49



Tgt Ion:149 Resp: 59604  
 Ion Ratio Lower Upper  
 149 100  
 91 77.9 61.7 92.5  
 206 22.0 16.6 24.8









#81  
Benzo(a)anthracene  
Concen: 10.348 ng  
RT: 21.538 min Scan# 3140  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010

Tgt Ion:228 Resp: 16598  
Ion Ratio Lower Upper  
228 100  
226 27.3 22.3 33.5  
229 20.8 16.3 24.5




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#82  
3,3'-Dichlorobenzidine  
Concen: 10.169 ng  
RT: 21.456 min Scan# 3126  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49


Tgt Ion:252 Resp: 55479  
Ion Ratio Lower Upper  
252 100  
254 62.8 51.0 76.6  
126 10.0 7.0 10.6





#83  
Chrysene  
Concen: 10.179 ng  
RT: 21.603 min Scan# 3152  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010



Tgt Ion:228 Resp: 156690  
Ion Ratio Lower Upper  
228 100  
226 31.4 24.8 37.2  
229 20.5 15.6 23.4




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#84  
Bis(2-ethylhexyl)phthalate  
Concen: 10.226 ng  
RT: 21.456 min Scan# 3126  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49


Tgt Ion:149 Resp: 94188  
Ion Ratio Lower Upper  
149 100  
167 27.7 21.8 32.8  
279 3.7 2.9 4.3





#85  
Di-n-octyl phthalate  
Concen: 9.703 ng  
RT: 22.649 min Scan# 31  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010




Tgt Ion:149 Resp: 15676  
Ion Ratio Lower Upper  
149 100  
167 1.4 1.2 1.8  
43 12.9 9.6 14.4

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 24.711 min Scan# 3680  
Delta R.T. 0.007 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:264 Resp: 282198  
Ion Ratio Lower Upper  
264 100  
260 23.9 18.5 27.7  
265 21.1 17.4 26.2





#87

Indeno(1,2,3-cd)pyrene

Concen: 9.952 ng

RT: 28.283 min Scan# 41

Delta R.T. -0.016 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010

Tgt Ion:276 Resp: 18362

Ion Ratio Lower Upper

276 100

138 14.7 11.3 16.9





277 25.3 20.4 30.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/20/2023

Supervised By :mohammad ahmed 06/20/2023



#88

Benzo(b)fluoranthene

Concen: 10.200 ng

RT: 23.706 min Scan# 3509

Delta R.T. 0.001 min

Lab File: BG057755.D

Acq: 19 Jun 2023 14:49




Tgt Ion:252 Resp: 157734


Ion Ratio Lower Upper

252 100

253 21.8 17.1 25.7

125 8.8 7.4 11.0






#89  
Benzo(k)fluoranthene  
Concen: 10.337 ng  
RT: 23.771 min Scan# 3521  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Instrument :

BNA\_G

ClientSampleId :

SSTDICC010






Tgt Ion:252 Resp: 16281  
Ion Ratio Lower Upper  
252 100  
253 21.9 18.0 27.0  
125 9.4 7.6 11.4


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023




#90  
Benzo(a)pyrene  
Concen: 10.092 ng  
RT: 24.558 min Scan# 3654  
Delta R.T. 0.001 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49



Tgt Ion:252 Resp: 153263  
Ion Ratio Lower Upper  
252 100  
253 22.1 17.4 26.2  
125 10.9 8.0 12.0








#91  
Dibenzo(a,h)anthracene  
Concen: 9.993 ng  
RT: 28.360 min Scan# 41  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC010



Tgt Ion:278 Resp: 152980  
Ion Ratio Lower Upper  
278 100  
139 13.2 10.8 16.2  
279 23.8 19.0 28.4

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/20/2023  
Supervised By :mohammad ahmed 06/20/2023



#92  
Benzo(g,h,i)perylene  
Concen: 9.963 ng  
RT: 29.411 min Scan# 4480  
Delta R.T. -0.004 min  
Lab File: BG057755.D  
Acq: 19 Jun 2023 14:49

Tgt Ion:276 Resp: 152712  
Ion Ratio Lower Upper  
276 100  
277 23.5 19.4 29.0  
138 17.5 14.2 21.4

