

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062023\
 Data File : BG057767.D
 Acq On : 20 Jun 2023 8:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 20 09:13:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 03:40:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.590	0.595	-0.8	98	0.00
3	Pyridine	1.798	1.842	-2.4	94	0.00
4	n-Nitrosodimethylamine	0.822	0.799	2.8	94	0.00
5 S	2-Fluorophenol	1.297	1.329	-2.5	97	0.00
6	Aniline	2.534	2.561	-1.1	94	0.00
7 S	Phenol-d6	1.979	2.032	-2.7	96	0.00
8	2-Chlorophenol	1.334	1.348	-1.0	96	0.00
9	Benzaldehyde	1.244	1.134	8.8	97	0.00
10 C	Phenol	1.949	1.999	-2.6	96	0.00
11	bis(2-Chloroethyl)ether	1.498	1.489	0.6	94	0.00
12	1,3-Dichlorobenzene	1.358	1.382	-1.8	96	0.00
13 C	1,4-Dichlorobenzene	1.376	1.385	-0.7	97	0.00
14	1,2-Dichlorobenzene	1.331	1.340	-0.7	95	0.00
15	Benzyl Alcohol	1.513	1.505	0.5	91	0.00
16	2,2'-oxybis(1-Chloropropane	2.878	2.856	0.8	94	0.00
17	2-Methylphenol	1.433	1.448	-1.0	95	0.00
18	Hexachloroethane	0.485	0.488	-0.6	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.380	1.368	0.9	90	0.00
20	3+4-Methylphenols	1.980	2.023	-2.2	93	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
22	Acetophenone	0.561	0.569	-1.4	94	0.00
23 S	Nitrobenzene-d5	0.390	0.403	-3.3	93	0.00
24	Nitrobenzene	0.406	0.413	-1.7	93	0.00
25	Isophorone	0.890	0.895	-0.6	91	0.00
26 C	2-Nitrophenol	0.129	0.150	-16.3	96	0.00
27	2,4-Dimethylphenol	0.319	0.321	-0.6	91	0.00
28	bis(2-Chloroethoxy)methane	0.479	0.485	-1.3	93	0.00
29 C	2,4-Dichlorophenol	0.303	0.309	-2.0	91	0.00
30	1,2,4-Trichlorobenzene	0.338	0.337	0.3	94	0.00
31	Naphthalene	1.064	1.071	-0.7	94	0.00
32	Benzoic acid	0.235	0.237	-0.9	93	-0.02
33	4-Chloroaniline	0.493	0.510	-3.4	95	0.00
34 C	Hexachlorobutadiene	0.206	0.206	0.0	95	0.00
35	Caprolactam	0.123	0.125	-1.6	90	0.00
36 C	4-Chloro-3-methylphenol	0.396	0.419	-5.8	95	0.00
37	2-Methylnaphthalene	0.776	0.792	-2.1	94	0.00
38	1-Methylnaphthalene	0.732	0.745	-1.8	93	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	0.554	0.552	0.4	92	0.00
41 P	Hexachlorocyclopentadiene	0.294	0.289	1.7	89	0.00
42 S	2,4,6-Tribromophenol	0.279	0.291	-4.3	94	0.00
43 C	2,4,6-Trichlorophenol	0.397	0.410	-3.3	93	0.00
44	2,4,5-Trichlorophenol	0.463	0.486	-5.0	96	0.00
45 S	2-Fluorobiphenyl	1.319	1.316	0.2	94	0.00
46	1,1'-Biphenyl	1.373	1.361	0.9	92	0.00
47	2-Chloronaphthalene	1.055	1.056	-0.1	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062023\
 Data File : BG057767.D
 Acq On : 20 Jun 2023 8:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 20 09:13:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 03:40:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.375	0.403	-7.5	94	0.00
49	Acenaphthylene	1.811	1.819	-0.4	93	0.00
50	Dimethylphthalate	1.536	1.556	-1.3	94	0.00
51	2,6-Dinitrotoluene	0.283	0.308	-8.8	93	0.00
52 C	Acenaphthene	1.121	1.128	-0.6	93	0.00
53	3-Nitroaniline	0.319	0.348	-9.1	95	0.00
54 P	2,4-Dinitrophenol	0.131	0.133	-1.5	92	0.00
55	Dibenzofuran	1.803	1.817	-0.8	93	0.00
56 P	4-Nitrophenol	0.253	0.275	-8.7	94	0.00
57	2,4-Dinitrotoluene	0.384	0.429	-11.7	96	0.00
58	Fluorene	1.427	1.432	-0.4	93	0.00
59	2,3,4,6-Tetrachlorophenol	0.394	0.410	-4.1	92	0.00
60	Diethylphthalate	1.576	1.597	-1.3	93	0.00
61	4-Chlorophenyl-phenylether	0.759	0.768	-1.2	94	0.00
62	4-Nitroaniline	0.325	0.342	-5.2	92	0.00
63	Azobenzene	1.612	1.613	-0.1	93	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	0.084	0.086	-2.4	93	0.00
66 c	n-Nitrosodiphenylamine	0.528	0.536	-1.5	94	0.00
67	4-Bromophenyl-phenylether	0.212	0.213	-0.5	92	0.00
68	Hexachlorobenzene	0.216	0.216	0.0	92	0.00
69	Atrazine	0.181	0.183	-1.1	93	0.00
70 C	Pentachlorophenol	0.160	0.169	-5.6	94	0.00
71	Phenanthrene	0.968	0.974	-0.6	94	0.00
72	Anthracene	0.989	0.988	0.1	92	0.00
73	Carbazole	0.912	0.914	-0.2	93	0.00
74	Di-n-butylphthalate	1.080	1.111	-2.9	94	0.00
75 C	Fluoranthene	1.154	1.153	0.1	93	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
77	Benzidine	0.469	0.584	-24.5	108	0.00
78	Pyrene	1.427	1.437	-0.7	93	0.00
79 S	Terphenyl-d14	1.073	1.093	-1.9	96	0.00
80	Butylbenzylphthalate	0.543	0.572	-5.3	94	0.00
81	Benzo(a)anthracene	1.372	1.398	-1.9	94	0.00
82	3,3'-Dichlorobenzidine	0.467	0.493	-5.6	95	0.00
83	Chrysene	1.317	1.341	-1.8	95	0.00
84	Bis(2-ethylhexyl)phthalate	0.788	0.824	-4.6	94	0.00
85 c	Di-n-octyl phthalate	1.382	1.466	-6.1	94	0.00
86 I	Perylene-d12	1.000	1.000	0.0	95	0.00
87	Indeno(1,2,3-cd)pyrene	1.308	1.327	-1.5	94	0.00
88	Benzo(b)fluoranthene	1.096	1.130	-3.1	96	0.00
89	Benzo(k)fluoranthene	1.116	1.117	-0.1	94	0.00
90 C	Benzo(a)pyrene	1.076	1.103	-2.5	94	0.00
91	Dibenzo(a,h)anthracene	1.085	1.117	-2.9	95	0.00
92	Benzo(g,h,i)perylene	1.086	1.109	-2.1	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062023\
Data File : BG057767.D
Acq On : 20 Jun 2023 8:35
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 20 09:13:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 03:40:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0