

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062023\
 Data File : BG057767.D
 Acq On : 20 Jun 2023 8:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 20 09:13:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 03:40:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	40.326	-0.8	98	0.00
3	Pyridine	40.000	40.969	-2.4	94	0.00
4	n-Nitrosodimethylamine	40.000	38.883	2.8	94	0.00
5 S	2-Fluorophenol	80.000	81.975	-2.5	97	0.00
6	Aniline	40.000	40.422	-1.1	94	0.00
7 S	Phenol-d6	80.000	82.134	-2.7	96	0.00
8	2-Chlorophenol	40.000	40.409	-1.0	96	0.00
9	Benzaldehyde	40.000	36.453	8.9	97	0.00
10 C	Phenol	40.000	41.034	-2.6	96	0.00
11	bis(2-Chloroethyl)ether	40.000	39.766	0.6	94	0.00
12	1,3-Dichlorobenzene	40.000	40.698	-1.7	96	0.00
13 C	1,4-Dichlorobenzene	40.000	40.259	-0.6	97	0.00
14	1,2-Dichlorobenzene	40.000	40.259	-0.6	95	0.00
15	Benzyl Alcohol	40.000	39.784	0.5	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.689	0.8	94	0.00
17	2-Methylphenol	40.000	40.403	-1.0	95	0.00
18	Hexachloroethane	40.000	40.210	-0.5	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.645	0.9	90	0.00
20	3+4-Methylphenols	40.000	40.880	-2.2	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	92	0.00
22	Acetophenone	40.000	40.567	-1.4	94	0.00
23 S	Nitrobenzene-d5	80.000	82.777	-3.5	93	0.00
24	Nitrobenzene	40.000	40.694	-1.7	93	0.00
25	Isophorone	40.000	40.215	-0.5	91	0.00
26 C	2-Nitrophenol	40.000	46.488	-16.2	96	0.00
27	2,4-Dimethylphenol	40.000	40.147	-0.4	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.456	-1.1	93	0.00
29 C	2,4-Dichlorophenol	40.000	40.770	-1.9	91	0.00
30	1,2,4-Trichlorobenzene	40.000	39.871	0.3	94	0.00
31	Naphthalene	40.000	40.238	-0.6	94	0.00
32	Benzoic acid	40.000	40.266	-0.7	93	-0.02
33	4-Chloroaniline	40.000	41.374	-3.4	95	0.00
34 C	Hexachlorobutadiene	40.000	39.916	0.2	95	0.00
35	Caprolactam	40.000	40.616	-1.5	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.305	-5.8	95	0.00
37	2-Methylnaphthalene	40.000	40.815	-2.0	94	0.00
38	1-Methylnaphthalene	40.000	40.679	-1.7	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.851	0.4	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.271	1.8	89	0.00
42 S	2,4,6-Tribromophenol	80.000	83.480	-4.4	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.351	-3.4	93	0.00
44	2,4,5-Trichlorophenol	40.000	41.986	-5.0	96	0.00
45 S	2-Fluorobiphenyl	80.000	79.860	0.2	94	0.00
46	1,1'-Biphenyl	40.000	39.635	0.9	92	0.00
47	2-Chloronaphthalene	40.000	40.020	-0.1	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062023\
 Data File : BG057767.D
 Acq On : 20 Jun 2023 8:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 20 09:13:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 03:40:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.049	-7.6	94	0.00
49	Acenaphthylene	40.000	40.178	-0.4	93	0.00
50	Dimethylphthalate	40.000	40.507	-1.3	94	0.00
51	2,6-Dinitrotoluene	40.000	43.524	-8.8	93	0.00
52 C	Acenaphthene	40.000	40.258	-0.6	93	0.00
53	3-Nitroaniline	40.000	43.718	-9.3	95	0.00
54 P	2,4-Dinitrophenol	40.000	38.419	4.0	92	0.00
55	Dibenzofuran	40.000	40.321	-0.8	93	0.00
56 P	4-Nitrophenol	40.000	43.449	-8.6	94	0.00
57	2,4-Dinitrotoluene	40.000	44.747	-11.9	96	0.00
58	Fluorene	40.000	40.131	-0.3	93	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.571	-3.9	92	0.00
60	Diethylphthalate	40.000	40.532	-1.3	93	0.00
61	4-Chlorophenyl-phenylether	40.000	40.515	-1.3	94	0.00
62	4-Nitroaniline	40.000	42.063	-5.2	92	0.00
63	Azobenzene	40.000	40.026	-0.1	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.490	3.8	93	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.649	-1.6	94	0.00
67	4-Bromophenyl-phenylether	40.000	40.095	-0.2	92	0.00
68	Hexachlorobenzene	40.000	39.903	0.2	92	0.00
69	Atrazine	40.000	40.483	-1.2	93	0.00
70 C	Pentachlorophenol	40.000	42.254	-5.6	94	0.00
71	Phenanthrene	40.000	40.239	-0.6	94	0.00
72	Anthracene	40.000	39.973	0.1	92	0.00
73	Carbazole	40.000	40.080	-0.2	93	0.00
74	Di-n-butylphthalate	40.000	41.151	-2.9	94	0.00
75 C	Fluoranthene	40.000	39.978	0.1	93	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	49.838	-24.6	108	0.00
78	Pyrene	40.000	40.294	-0.7	93	0.00
79 S	Terphenyl-d14	80.000	81.517	-1.9	96	0.00
80	Butylbenzylphthalate	40.000	42.126	-5.3	94	0.00
81	Benzo(a)anthracene	40.000	40.739	-1.8	94	0.00
82	3,3'-Dichlorobenzidine	40.000	42.274	-5.7	95	0.00
83	Chrysene	40.000	40.725	-1.8	95	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.836	-4.6	94	0.00
85 c	Di-n-octyl phthalate	40.000	42.429	-6.1	94	0.00
86 I	Perylene-d12	20.000	20.000	0.0	95	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.579	-1.4	94	0.00
88	Benzo(b)fluoranthene	40.000	41.247	-3.1	96	0.00
89	Benzo(k)fluoranthene	40.000	40.027	-0.1	94	0.00
90 C	Benzo(a)pyrene	40.000	40.985	-2.5	94	0.00
91	Dibenzo(a,h)anthracene	40.000	41.176	-2.9	95	0.00
92	Benzo(g,h,i)perylene	40.000	40.830	-2.1	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062023\
Data File : BG057767.D
Acq On : 20 Jun 2023 8:35
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 20 09:13:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 03:40:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0