

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG062118\  
 Data File : BG035329.D  
 Acq On : 22 Jun 2018 11:29  
 Operator : JU/SJ  
 Sample : J3622-11  
 Misc :  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SP-B

Quant Time: Jun 22 12:57:17 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG060718.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jun 08 17:16:28 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)      |
|-----------------------------|-------|------|----------|--------|-------|---------------|
| 1) 1,4-Dichlorobenzene-d4   | 8.06  | 152  | 49841    | 20.00  | ng    | -0.02         |
| 21) Naphthalene-d8          | 10.86 | 136  | 173752   | 20.00  | ng    | -0.02         |
| 38) Acenaphthene-d10        | 14.68 | 164  | 114446   | 20.00  | ng    | -0.02         |
| 63) Phenanthrene-d10        | 17.42 | 188  | 327596   | 20.00  | ng    | -0.02         |
| 75) Chrysene-d12            | 21.68 | 240  | 365144   | 20.00  | ng    | -0.04         |
| 86) Perylene-d12            | 24.90 | 264  | 347122   | 20.00  | ng    | -0.05         |
| System Monitoring Compounds |       |      |          |        |       |               |
| 5) 2-Fluorophenol           | 5.63  | 112  | 368930   | 124.92 | ng    | -0.01         |
| 7) Phenol-d6                | 7.21  | 99   | 467025   | 107.14 | ng    | -0.01         |
| 23) Nitrobenzene-d5         | 9.21  | 82   | 388353   | 106.24 | ng    | -0.02         |
| 41) 2,4,6-Tribromophenol    | 16.16 | 330  | 266431   | 181.86 | ng    | -0.02         |
| 44) 2-Fluorobiphenyl        | 13.30 | 172  | 879225   | 99.87  | ng    | -0.02         |
| 78) Terphenyl-d14           | 20.02 | 244  | 1623074  | 93.11  | ng    | -0.02         |
| Target Compounds            |       |      |          |        |       |               |
| 31) Naphthalene             | 10.91 | 128  | 29439    | 3.261  | ng    | Qvalue<br>100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG062118\  
Data File : BG035329.D  
Acq On : 22 Jun 2018 11:29  
Operator : JU/SJ  
Sample : J3622-11  
Misc :  
ALS Vial : 33 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
SP-B

Quant Time: Jun 22 12:57:17 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG060718.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Jun 08 17:16:28 2018  
Response via : Initial Calibration

