


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG062119\
 Data File : BG041395.D
 Acq On : 22 Jun 2019 5:25
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 17 14:58:02 2019



AutoFind: Scans 2509, 2510, 2511; Background Corrected with Scan 2502

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.5	20744	PASS
68	69	0.00	2	0.8	190	PASS
69	198	0.00	100	33.5	22759	PASS
70	69	0.00	2	0.5	121	PASS
127	198	10	80	39.6	26901	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	67965	PASS
199	198	5	9	6.9	4659	PASS
275	198	10	60	30.6	20826	PASS
365	198	1	100	3.4	2324	PASS
441	198	0.01	100	9.6	6522	PASS
442	198	50	100	59.6	40488	PASS
443	442	15	24	19.0	7689	PASS