

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062122\
 Data File : BGCCC985.D
 Acq On : 21 Jun 2022 20:00
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Jun 21 23:08:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 15:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.078	152	20810	20.000	ng/u1	0.00
20) Naphthalene-d8	10.887	136	83130	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.706	164	65036	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.444	188	174358	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.721	240	198064	20.000	ng/u1	# 0.00
88) Perylene-d12	24.976	264	202786	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.507	96	3716	8.496	ng/uL	0.00
4) Pyridine-d5	3.918	84	23122	19.437	ng/u1	0.00
7) Phenol-d5	7.232	99	30428	20.410	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.396	67	18466	20.533	ng/u1	0.00
11) 2-Chlorophenol-d4	7.608	132	25135	21.142	ng/u1	0.00
15) 4-Methylphenol-d8	8.777	113	27049	21.279	ng/u1	0.00
21) Nitrobenzene-d5	9.236	128	12763	20.684	ng/u1	0.00
24) 2-Nitrophenol-d4	9.964	143	14497	22.112	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.505	165	31399	21.647	ng/u1	0.00
31) 4-Chloroaniline-d4	11.016	131	37645	20.543	ng/u1	0.00
46) Dimethylphthalate-d6	14.100	166	103259	20.182	ng/u1	0.00
49) Acenaphthylene-d8	14.394	160	112387	20.201	ng/u1	0.00
54) 4-Nitrophenol-d4	14.882	143	14245	20.334	ng/u1	0.00
60) Fluorene-d10	15.693	176	93894	20.431	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	18932	19.905	ng/u1	0.00
73) Anthracene-d10	17.543	188	161189	20.470	ng/u1	0.00
81) Pyrene-d10	19.829	212	210686	20.402	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.753	264	208059	20.326	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.542	88	4421	9.263	ng/uL#	81
5) Pyridine	3.936	79	25573	20.729	ng/u1	95
6) Benzaldehyde	7.214	77	18948	24.958	ng/u1	90
8) Phenol	7.261	94	32405	20.831	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.490	93	24500	20.792	ng/u1	87
12) 2-Chlorophenol	7.643	128	25944	21.374	ng/u1	99
13) 2-Methylphenol	8.513	108	25869	21.423	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.607	45	36911	21.089	ng/u1	97
16) Acetophenone	8.901	105	40987	20.849	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.883	70	21289	21.243	ng/u1	98
18) 4-Methylphenol	8.842	108	27159	21.024	ng/u1	95
19) Hexachloroethane	9.171	117	10606	20.620	ng/u1#	93
22) Nitrobenzene	9.283	77	30421	20.698	ng/u1	93
23) Isophorone	9.805	82	59465	20.868	ng/u1	97
25) 2-Nitrophenol	9.993	139	15219	22.138	ng/u1	97
26) 2,4-Dimethylphenol	10.046	107	31778	21.169	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.287	93	33667	20.987	ng/u1	99
29) 2,4-Dichlorophenol	10.534	162	30486	21.017	ng/u1	95
30) Naphthalene	10.939	128	84378	20.461	ng/u1	99
32) 4-Chloroaniline	11.039	127	37882	20.847	ng/u1	99
33) Hexachlorobutadiene	11.227	225	28644	20.217	ng/u1	95
34) Caprolactam	11.785	113	8557	19.768	ng/u1	78
35) 4-Chloro-3-methylphenol	12.156	107	30216	21.661	ng/u1	96
36) 2-Methylnaphthalene	12.538	142	60629	19.875	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062122\
 Data File : BGCCC985.D
 Acq On : 21 Jun 2022 20:00
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :


Quant Time: Jun 21 23:08:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 15:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.755	142	62205	20.447	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.902	216	53710	20.318	ng/ul	96
40) Hexachlorocyclopentadiene	12.890	237	24938	16.821	ng/ul#	91
41) 2,4,6-Trichlorophenol	13.137	196	29539	21.530	ng/ul	93
42) 2,4,5-Trichlorophenol	13.213	196	33561	21.637	ng/ul	98
43) 1,1'-Biphenyl	13.536	154	91749	20.354	ng/ul	98
44) 2-Chloronaphthalene	13.583	162	74360	19.865	ng/ul	97
45) 2-Nitroaniline	13.777	65	20841	22.131	ng/ul	96
47) Dimethylphthalate	14.147	163	98396	20.009	ng/ul	99
48) 2,6-Dinitrotoluene	14.265	165	20788	21.019	ng/ul	99
50) Acenaphthylene	14.424	152	117476	20.565	ng/ul	100
51) 3-Nitroaniline	14.594	138	17487	21.781	ng/ul#	92
52) Acenaphthene	14.764	153	77783	20.390	ng/ul	99
53) 2,4-Dinitrophenol	14.805	184	7674	16.749	ng/ul#	66
55) 4-Nitrophenol	14.899	109	13984	20.363	ng/ul	93
56) Dibenzofuran	15.099	168	120470	20.640	ng/ul	100
57) 2,4-Dinitrotoluene	15.052	165	31502	21.958	ng/ul#	93
58) 2,3,4,6-Tetrachlorophenol	15.328	232	30528	22.650	ng/ul#	88
59) Diethylphthalate	15.511	149	99292	20.593	ng/ul	98
61) Fluorene	15.746	166	98591	20.740	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.740	204	62102	20.546	ng/ul	95
63) 4-Nitroaniline	15.757	138	16544	21.430	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.816	198	18389	20.141	ng/ul	97
67) N-Nitrosodiphenylamine	15.945	169	89278	20.944	ng/ul	97
68) 4-Bromophenyl-phenylether	16.633	248	44307	20.458	ng/ul	96
69) Hexachlorobenzene	16.756	284	51650	20.348	ng/ul	95
70) Atrazine	16.891	200	42791	20.438	ng/ul	96
71) Pentachlorophenol	17.097	266	22155	20.592	ng/ul	93
72) Phenanthrene	17.485	178	179923	20.829	ng/ul	97
74) Anthracene	17.579	178	179341	20.600	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.507	216	56786	19.122	ng/ul	97
76) Pentachlorobenzene	15.023	250	54992	20.134	ng/ul	99
77) Carbazole	17.843	167	150395	20.953	ng/ul	99
78) Di-n-butylphthalate	18.401	149	171232	20.954	ng/ul	99
80) Fluoranthene	19.494	202	237921	20.638	ng/ul#	93
82) Pyrene	19.858	202	233276	20.353	ng/ul#	94
83) Butylbenzylphthalate	20.740	149	74213	21.485	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.609	252	91942	22.431	ng/ul	96
85) Benzo(a)anthracene	21.697	228	246459	20.188	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.609	149	105529	20.908	ng/ul#	97
87) Chrysene	21.768	228	237327	20.263	ng/ul	99
89) Di-n-octyl phthalate	22.831	149	185263	21.272	ng/ul	100
90) Benzo(b)fluoranthene	23.942	252	264237	21.102	ng/ul#	97
91) Benzo(k)fluoranthene	24.006	252	243424	20.303	ng/ul#	97
93) Benzo(a)pyrene	24.823	252	252274	20.915	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	28.701	276	301011	20.742	ng/ul#	96
95) Dibenzo(a,h)anthracene	28.766	278	254915	20.873	ng/ul#	96
96) Benzo(g,h,i)perylene	29.858	276	251342	20.513	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 21 23:08:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 20 15:29:13 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.263 ng/uL
RT: 3.542 min Scan# 20
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 88 Resp: 4421
Ion Ratio Lower Upper
88 100
43 32.5 48.1 72.1#
58 70.5 59.8 89.6



#5
Pyridine
Concen: 20.729 ng/uL
RT: 3.936 min Scan# 93
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion: 79 Resp: 25573
Ion Ratio Lower Upper
79 100
52 88.0 75.4 113.2
51 43.4 35.2 52.8









#6
Benzaldehyde
Concen: 24.958 ng/ul
RT: 7.214 min Scan# 611
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 77 Resp: 18948
Ion Ratio Lower Upper
77 100
105 114.5 79.7 119.5
106 92.2 71.0 106.4



#8
Phenol
Concen: 20.831 ng/ul
RT: 7.261 min Scan# 659
Delta R.T. 0.007 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion: 94 Resp: 32405
Ion Ratio Lower Upper
94 100
65 30.4 24.1 36.1
66 45.1 32.5 48.7








#10
 Bis(2-Chloroethyl)ether
 Concen: 20.792 ng/ul
 RT: 7.490 min Scan# 698
 Delta R.T. -0.005 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00


Instrument :
 BNA_G
 ClientSampleId :





Tgt Ion: 93	Resp:	24500
Ion	Ratio	Lower Upper
93	100	
63	76.2	72.3 108.5
95	31.3	29.4 44.0




#12
 2-Chlorophenol
 Concen: 21.374 ng/ul
 RT: 7.643 min Scan# 724
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00



Tgt Ion: 128	Resp:	25944
Ion	Ratio	Lower Upper
128	100	
64	50.4	41.4 62.2
130	33.9	27.0 40.6





#13

2-Methylphenol

Concen: 21.423 ng/ul

RT: 8.513 min Scan# 81

Delta R.T. 0.001 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00

Instrument :

BNA_G





ClientSampleId :

Tgt Ion:108 Resp: 25869

Ion Ratio Lower Upper

108 100

107 88.5 75.8 113.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.089 ng/ul

RT: 8.607 min Scan# 888

Delta R.T. 0.001 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00




Tgt Ion: 45 Resp: 36911


Ion Ratio Lower Upper

45 100

77 11.2 10.2 15.4

79 10.5 8.1 12.1









#16
Acetophenone
Concen: 20.849 ng/ul
RT: 8.901 min Scan# 91
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:105 Resp: 40987
Ion Ratio Lower Upper
105 100
77 77.2 61.0 91.6
51 27.7 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 21.243 ng/ul
RT: 8.883 min Scan# 935
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion: 70 Resp: 21289
Ion Ratio Lower Upper
70 100
42 61.9 50.2 75.4
101 10.2 7.1 10.7
130 22.1 15.5 23.3









#18
4-Methylphenol
Concen: 21.024 ng/ul
RT: 8.842 min Scan# 91
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:108 Resp: 27159
Ion Ratio Lower Upper
108 100
107 107.7 90.7 136.1



#19
Hexachloroethane
Concen: 20.620 ng/ul
RT: 9.171 min Scan# 984
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:117 Resp: 10606
Ion Ratio Lower Upper
117 100
201 147.9 117.4 176.0
199 96.2 64.0 96.0#









#22
Nitrobenzene
Concen: 20.698 ng/ul
RT: 9.283 min Scan# 1003
Delta R.T. 0.007 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 77 Resp: 30421
Ion Ratio Lower Upper
77 100
123 43.6 31.2 46.8
65 14.5 10.5 15.7



#23
Isophorone
Concen: 20.868 ng/ul
RT: 9.805 min Scan# 1092
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion: 82 Resp: 59465
Ion Ratio Lower Upper
82 100
95 8.8 6.1 9.1
138 18.3 13.6 20.4









#25
2-Nitrophenol
Concen: 22.138 ng/ul
RT: 9.993 min Scan# 1124
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:139 Resp: 15219
Ion Ratio Lower Upper
139 100
65 57.3 47.7 71.5
109 35.2 27.1 40.7



#26
2,4-Dimethylphenol
Concen: 21.169 ng/ul
RT: 10.046 min Scan# 1133
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:107 Resp: 31778
Ion Ratio Lower Upper
107 100
121 49.1 39.9 59.9
122 81.6 66.7 100.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.987 ng/ul
 RT: 10.287 min Scan# 1174
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion: 93 Resp: 33667
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.5 39.7
 123 11.9 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 21.017 ng/ul
 RT: 10.534 min Scan# 1216
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Tgt Ion: 162 Resp: 30486
 Ion Ratio Lower Upper
 162 100
 164 67.5 50.9 76.3
 98 33.1 28.5 42.7









#30
Naphthalene
Concen: 20.461 ng/ul
RT: 10.939 min Scan# 1302
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:128 Resp: 84378
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.6
127 13.5 11.1 16.7



#32
4-Chloroaniline
Concen: 20.847 ng/ul
RT: 11.039 min Scan# 1302
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion:127 Resp: 37882
Ion Ratio Lower Upper
127 100
129 32.5 25.4 38.2








#33
Hexachlorobutadiene
Concen: 20.217 ng/ul
RT: 11.227 min Scan# 11
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:225 Resp: 28644
Ion Ratio Lower Upper
225 100
223 67.7 50.2 75.2
227 61.6 51.8 77.6



#34
Caprolactam
Concen: 19.768 ng/ul
RT: 11.785 min Scan# 1429
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion:113 Resp: 8557
Ion Ratio Lower Upper
113 100
55 168.4 167.8 251.6
56 138.2 127.6 191.4








#35
 4-Chloro-3-methylphenol
 Concen: 21.661 ng/ul
 RT: 12.156 min Scan# 1492
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Instrument :
 BNA_G
 ClientSampleId :







Tgt Ion:107 Resp: 30216
 Ion Ratio Lower Upper
 107 100
 144 27.2 21.3 31.9
 142 76.5 65.0 97.4



#36
 2-Methylnaphthalene
 Concen: 19.875 ng/ul
 RT: 12.538 min Scan# 1557
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Tgt Ion:142 Resp: 60629
 Ion Ratio Lower Upper
 142 100
 141 92.8 76.3 114.5





#37

1-Methylnaphthalene

Concen: 20.447 ng/ul

RT: 12.755 min Scan# 1

Delta R.T. 0.001 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:142 Resp: 62205

Ion Ratio Lower Upper

142 100

141 92.7 73.4 110.2

116 4.1 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.318 ng/ul

RT: 12.902 min Scan# 1619

Delta R.T. -0.005 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00

Tgt Ion:216 Resp: 53710




Ion Ratio Lower Upper


216 100

214 76.1 58.1 87.1

179 17.1 14.8 22.2

108 11.4 10.4 15.6









#40
Hexachlorocyclopentadiene
Concen: 16.821 ng/ul
RT: 12.890 min Scan# 1617
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:237 Resp: 24938
Ion Ratio Lower Upper
237 100
235 65.5 47.0 70.6
272 13.5 8.6 12.8#



#41
2,4,6-Trichlorophenol
Concen: 21.530 ng/ul
RT: 13.137 min Scan# 1659
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:196 Resp: 29539
Ion Ratio Lower Upper
196 100
198 93.8 81.9 122.9
200 29.8 25.7 38.5









#42
2,4,5-Trichlorophenol
Concen: 21.637 ng/ul
RT: 13.213 min Scan# 1
Delta R.T. 0.007 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:196 Resp: 33561
Ion Ratio Lower Upper
196 100
198 93.5 76.3 114.5
200 30.8 25.5 38.3



#43
1,1'-Biphenyl
Concen: 20.354 ng/ul
RT: 13.536 min Scan# 1727
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion:154 Resp: 91749
Ion Ratio Lower Upper
154 100
153 41.8 32.7 49.1
76 9.4 9.0 13.6








#44
2-Chloronaphthalene
Concen: 19.865 ng/ul
RT: 13.583 min Scan# 1735
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:162 Resp: 74360
Ion Ratio Lower Upper
162 100
164 33.1 25.7 38.5
127 33.4 28.3 42.5



#45
2-Nitroaniline
Concen: 22.131 ng/ul
RT: 13.777 min Scan# 1768
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion: 65 Resp: 20841
Ion Ratio Lower Upper
65 100
92 59.1 52.2 78.4
138 97.9 79.0 118.4









#47
Dimethylphthalate
Concen: 20.009 ng/ul
RT: 14.147 min Scan# 1831
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:163 Resp: 98396
Ion Ratio Lower Upper
163 100
194 4.8 3.7 5.5
164 9.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 21.019 ng/ul
RT: 14.265 min Scan# 1851
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion:165 Resp: 20788
Ion Ratio Lower Upper
165 100
89 55.7 44.7 67.1
121 18.9 16.9 25.3








#50
Acenaphthylene
Concen: 20.565 ng/ul
RT: 14.424 min Scan# 117476
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.5	16.2	24.2
153	13.1	10.5	15.7



#51
3-Nitroaniline
Concen: 21.781 ng/ul
RT: 14.594 min Scan# 1907
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion	Ratio	Lower	Upper
138	100		
108	9.0	17.8	26.6
92	112.9	94.2	141.4








#52
Acenaphthene
Concen: 20.390 ng/ul
RT: 14.764 min Scan# 1936
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Instrument :
BNA_G
ClientSampleId :





Tgt Ion:153 Resp: 77783
Ion Ratio Lower Upper
153 100
152 46.7 37.5 56.3
154 86.6 68.8 103.2




#53
2,4-Dinitrophenol
Concen: 16.749 ng/ul
RT: 14.805 min Scan# 1943
Delta R.T. 0.007 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00



Tgt Ion:184 Resp: 7674
Ion Ratio Lower Upper
184 100
63 48.5 34.7 52.1
154 62.1 87.3 130.9#









#55
4-Nitrophenol
Concen: 20.363 ng/ul
RT: 14.899 min Scan# 1958
Delta R.T. 0.007 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:109 Resp: 13984
Ion Ratio Lower Upper
109 100
139 107.0 92.9 139.3
65 116.2 97.8 146.6



#56
Dibenzofuran
Concen: 20.640 ng/ul
RT: 15.099 min Scan# 1993
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:168 Resp: 120470
Ion Ratio Lower Upper
168 100
139 35.1 28.2 42.2









#57
2,4-Dinitrotoluene
Concen: 21.958 ng/ul
RT: 15.052 min Scan# 1985
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:165 Resp: 31502
Ion Ratio Lower Upper
165 100
63 36.8 33.0 49.4
182 2.3 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.650 ng/ul
RT: 15.328 min Scan# 2032
Delta R.T. 0.007 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion:232 Resp: 30528
Ion Ratio Lower Upper
232 100
131 26.6 27.5 41.3#
130 1.7 1.8 2.8#
166 21.0 20.7 31.1








#59
Diethylphthalate
Concen: 20.593 ng/ul
RT: 15.511 min Scan# 2063
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:149 Resp: 99292
Ion Ratio Lower Upper
149 100
177 24.4 18.9 28.3
150 12.6 10.6 15.8



#61
Fluorene
Concen: 20.740 ng/ul
RT: 15.746 min Scan# 2103
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:166 Resp: 98591
Ion Ratio Lower Upper
166 100
165 99.9 81.7 122.5
167 13.5 11.0 16.4





#62

4-Chlorophenyl-phenylether

Concen: 20.546 ng/ul

RT: 15.740 min Scan# 2102

Delta R.T. 0.001 min


Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:204 Resp: 62102

Ion Ratio Lower Upper

204 100

206 32.3 27.1 40.7

141 46.4 40.7 61.1



#63

4-Nitroaniline

Concen: 21.430 ng/ul

RT: 15.757 min Scan# 2105

Delta R.T. 0.007 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00




Tgt Ion:138 Resp: 16544


Ion Ratio Lower Upper

138 100

92 49.5 42.0 63.0

108 96.4 79.0 118.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.141 ng/ul
 RT: 15.816 min Scan# 2115
 Delta R.T. -0.005 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion:198 Resp: 18389
 Ion Ratio Lower Upper
 198 100
 182 3.8 3.4 5.0
 77 16.9 14.7 22.1



#67
 N-Nitrosodiphenylamine
 Concen: 20.944 ng/ul
 RT: 15.945 min Scan# 2137
 Delta R.T. -0.005 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Tgt Ion:169 Resp: 89278
 Ion Ratio Lower Upper
 169 100
 168 65.8 51.0 76.4
 167 37.2 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 20.458 ng/ul
RT: 16.633 min Scan# 21
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:248 Resp: 44307
Ion Ratio Lower Upper
248 100
250 98.2 78.2 117.2
141 50.9 46.7 70.1



#69
Hexachlorobenzene
Concen: 20.348 ng/ul
RT: 16.756 min Scan# 2275
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:284 Resp: 51650
Ion Ratio Lower Upper
284 100
142 22.1 20.8 31.2
249 32.3 27.4 41.0





#70

Atrazine

Concen: 20.438 ng/ul

RT: 16.891 min Scan# 21

Delta R.T. 0.001 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:200 Resp: 42791

Ion Ratio Lower Upper

200 100

173 23.4 19.8 29.6

215 51.8 38.8 58.2



#71

Pentachlorophenol

Concen: 20.592 ng/ul

RT: 17.097 min Scan# 2333

Delta R.T. 0.001 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00




Tgt Ion:266 Resp: 22155


Ion Ratio Lower Upper

266 100

264 66.6 56.8 85.2

268 66.5 48.0 72.0









#72
 Phenanthrene
 Concen: 20.829 ng/u1
 RT: 17.485 min Scan# 21
 Delta R.T. -0.005 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion:178 Resp: 179923
 Ion Ratio Lower Upper
 178 100
 179 14.9 12.6 19.0
 176 18.0 15.7 23.5



#74
 Anthracene
 Concen: 20.600 ng/u1
 RT: 17.579 min Scan# 2415
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Tgt Ion:178 Resp: 179341
 Ion Ratio Lower Upper
 178 100
 179 15.0 12.4 18.6
 176 18.4 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.122 ng/uL
RT: 13.507 min Scan# 1722
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:216 Resp: 56786
Ion Ratio Lower Upper
216 100
214 77.3 64.2 96.4
218 47.8 39.0 58.4



#76
Pentachlorobenzene
Concen: 20.134 ng/uL
RT: 15.023 min Scan# 1980
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00


Tgt Ion:250 Resp: 54992
Ion Ratio Lower Upper
250 100
248 60.8 47.7 71.5
252 62.4 50.0 75.0








#77
 Carbazole
 Concen: 20.953 ng/ul
 RT: 17.843 min Scan# 2460
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Instrument :
 BNA_G
 ClientSampleId :







Tgt Ion:167 Resp: 150395
 Ion Ratio Lower Upper
 167 100
 166 22.6 18.2 27.2
 139 10.8 9.0 13.6



#78
 Di-n-butylphthalate
 Concen: 20.954 ng/ul
 RT: 18.401 min Scan# 2555
 Delta R.T. -0.005 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00


Tgt Ion:149 Resp: 171232
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.1 10.7
 104 4.9 4.2 6.2








#80
Fluoranthene
Concen: 20.638 ng/u1
RT: 19.494 min Scan# 21
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:202 Resp: 237921
Ion Ratio Lower Upper
202 100
101 6.1 6.6 9.8#
100 4.6 5.8 8.8#



#82
Pyrene
Concen: 20.353 ng/u1
RT: 19.858 min Scan# 2803
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:202 Resp: 233276
Ion Ratio Lower Upper
202 100
101 6.7 7.3 10.9#
100 5.6 6.2 9.4#





#83

Butylbenzylphthalate

Concen: 21.485 ng/ul

RT: 20.740 min Scan# 2953

Delta R.T. 0.001 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:149 Resp: 74213

Ion Ratio Lower Upper

149 100

91 62.1 54.0 81.0

206 27.2 19.5 29.3



#84

3,3'-Dichlorobenzidine

Concen: 22.431 ng/ul

RT: 21.609 min Scan# 3101

Delta R.T. -0.005 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00




Tgt Ion:252 Resp: 91942


Ion Ratio Lower Upper

252 100

254 65.4 49.6 74.4

126 6.5 6.0 9.0









#85
Benzo(a)anthracene
Concen: 20.188 ng/ul
RT: 21.697 min Scan# 3116
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:228 Resp: 246459
Ion Ratio Lower Upper
228 100
229 20.2 16.5 24.7
226 27.9 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.908 ng/ul
RT: 21.609 min Scan# 3101
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:149 Resp: 105529
Ion Ratio Lower Upper
149 100
167 28.7 21.7 32.5
279 6.3 3.9 5.9#









#87
Chrysene
Concen: 20.263 ng/ul
RT: 21.768 min Scan# 3128
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:228 Resp: 237327
Ion Ratio Lower Upper
228 100
226 30.2 24.2 36.4
229 19.2 16.3 24.5



#89
Di-n-octyl phthalate
Concen: 21.272 ng/ul
RT: 22.831 min Scan# 3309
Delta R.T. -0.005 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:149 Resp: 185263









#90
Benzo(b)fluoranthene
Concen: 21.102 ng/ul
RT: 23.942 min Scan# 3498
Delta R.T. 0.001 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:252 Resp: 264237
Ion Ratio Lower Upper
252 100
253 22.4 18.2 27.4
125 4.6 6.1 9.1#



#91
Benzo(k)fluoranthene
Concen: 20.303 ng/ul
RT: 24.006 min Scan# 3509
Delta R.T. -0.011 min
Lab File: BGCCC985.D
Acq: 21 Jun 2022 20:00

Tgt Ion:252 Resp: 243424
Ion Ratio Lower Upper
252 100
253 22.7 17.4 26.2
125 4.8 5.4 8.0#





#93

Benzo(a)pyrene

Concen: 20.915 ng/ul

RT: 24.823 min Scan# 3

Delta R.T. -0.005 min


Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:252 Resp: 252274

Ion Ratio Lower Upper

252 100

253 21.6 17.6 26.4

125 6.2 6.8 10.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.742 ng/ul

RT: 28.701 min Scan# 4308

Delta R.T. 0.001 min

Lab File: BGCCC985.D

Acq: 21 Jun 2022 20:00




Tgt Ion:276 Resp: 301011


Ion Ratio Lower Upper

276 100

138 8.8 10.1 15.1#


277 23.9 19.9 29.9








#95
 Dibenzo(a,h)anthracene
 Concen: 20.873 ng/ul
 RT: 28.766 min Scan# 41
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Instrument :
 BNA_G
 ClientSampleId :



Tgt Ion:278 Resp: 254915
 Ion Ratio Lower Upper
 278 100
 139 8.1 8.6 13.0#
 279 23.1 19.4 29.0



#96
 Benzo(g,h,i)perylene
 Concen: 20.513 ng/ul
 RT: 29.858 min Scan# 4505
 Delta R.T. 0.001 min
 Lab File: BGCCC985.D
 Acq: 21 Jun 2022 20:00

Tgt Ion:276 Resp: 251342
 Ion Ratio Lower Upper
 276 100
 138 11.5 11.3 16.9
 277 24.1 18.6 27.8

