

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062122\
 Data File : BG053992.D
 Acq On : 22 Jun 2022 00:33
 Operator : CG/JU
 Sample : N3400-06
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0AD3

Quant Time: Jun 22 03:10:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 15:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	21395	20.000	ng/u1	0.00
20) Naphthalene-d8	10.885	136	86558	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.698	164	67207	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.442	188	182076	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.720	240	199024	20.000	ng/u1	# 0.00
88) Perylene-d12	24.980	264	204464	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.506	96	1913	4.254	ng/uL	0.00
4) Pyridine-d5	3.923	84	11100	9.076	ng/u1	0.00
7) Phenol-d5	7.236	99	6934	4.524	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.401	67	19115	20.674	ng/u1	0.00
11) 2-Chlorophenol-d4	7.613	132	14960	12.239	ng/u1	0.00
15) 4-Methylphenol-d8	8.776	113	12064	9.231	ng/u1	0.00
21) Nitrobenzene-d5	9.234	128	13258	20.636	ng/u1	0.00
24) 2-Nitrophenol-d4	9.963	143	9849	14.428	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.503	165	21490	14.229	ng/u1	0.00
31) 4-Chloroaniline-d4	11.014	131	37117	19.453	ng/u1	0.00
46) Dimethylphthalate-d6	14.099	166	127972	24.204	ng/u1	0.00
49) Acenaphthylene-d8	14.393	160	128378	22.330	ng/u1	0.00
54) 4-Nitrophenol-d4	14.892	143	2380	3.288	ng/u1	0.01
60) Fluorene-d10	15.691	176	112860	23.764	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.803	200	12781	12.868	ng/u1	0.00
73) Anthracene-d10	17.542	188	201021	24.447	ng/u1	0.00
81) Pyrene-d10	19.828	212	272186	26.230	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.751	264	264754	25.652	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062122\
Data File : BG053992.D
Acq On : 22 Jun 2022 00:33
Operator : CG/JU
Sample : N3400-06
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0AD3

Quant Time: Jun 22 03:10:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 20 15:29:13 2022
Response via : Initial Calibration

