

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062223\
 Data File : BG057820.D
 Acq On : 22 Jun 2023 23:03
 Operator : CG/JU
 Sample : 03320-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-19

Quant Time: Jun 23 03:47:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 03:40:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	39025	20.000	ng	0.00
21) Naphthalene-d8	10.725	136	166501	20.000	ng	0.00
39) Acenaphthene-d10	14.556	164	116482	20.000	ng	0.00
64) Phenanthrene-d10	17.294	188	293340	20.000	ng	-0.01
76) Chrysene-d12	21.548	240	270840	20.000	ng	-0.01
86) Perylene-d12	24.697	264	327395	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.496	112	358459	141.630	ng	0.00
7) Phenol-d6	7.082	99	486101	125.872	ng	0.00
23) Nitrobenzene-d5	9.080	82	309736	95.475	ng	-0.01
42) 2,4,6-Tribromophenol	16.042	330	246337	151.818	ng	0.00
45) 2-Fluorobiphenyl	13.181	172	693200	90.263	ng	0.00
79) Terphenyl-d14	19.914	244	1367381	94.141	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062223\
Data File : BG057820.D
Acq On : 22 Jun 2023 23:03
Operator : CG/JU
Sample : 03320-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-19

Quant Time: Jun 23 03:47:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 03:40:38 2023
Response via : Initial Calibration

