

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062322\
 Data File : BG054043.D
 Acq On : 24 Jun 2022 11:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020572

Quant Time: Jun 24 23:52:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 15:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.073	152	26406	20.000	ng/u1	0.00
20) Naphthalene-d8	10.881	136	119677	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.700	164	96056	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.438	188	244180	20.000	ng/u1	0.00
79) Chrysene-d12	21.716	240	269109	20.000	ng/u1	0.00
88) Perylene-d12	24.971	264	270293	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.496	96	4503	8.114	ng/uL	-0.01
4) Pyridine-d5	3.907	84	31160	20.643	ng/u1	-0.01
7) Phenol-d5	7.233	99	42592	22.515	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.391	67	25806	22.614	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.609	132	33029	21.894	ng/u1	0.00
15) 4-Methylphenol-d8	8.778	113	37364	23.164	ng/u1	0.00
21) Nitrobenzene-d5	9.230	128	17865	20.111	ng/u1	0.00
24) 2-Nitrophenol-d4	9.959	143	19800	20.978	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.499	165	43298	20.735	ng/u1	0.00
31) 4-Chloroaniline-d4	11.011	131	55905	21.192	ng/u1	0.00
46) Dimethylphthalate-d6	14.095	166	145253	19.222	ng/u1	0.00
49) Acenaphthylene-d8	14.395	160	161625	19.670	ng/u1	0.00
54) 4-Nitrophenol-d4	14.888	143	21574	20.851	ng/u1	0.00
60) Fluorene-d10	15.687	176	131573	19.384	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.799	200	25874	19.425	ng/u1	0.00
73) Anthracene-d10	17.538	188	216776	19.658	ng/u1	0.00
81) Pyrene-d10	19.824	212	281158	20.038	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.747	264	267636	19.616	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.531	88	5177	8.548	ng/uL#	84
5) Pyridine	3.931	79	32326	20.650	ng/u1	95
6) Benzaldehyde	7.203	77	25454	26.423	ng/u1	96
8) Phenol	7.256	94	45502	23.052	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.485	93	33919	22.685	ng/u1	89
12) 2-Chlorophenol	7.638	128	34153	22.174	ng/u1	95
13) 2-Methylphenol	8.514	108	34941	22.804	ng/u1	91
14) 2,2'-oxybis(1-Chloropr...	8.596	45	54140	24.377	ng/u1	97
16) Acetophenone	8.890	105	56741	22.746	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.872	70	30106	23.675	ng/u1	98
18) 4-Methylphenol	8.837	108	39120	23.866	ng/u1	99
19) Hexachloroethane	9.160	117	12894	19.756	ng/u1	89
22) Nitrobenzene	9.277	77	43202	20.418	ng/u1#	93
23) Isophorone	9.794	82	83586	20.375	ng/u1	96
25) 2-Nitrophenol	9.988	139	21327	21.549	ng/u1	94
26) 2,4-Dimethylphenol	10.047	107	45632	21.115	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.276	93	49587	21.471	ng/u1	95
29) 2,4-Dichlorophenol	10.529	162	43881	21.013	ng/u1	94
30) Naphthalene	10.934	128	119560	20.139	ng/u1	98
32) 4-Chloroaniline	11.034	127	55120	21.070	ng/u1	97
33) Hexachlorobutadiene	11.222	225	31612	15.498	ng/u1	93
34) Caprolactam	11.780	113	12456	19.988	ng/u1	80
35) 4-Chloro-3-methylphenol	12.156	107	43852	21.837	ng/u1	99
36) 2-Methylnaphthalene	12.532	142	87941	20.024	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062322\
 Data File : BG054043.D
 Acq On : 24 Jun 2022 11:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 31 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD020572


Quant Time: Jun 24 23:52:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 15:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.750	142	89907	20.528	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.897	216	66884	17.130	ng/ul	97
40) Hexachlorocyclopentadiene	12.879	237	42901	19.592	ng/ul	97
41) 2,4,6-Trichlorophenol	13.132	196	39143	19.316	ng/ul	92
42) 2,4,5-Trichlorophenol	13.208	196	43196	18.856	ng/ul	97
43) 1,1'-Biphenyl	13.531	154	131518	19.755	ng/ul	99
44) 2-Chloronaphthalene	13.578	162	104345	18.873	ng/ul	99
45) 2-Nitroaniline	13.772	65	31173	22.412	ng/ul	98
47) Dimethylphthalate	14.142	163	140228	19.307	ng/ul	99
48) 2,6-Dinitrotoluene	14.266	165	30271	20.723	ng/ul	91
50) Acenaphthylene	14.418	152	168765	20.003	ng/ul	98
51) 3-Nitroaniline	14.589	138	26473	22.325	ng/ul#	96
52) Acenaphthene	14.759	153	111719	19.829	ng/ul	99
53) 2,4-Dinitrophenol	14.806	184	12658	18.705	ng/ul#	62
55) 4-Nitrophenol	14.900	109	18674	18.411	ng/ul	96
56) Dibenzofuran	15.094	168	169421	19.653	ng/ul	97
57) 2,4-Dinitrotoluene	15.047	165	43755	20.650	ng/ul#	96
58) 2,3,4,6-Tetrachlorophenol	15.323	232	38966	19.574	ng/ul#	93
59) Diethylphthalate	15.505	149	140444	19.721	ng/ul	99
61) Fluorene	15.740	166	138842	19.775	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.734	204	80302	17.988	ng/ul	98
63) 4-Nitroaniline	15.752	138	25957	22.765	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.811	198	25365	19.837	ng/ul	95
67) N-Nitrosodiphenylamine	15.946	169	123674	20.717	ng/ul	94
68) 4-Bromophenyl-phenylether	16.628	248	55265	18.221	ng/ul	96
69) Hexachlorobenzene	16.751	284	61784	17.380	ng/ul	94
70) Atrazine	16.886	200	56498	19.269	ng/ul	98
71) Pentachlorophenol	17.092	266	26841	17.814	ng/ul	96
72) Phenanthrene	17.485	178	243221	20.105	ng/ul	99
74) Anthracene	17.573	178	249124	20.434	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.502	216	72567	17.448	ng/ul	98
76) Pentachlorobenzene	15.018	250	69053	18.053	ng/ul	96
77) Carbazole	17.838	167	212539	21.144	ng/ul	99
78) Di-n-butylphthalate	18.396	149	254287	22.219	ng/ul	100
80) Fluoranthene	19.489	202	320229	20.444	ng/ul	97
82) Pyrene	19.853	202	318824	20.473	ng/ul	99
83) Butylbenzylphthalate	20.734	149	111982	23.861	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.610	252	115273	20.699	ng/ul	99
85) Benzo(a)anthracene	21.698	228	329084	19.839	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.604	149	158581	23.124	ng/ul	99
87) Chrysene	21.763	228	316249	19.873	ng/ul	99
89) Di-n-octyl phthalate	22.826	149	271954	23.428	ng/ul	100
90) Benzo(b)fluoranthene	23.937	252	342736	20.535	ng/ul	99
91) Benzo(k)fluoranthene	24.001	252	313316	19.605	ng/ul	99
93) Benzo(a)pyrene	24.818	252	325563	20.250	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	28.696	276	375235	19.399	ng/ul	98
95) Dibenzo(a,h)anthracene	28.749	278	322089	19.786	ng/ul	97
96) Benzo(g,h,i)perylene	29.847	276	309815	18.970	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 24 23:52:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 20 15:29:13 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.548 ng/uL
RT: 3.531 min Scan# 24
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion: 88 Resp: 5177
Ion Ratio Lower Upper
88 100
43 34.4 48.1 72.1#
58 76.4 59.8 89.6



#5
Pyridine
Concen: 20.650 ng/uL
RT: 3.931 min Scan# 92
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion: 79 Resp: 32326
Ion Ratio Lower Upper
79 100
52 88.8 75.4 113.2
51 41.1 35.2 52.8









#6
Benzaldehyde
Concen: 26.423 ng/ul
RT: 7.203 min Scan# 649
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion: 77 Resp: 25454
Ion Ratio Lower Upper
77 100
105 96.1 79.7 119.5
106 93.7 71.0 106.4



#8
Phenol
Concen: 23.052 ng/ul
RT: 7.256 min Scan# 658
Delta R.T. 0.002 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion: 94 Resp: 45502
Ion Ratio Lower Upper
94 100
65 30.0 24.1 36.1
66 41.2 32.5 48.7









#10
Bis(2-Chloroethyl)ether
Concen: 22.685 ng/ul
RT: 7.485 min Scan# 698
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion: 93 Resp: 33919
Ion Ratio Lower Upper
93 100
63 78.8 72.3 108.5
95 31.4 29.4 44.0



#12
2-Chlorophenol
Concen: 22.174 ng/ul
RT: 7.638 min Scan# 723
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:128 Resp: 34153
Ion Ratio Lower Upper
128 100
64 56.1 41.4 62.2
130 32.0 27.0 40.6





#13

2-Methylphenol

Concen: 22.804 ng/ul

RT: 8.514 min Scan# 81

Delta R.T. 0.002 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09

Instrument :

BNA_G

ClientSampleId :





SSTD020572

Tgt Ion:108 Resp: 34941

Ion Ratio Lower Upper

108 100

107 86.1 75.8 113.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.377 ng/ul

RT: 8.596 min Scan# 886

Delta R.T. -0.010 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09




Tgt Ion: 45 Resp: 54140


Ion Ratio Lower Upper

45 100

77 11.6 10.2 15.4

79 10.8 8.1 12.1









#16
Acetophenone
Concen: 22.746 ng/ul
RT: 8.890 min Scan# 91
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:105 Resp: 56741
Ion Ratio Lower Upper
105 100
77 75.1 61.0 91.6
51 29.4 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 23.675 ng/ul
RT: 8.872 min Scan# 933
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion: 70 Resp: 30106
Ion Ratio Lower Upper
70 100
42 61.1 50.2 75.4
101 8.8 7.1 10.7
130 21.2 15.5 23.3









#18
4-Methylphenol
Concen: 23.866 ng/ul
RT: 8.837 min Scan# 91
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:108 Resp: 39120
Ion Ratio Lower Upper
108 100
107 112.7 90.7 136.1



#19
Hexachloroethane
Concen: 19.756 ng/ul
RT: 9.160 min Scan# 982
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:117 Resp: 12894
Ion Ratio Lower Upper
117 100
201 134.3 117.4 176.0
199 92.0 64.0 96.0









#22
Nitrobenzene
Concen: 20.418 ng/ul
RT: 9.277 min Scan# 1002
Delta R.T. 0.002 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion: 77 Resp: 43202
Ion Ratio Lower Upper
77 100
123 43.1 31.2 46.8
65 16.0 10.5 15.7#



#23
Isophorone
Concen: 20.375 ng/ul
RT: 9.794 min Scan# 1090
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion: 82 Resp: 83586
Ion Ratio Lower Upper
82 100
95 6.4 6.1 9.1
138 18.7 13.6 20.4





#25

2-Nitrophenol

Concen: 21.549 ng/ul

RT: 9.988 min Scan# 1123

Delta R.T. -0.010 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09

Instrument :

BNA_G

ClientSampleId :

SSTD020572





Tgt Ion:139 Resp: 21327

Ion Ratio Lower Upper

139 100

65 54.1 47.7 71.5

109 31.4 27.1 40.7



#26

2,4-Dimethylphenol

Concen: 21.115 ng/ul

RT: 10.047 min Scan# 1133

Delta R.T. -0.004 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09




Tgt Ion:107 Resp: 45632


Ion Ratio Lower Upper

107 100

121 47.4 39.9 59.9

122 81.0 66.7 100.1





#27

Bis(2-Chloroethoxy)methane

Concen: 21.471 ng/ul

RT: 10.276 min Scan# 1172

Delta R.T. -0.010 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09

Instrument :

BNA_G

ClientSampleId :

SSTD020572





Tgt Ion: 93 Resp: 49587

Ion Ratio Lower Upper

93 100

95 29.5 26.5 39.7

123 10.6 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 21.013 ng/ul

RT: 10.529 min Scan# 1215

Delta R.T. -0.004 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09




Tgt Ion:162 Resp: 43881


Ion Ratio Lower Upper

162 100

164 67.6 50.9 76.3

98 30.9 28.5 42.7









#30
Naphthalene
Concen: 20.139 ng/ul
RT: 10.934 min Scan# 1101
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:128 Resp: 119560
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.6
127 13.0 11.1 16.7



#32
4-Chloroaniline
Concen: 21.070 ng/ul
RT: 11.034 min Scan# 1301
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:127 Resp: 55120
Ion Ratio Lower Upper
127 100
129 33.6 25.4 38.2









#33
Hexachlorobutadiene
Concen: 15.498 ng/ul
RT: 11.222 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:225 Resp: 31612
Ion Ratio Lower Upper
225 100
223 68.6 50.2 75.2
227 69.1 51.8 77.6



#34
Caprolactam
Concen: 19.988 ng/ul
RT: 11.780 min Scan# 1428
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:113 Resp: 12456
Ion Ratio Lower Upper
113 100
55 173.0 167.8 251.6
56 137.8 127.6 191.4









#35
4-Chloro-3-methylphenol
Concen: 21.837 ng/ul
RT: 12.156 min Scan# 1492
Delta R.T. 0.002 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:	107	Resp:	43852
Ion	Ratio	Lower	Upper
107	100		
144	26.2	21.3	31.9
142	79.9	65.0	97.4



#36
2-Methylnaphthalene
Concen: 20.024 ng/ul
RT: 12.532 min Scan# 1556
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:	142	Resp:	87941
Ion	Ratio	Lower	Upper
142	100		
141	93.3	76.3	114.5





#37

1-Methylnaphthalene

Concen: 20.528 ng/ul

RT: 12.750 min Scan# 1

Delta R.T. -0.004 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09

Instrument :

BNA_G

ClientSampleId :

SSTD020572





Tgt Ion:142 Resp: 89907

Ion Ratio Lower Upper

142 100

141 92.4 73.4 110.2

116 3.8 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.130 ng/ul

RT: 12.897 min Scan# 1618

Delta R.T. -0.010 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09

Tgt Ion:216 Resp: 66884




Ion Ratio Lower Upper


216 100

214 75.6 58.1 87.1

179 17.6 14.8 22.2

108 12.4 10.4 15.6









#40
Hexachlorocyclopentadiene
Concen: 19.592 ng/ul
RT: 12.879 min Scan# 1615
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:237 Resp: 42901
Ion Ratio Lower Upper
237 100
235 60.7 47.0 70.6
272 12.7 8.6 12.8



#41
2,4,6-Trichlorophenol
Concen: 19.316 ng/ul
RT: 13.132 min Scan# 1658
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:196 Resp: 39143
Ion Ratio Lower Upper
196 100
198 93.3 81.9 122.9
200 28.4 25.7 38.5









#42
2,4,5-Trichlorophenol
Concen: 18.856 ng/ul
RT: 13.208 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:196 Resp: 43196
Ion Ratio Lower Upper
196 100
198 92.9 76.3 114.5
200 30.0 25.5 38.3



#43
1,1'-Biphenyl
Concen: 19.755 ng/ul
RT: 13.531 min Scan# 1726
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:154 Resp: 131518
Ion Ratio Lower Upper
154 100
153 41.2 32.7 49.1
76 10.7 9.0 13.6









#44
2-Chloronaphthalene
Concen: 18.873 ng/ul
RT: 13.578 min Scan# 1734
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:162 Resp: 104345
Ion Ratio Lower Upper
162 100
164 32.9 25.7 38.5
127 35.5 28.3 42.5



#45
2-Nitroaniline
Concen: 22.412 ng/ul
RT: 13.772 min Scan# 1767
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion: 65 Resp: 31173
Ion Ratio Lower Upper
65 100
92 65.6 52.2 78.4
138 101.8 79.0 118.4









#47
Dimethylphthalate
Concen: 19.307 ng/ul
RT: 14.142 min Scan# 1830
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:163 Resp: 140228
Ion Ratio Lower Upper
163 100
194 4.4 3.7 5.5
164 10.2 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.723 ng/ul
RT: 14.266 min Scan# 1851
Delta R.T. 0.002 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:165 Resp: 30271
Ion Ratio Lower Upper
165 100
89 48.4 44.7 67.1
121 17.8 16.9 25.3









#50
Acenaphthylene
Concen: 20.003 ng/ul
RT: 14.418 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:152 Resp: 168765
Ion Ratio Lower Upper
152 100
151 19.0 16.2 24.2
153 13.8 10.5 15.7



#51
3-Nitroaniline
Concen: 22.325 ng/ul
RT: 14.589 min Scan# 1906
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:138 Resp: 26473
Ion Ratio Lower Upper
138 100
108 12.2 17.8 26.6#
92 116.4 94.2 141.4









#52
Acenaphthene
Concen: 19.829 ng/ul
RT: 14.759 min Scan# 1935
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:153 Resp: 111719
Ion Ratio Lower Upper
153 100
152 47.1 37.5 56.3
154 87.6 68.8 103.2



#53
2,4-Dinitrophenol
Concen: 18.705 ng/ul
RT: 14.806 min Scan# 1943
Delta R.T. 0.008 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:184 Resp: 12658
Ion Ratio Lower Upper
184 100
63 45.3 34.7 52.1
154 55.1 87.3 130.9#









#55
4-Nitrophenol
Concen: 18.411 ng/ul
RT: 14.900 min Scan# 1959
Delta R.T. 0.008 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:109 Resp: 18674
Ion Ratio Lower Upper
109 100
139 112.7 92.9 139.3
65 116.7 97.8 146.6



#56
Dibenzofuran
Concen: 19.653 ng/ul
RT: 15.094 min Scan# 1992
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:168 Resp: 169421
Ion Ratio Lower Upper
168 100
139 33.6 28.2 42.2









#57
2,4-Dinitrotoluene
Concen: 20.650 ng/ul
RT: 15.047 min Scan# 1984
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:165 Resp: 43755
Ion Ratio Lower Upper
165 100
63 38.8 33.0 49.4
182 2.5 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.574 ng/ul
RT: 15.323 min Scan# 2031
Delta R.T. 0.002 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:232 Resp: 38966
Ion Ratio Lower Upper
232 100
131 29.7 27.5 41.3
130 1.8 1.8 2.8#
166 23.2 20.7 31.1









#59
Diethylphthalate
Concen: 19.721 ng/ul
RT: 15.505 min Scan# 2062
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:149 Resp: 140444
Ion Ratio Lower Upper
149 100
177 23.3 18.9 28.3
150 12.5 10.6 15.8



#61
Fluorene
Concen: 19.775 ng/ul
RT: 15.740 min Scan# 2102
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:166 Resp: 138842
Ion Ratio Lower Upper
166 100
165 98.2 81.7 122.5
167 13.1 11.0 16.4





#62

4-Chlorophenyl-phenylether

Concen: 17.988 ng/ul

RT: 15.734 min Scan# 2101

Delta R.T. -0.004 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09

Instrument :

BNA_G

ClientSampleId :

SSTD020572





Tgt Ion:204 Resp: 80302

Ion Ratio Lower Upper

204 100

206 33.5 27.1 40.7

141 48.9 40.7 61.1



#63

4-Nitroaniline

Concen: 22.765 ng/ul

RT: 15.752 min Scan# 2104

Delta R.T. 0.002 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09




Tgt Ion:138 Resp: 25957


Ion Ratio Lower Upper

138 100

92 51.5 42.0 63.0

108 96.3 79.0 118.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.837 ng/ul
 RT: 15.811 min Scan# 2115
 Delta R.T. -0.010 min
 Lab File: BG054043.D
 Acq: 24 Jun 2022 11:09




Instrument :
 BNA_G
 ClientSampleId :
 SST020572


Tgt Ion:198 Resp: 25365
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.4 5.0
 77 21.3 14.7 22.1



#67
 N-Nitrosodiphenylamine
 Concen: 20.717 ng/ul
 RT: 15.946 min Scan# 2137
 Delta R.T. -0.004 min
 Lab File: BG054043.D
 Acq: 24 Jun 2022 11:09

Tgt Ion:169 Resp: 123674
 Ion Ratio Lower Upper
 169 100
 168 68.8 51.0 76.4
 167 38.2 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 18.221 ng/ul
RT: 16.628 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:248 Resp: 55265
Ion Ratio Lower Upper
248 100
250 95.1 78.2 117.2
141 53.6 46.7 70.1



#69
Hexachlorobenzene
Concen: 17.380 ng/ul
RT: 16.751 min Scan# 2274
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:284 Resp: 61784
Ion Ratio Lower Upper
284 100
142 23.3 20.8 31.2
249 30.3 27.4 41.0









#70
 Atrazine
 Concen: 19.269 ng/ul
 RT: 16.886 min Scan# 21
 Delta R.T. -0.004 min
 Lab File: BG054043.D
 Acq: 24 Jun 2022 11:09




Instrument :
 BNA_G
 ClientSampleId :
 SST020572


Tgt Ion:200 Resp: 56498
 Ion Ratio Lower Upper
 200 100
 173 22.4 19.8 29.6
 215 49.2 38.8 58.2



#71
 Pentachlorophenol
 Concen: 17.814 ng/ul
 RT: 17.092 min Scan# 2332
 Delta R.T. -0.004 min
 Lab File: BG054043.D
 Acq: 24 Jun 2022 11:09

Tgt Ion:266 Resp: 26841
 Ion Ratio Lower Upper
 266 100
 264 65.9 56.8 85.2
 268 61.2 48.0 72.0









#72
Phenanthrene
Concen: 20.105 ng/ul
RT: 17.485 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:178 Resp: 243221
Ion Ratio Lower Upper
178 100
179 15.0 12.6 19.0
176 19.2 15.7 23.5



#74
Anthracene
Concen: 20.434 ng/ul
RT: 17.573 min Scan# 2414
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:178 Resp: 249124
Ion Ratio Lower Upper
178 100
179 15.6 12.4 18.6
176 18.4 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.448 ng/uL
RT: 13.502 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:216 Resp: 72567
Ion Ratio Lower Upper
216 100
214 78.0 64.2 96.4
218 48.4 39.0 58.4



#76
Pentachlorobenzene
Concen: 18.053 ng/uL
RT: 15.018 min Scan# 1979
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:250 Resp: 69053
Ion Ratio Lower Upper
250 100
248 65.3 47.7 71.5
252 63.0 50.0 75.0









#77
 Carbazole
 Concen: 21.144 ng/ul
 RT: 17.838 min Scan# 2460
 Delta R.T. -0.004 min
 Lab File: BG054043.D
 Acq: 24 Jun 2022 11:09




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020572


Tgt Ion:167 Resp: 212539
 Ion Ratio Lower Upper
 167 100
 166 23.0 18.2 27.2
 139 11.5 9.0 13.6



#78
 Di-n-butylphthalate
 Concen: 22.219 ng/ul
 RT: 18.396 min Scan# 2554
 Delta R.T. -0.010 min
 Lab File: BG054043.D
 Acq: 24 Jun 2022 11:09

Tgt Ion:149 Resp: 254287
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.1 10.7
 104 5.0 4.2 6.2









#80
Fluoranthene
Concen: 20.444 ng/ul
RT: 19.489 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:202 Resp: 320229
Ion Ratio Lower Upper
202 100
101 7.5 6.6 9.8
100 6.1 5.8 8.8



#82
Pyrene
Concen: 20.473 ng/ul
RT: 19.853 min Scan# 2802
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:202 Resp: 318824
Ion Ratio Lower Upper
202 100
101 9.3 7.3 10.9
100 7.6 6.2 9.4









#83
Butylbenzylphthalate
Concen: 23.861 ng/ul
RT: 20.734 min Scan# 2952
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:149 Resp: 111982
Ion Ratio Lower Upper
149 100
91 65.2 54.0 81.0
206 24.0 19.5 29.3



#84
3,3'-Dichlorobenzidine
Concen: 20.699 ng/ul
RT: 21.610 min Scan# 3101
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:252 Resp: 115273
Ion Ratio Lower Upper
252 100
254 62.5 49.6 74.4
126 7.3 6.0 9.0





#85

Benzo(a)anthracene

Concen: 19.839 ng/ul

RT: 21.698 min Scan# 3116

Delta R.T. -0.004 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09

Instrument :

BNA_G

ClientSampleId :

SSTD020572





Tgt Ion:228 Resp: 329084

Ion Ratio Lower Upper

228 100

229 19.8 16.5 24.7

226 27.6 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.124 ng/ul

RT: 21.604 min Scan# 3100

Delta R.T. -0.004 min

Lab File: BG054043.D

Acq: 24 Jun 2022 11:09




Tgt Ion:149 Resp: 158581


Ion Ratio Lower Upper

149 100

167 28.0 21.7 32.5

279 4.9 3.9 5.9









#87
Chrysene
Concen: 19.873 ng/ul
RT: 21.763 min Scan# 3127
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:228 Resp: 316249
Ion Ratio Lower Upper
228 100
226 30.1 24.2 36.4
229 20.1 16.3 24.5



#89
Di-n-octyl phthalate
Concen: 23.428 ng/ul
RT: 22.826 min Scan# 3308
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:149 Resp: 271954









#90
Benzo(b)fluoranthene
Concen: 20.535 ng/ul
RT: 23.937 min Scan# 3497
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:252 Resp: 342736
Ion Ratio Lower Upper
252 100
253 22.4 18.2 27.4
125 6.4 6.1 9.1



#91
Benzo(k)fluoranthene
Concen: 19.605 ng/ul
RT: 24.001 min Scan# 3508
Delta R.T. -0.016 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:252 Resp: 313316
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.2
125 6.3 5.4 8.0









#93
Benzo(a)pyrene
Concen: 20.250 ng/ul
RT: 24.818 min Scan# 3647
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09




Instrument :
BNA_G
ClientSampleId :
SSTD020572


Tgt Ion:252 Resp: 325563
Ion Ratio Lower Upper
252 100
253 20.6 17.6 26.4
125 7.1 6.8 10.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.399 ng/ul
RT: 28.696 min Scan# 4307
Delta R.T. -0.004 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:276 Resp: 375235
Ion Ratio Lower Upper
276 100
138 12.8 10.1 15.1
277 23.7 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 19.786 ng/ul
RT: 28.749 min Scan# 4316
Delta R.T. -0.016 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Instrument :
BNA_G
ClientSampleId :
SSTD020572

Tgt Ion:278 Resp: 322089

Ion	Ratio	Lower	Upper
278	100		
139	9.8	8.6	13.0
279	22.6	19.4	29.0



#96
Benzo(g,h,i)perylene
Concen: 18.970 ng/ul
RT: 29.847 min Scan# 4503
Delta R.T. -0.010 min
Lab File: BG054043.D
Acq: 24 Jun 2022 11:09

Tgt Ion:276 Resp: 309815

Ion	Ratio	Lower	Upper
276	100		
138	12.5	11.3	16.9
277	24.0	18.6	27.8

