

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062322\
 Data File : BG054021.D
 Acq On : 23 Jun 2022 20:14
 Operator : CG/JU
 Sample : N3400-17
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0AF0

Quant Time: Jun 23 23:05:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 15:29:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.071	152	19992	20.000	ng/u1	0.00
20) Naphthalene-d8	10.880	136	93299	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.699	164	76549	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.443	188	194183	20.000	ng/u1	0.00
79) Chrysene-d12	21.714	240	218845	20.000	ng/u1	0.00
88) Perylene-d12	24.975	264	215435	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.494	96	1924	4.579	ng/uL	-0.01
4) Pyridine-d5	3.917	84	8738	7.646	ng/u1	0.00
7) Phenol-d5	7.231	99	13368	9.334	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.390	67	30406	35.194	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.607	132	33843	29.631	ng/u1	0.00
15) 4-Methylphenol-d8	8.776	113	25885	21.196	ng/u1	0.00
21) Nitrobenzene-d5	9.235	128	21873	31.585	ng/u1	0.00
24) 2-Nitrophenol-d4	9.957	143	23764	32.297	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.498	165	50096	30.773	ng/u1	0.00
31) 4-Chloroaniline-d4	11.009	131	53293	25.913	ng/u1	0.00
46) Dimethylphthalate-d6	14.099	166	204550	33.967	ng/u1	0.00
49) Acenaphthylene-d8	14.393	160	215269	32.874	ng/u1	0.00
54) 4-Nitrophenol-d4	14.887	143	5895	7.149	ng/u1	0.00
60) Fluorene-d10	15.686	176	181925	33.632	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.797	200	33732	31.844	ng/u1	0.00
73) Anthracene-d10	17.542	188	318913	36.366	ng/u1	0.00
81) Pyrene-d10	19.822	212	419933	36.803	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.746	264	402731	37.034	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062322\
Data File : BG054021.D
Acq On : 23 Jun 2022 20:14
Operator : CG/JU
Sample : N3400-17
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0AF0

Quant Time: Jun 23 23:05:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 20 15:29:13 2022
Response via : Initial Calibration

