

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
 Data File : BG054088.D
 Acq On : 27 Jun 2022 10:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 04:51:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	37.767	5.6	97	0.00
3	Pyridine	40.000	41.655	-4.1	104	0.00
4	n-Nitrosodimethylamine	40.000	44.828	-12.1	110	0.00
5 S	2-Fluorophenol	80.000	82.318	-2.9	105	0.00
6	Aniline	40.000	43.836	-9.6	111	0.00
7 S	Phenol-d6	80.000	90.723	-13.4	116	0.00
8	2-Chlorophenol	40.000	43.926	-9.8	113	0.00
9	Benzaldehyde	40.000	40.828	-2.1	114	0.00
10 C	Phenol	40.000	44.926	-12.3	113	0.00
11	bis(2-Chloroethyl)ether	40.000	42.010	-5.0	110	0.00
12	1,3-Dichlorobenzene	40.000	39.435	1.4	104	0.00
13 C	1,4-Dichlorobenzene	40.000	39.823	0.4	106	0.00
14	1,2-Dichlorobenzene	40.000	40.222	-0.6	105	0.00
15	Benzyl Alcohol	40.000	46.177	-15.4	115	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	46.991	-17.5	122	0.00
17	2-Methylphenol	40.000	45.281	-13.2	116	0.00
18	Hexachloroethane	40.000	41.372	-3.4	107	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	45.583	-14.0	116	0.00
20	3+4-Methylphenols	40.000	46.230	-15.6	118	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	119	0.00
22	Acetophenone	40.000	38.207	4.5	113	0.00
23 S	Nitrobenzene-d5	80.000	79.553	0.6	116	0.00
24	Nitrobenzene	40.000	41.132	-2.8	121	0.00
25	Isophorone	40.000	39.722	0.7	118	0.00
26 C	2-Nitrophenol	40.000	43.190	-8.0	124	0.00
27	2,4-Dimethylphenol	40.000	41.255	-3.1	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.854	0.4	118	0.00
29 C	2,4-Dichlorophenol	40.000	40.616	-1.5	119	0.00
30	1,2,4-Trichlorobenzene	40.000	36.173	9.6	107	0.00
31	Naphthalene	40.000	38.969	2.6	115	0.00
32	Benzoic acid	40.000	53.311	-33.3#	195	0.00
33	4-Chloroaniline	40.000	40.750	-1.9	118	0.00
34 C	Hexachlorobutadiene	40.000	34.265	14.3	100	0.00
35	Caprolactam	40.000	44.989	-12.5	126	0.01
36 C	4-Chloro-3-methylphenol	40.000	42.495	-6.2	122	0.00
37	2-Methylnaphthalene	40.000	39.436	1.4	116	0.00
38	1-Methylnaphthalene	40.000	39.235	1.9	117	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	122	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.353	11.6	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.534	-3.8	122	0.00
42 S	2,4,6-Tribromophenol	80.000	70.472	11.9	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.439	1.4	115	0.00
44	2,4,5-Trichlorophenol	40.000	39.312	1.7	116	0.00
45 S	2-Fluorobiphenyl	80.000	74.982	6.3	114	0.00
46	1,1'-Biphenyl	40.000	38.293	4.3	116	0.00
47	2-Chloronaphthalene	40.000	38.197	4.5	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
 Data File : BG054088.D
 Acq On : 27 Jun 2022 10:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 04:51:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.694	-11.7	130	0.00
49	Acenaphthylene	40.000	38.950	2.6	117	0.00
50	Dimethylphthalate	40.000	37.803	5.5	115	0.00
51	2,6-Dinitrotoluene	40.000	40.575	-1.4	119	0.00
52 C	Acenaphthene	40.000	39.668	0.8	119	0.00
53	3-Nitroaniline	40.000	42.384	-6.0	123	0.00
54 P	2,4-Dinitrophenol	40.000	39.803	0.5	129	0.00
55	Dibenzofuran	40.000	38.186	4.5	116	0.00
56 P	4-Nitrophenol	40.000	39.743	0.6	116	0.00
57	2,4-Dinitrotoluene	40.000	41.109	-2.8	120	0.00
58	Fluorene	40.000	38.425	3.9	117	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.262	1.8	116	0.00
60	Diethylphthalate	40.000	39.086	2.3	119	0.00
61	4-Chlorophenyl-phenylether	40.000	36.652	8.4	112	0.00
62	4-Nitroaniline	40.000	41.702	-4.3	124	0.00
63	Azobenzene	40.000	41.449	-3.6	127	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.564	-8.9	124	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.159	-0.4	116	0.00
67	4-Bromophenyl-phenylether	40.000	37.861	5.3	110	0.00
68	Hexachlorobenzene	40.000	35.253	11.9	104	0.00
69	Atrazine	40.000	37.969	5.1	108	0.00
70 C	Pentachlorophenol	40.000	39.789	0.5	112	0.00
71	Phenanthrene	40.000	38.603	3.5	113	0.00
72	Anthracene	40.000	39.349	1.6	115	0.00
73	Carbazole	40.000	39.324	1.7	115	0.00
74	Di-n-butylphthalate	40.000	40.953	-2.4	118	0.00
75 C	Fluoranthene	40.000	36.573	8.6	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	31.860	20.4	82	0.00
78	Pyrene	40.000	40.895	-2.2	107	0.00
79 S	Terphenyl-d14	80.000	77.780	2.8	101	0.00
80	Butylbenzylphthalate	40.000	46.094	-15.2	119	0.00
81	Benzo(a)anthracene	40.000	39.078	2.3	104	0.00
82	3,3'-Dichlorobenzidine	40.000	38.681	3.3	100	0.00
83	Chrysene	40.000	38.734	3.2	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.895	-14.7	118	0.00
85 c	Di-n-octyl phthalate	40.000	45.698	-14.2	117	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	37.705	5.7	96	0.02
88	Benzo(b)fluoranthene	40.000	38.714	3.2	98	0.00
89	Benzo(k)fluoranthene	40.000	38.990	2.5	102	0.00
90 C	Benzo(a)pyrene	40.000	39.275	1.8	100	0.01
91	Dibenzo(a,h)anthracene	40.000	38.027	4.9	98	0.02
92	Benzo(g,h,i)perylene	40.000	36.851	7.9	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
Data File : BG054088.D
Acq On : 27 Jun 2022 10:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 28 04:51:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0