

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
 Data File : BG054093.D
 Acq On : 27 Jun 2022 15:48
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 04:54:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.457	0.432	5.5	111	0.00
3	Pyridine	1.220	1.221	-0.1	114	0.00
4	n-Nitrosodimethylamine	0.537	0.588	-9.5	123	0.00
5 S	2-Fluorophenol	1.046	1.048	-0.2	117	0.00
6	Aniline	1.682	1.725	-2.6	119	0.00
7 S	Phenol-d6	1.419	1.526	-7.5	126	0.00
8	2-Chlorophenol	1.117	1.165	-4.3	122	0.00
9	Benzaldehyde	0.838	0.805	3.9	122	0.00
10 C	Phenol	1.438	1.554	-8.1	124	0.00
11	bis(2-Chloroethyl)ether	1.104	1.115	-1.0	120	0.00
12	1,3-Dichlorobenzene	1.400	1.347	3.8	116	0.00
13 C	1,4-Dichlorobenzene	1.406	1.364	3.0	118	0.00
14	1,2-Dichlorobenzene	1.350	1.313	2.7	115	0.00
15	Benzyl Alcohol	1.061	1.196	-12.7	129	0.00
16	2,2'-oxybis(1-Chloropropane	1.650	1.824	-10.5	130	0.00
17	2-Methylphenol	0.999	1.085	-8.6	127	0.00
18	Hexachloroethane	0.466	0.469	-0.6	119	0.00
19 P	n-Nitroso-di-n-propylamine	0.953	1.019	-6.9	124	0.00
20	3+4-Methylphenols	1.401	1.555	-11.0	130	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
22	Acetophenone	0.486	0.463	4.7	120	0.00
23 S	Nitrobenzene-d5	0.356	0.361	-1.4	126	0.00
24	Nitrobenzene	0.350	0.358	-2.3	127	0.00
25	Isophorone	0.665	0.668	-0.5	127	0.00
26 C	2-Nitrophenol	0.155	0.174	-12.3	137	0.00
27	2,4-Dimethylphenol	0.251	0.255	-1.6	128	0.00
28	bis(2-Chloroethoxy)methane	0.358	0.360	-0.6	128	0.00
29 C	2,4-Dichlorophenol	0.336	0.355	-5.7	132	0.00
30	1,2,4-Trichlorobenzene	0.416	0.383	7.9	116	0.00
31	Naphthalene	1.010	0.984	2.6	123	0.00
32	Benzoic acid	0.087	0.139	-59.8#	193#	0.00
33	4-Chloroaniline	0.419	0.440	-5.0	130	0.00
34 C	Hexachlorobutadiene	0.320	0.270	15.6	105	0.00
35	Caprolactam	0.092	0.110	-19.6	142	0.01
36 C	4-Chloro-3-methylphenol	0.327	0.372	-13.8	140	0.00
37	2-Methylnaphthalene	0.747	0.749	-0.3	126	0.00
38	1-Methylnaphthalene	0.727	0.726	0.1	127	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	135	0.00
40	1,2,4,5-Tetrachlorobenzene	0.771	0.652	15.4	112	0.00
41 P	Hexachlorocyclopentadiene	0.359	0.386	-7.5	139	0.00
42 S	2,4,6-Tribromophenol	0.321	0.295	8.1	120	0.00
43 C	2,4,6-Trichlorophenol	0.433	0.421	2.8	125	0.00
44	2,4,5-Trichlorophenol	0.483	0.477	1.2	128	0.00
45 S	2-Fluorobiphenyl	1.375	1.261	8.3	123	0.00
46	1,1'-Biphenyl	1.393	1.327	4.7	127	0.00
47	2-Chloronaphthalene	1.127	1.064	5.6	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
 Data File : BG054093.D
 Acq On : 27 Jun 2022 15:48
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 04:54:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.300	0.349	-16.3	149	0.00
49	Acenaphthylene	1.753	1.714	2.2	129	0.00
50	Dimethylphthalate	1.523	1.425	6.4	125	0.00
51	2,6-Dinitrotoluene	0.308	0.317	-2.9	134	0.00
52 C	Acenaphthene	1.118	1.119	-0.1	133	0.00
53	3-Nitroaniline	0.298	0.321	-7.7	137	0.00
54 P	2,4-Dinitrophenol	0.140	0.166	-18.6	155#	0.00
55	Dibenzofuran	1.801	1.746	3.1	130	0.00
56 P	4-Nitrophenol	0.212	0.233	-9.9	141	0.00
57	2,4-Dinitrotoluene	0.435	0.464	-6.7	137	0.00
58	Fluorene	1.478	1.448	2.0	131	0.00
59	2,3,4,6-Tetrachlorophenol	0.473	0.469	0.8	129	0.00
60	Diethylphthalate	1.475	1.419	3.8	129	0.00
61	4-Chlorophenyl-phenylether	0.885	0.805	9.0	122	0.00
62	4-Nitroaniline	0.311	0.333	-7.1	140	0.00
63	Azobenzene	1.187	1.254	-5.6	143	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	135	0.00
65	4,6-Dinitro-2-methylphenol	0.100	0.115	-15.0	149	0.00
66 c	n-Nitrosodiphenylamine	0.521	0.520	0.2	132	0.00
67	4-Bromophenyl-phenylether	0.250	0.231	7.6	123	0.00
68	Hexachlorobenzene	0.286	0.249	12.9	117	0.00
69	Atrazine	0.216	0.210	2.8	125	0.00
70 C	Pentachlorophenol	0.133	0.136	-2.3	132	0.00
71	Phenanthrene	1.028	1.001	2.6	130	0.00
72	Anthracene	1.003	0.984	1.9	130	0.00
73	Carbazole	0.861	0.865	-0.5	134	0.00
74	Di-n-butylphthalate	0.996	1.038	-4.2	137	0.00
75 C	Fluoranthene	1.349	1.266	6.2	125	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
77	Benzydine	0.414	0.339	18.1	96	0.00
78	Pyrene	1.208	1.263	-4.6	125	0.00
79 S	Terphenyl-d14	0.993	0.977	1.6	117	0.00
80	Butylbenzylphthalate	0.387	0.456	-17.8	139	0.00
81	Benzo(a)anthracene	1.295	1.280	1.2	120	0.00
82	3,3'-Dichlorobenzidine	0.387	0.381	1.6	116	0.00
83	Chrysene	1.234	1.203	2.5	119	0.00
84	Bis(2-ethylhexyl)phthalate	0.554	0.641	-15.7	137	0.00
85 c	Di-n-octyl phthalate	0.952	1.115	-17.1	138	0.00
86 I	Perylene-d12	1.000	1.000	0.0	119	0.01
87	Indeno(1,2,3-cd)pyrene	1.509	1.432	5.1	111	0.02
88	Benzo(b)fluoranthene	1.216	1.198	1.5	115	0.00
89	Benzo(k)fluoranthene	1.204	1.164	3.3	117	0.00
90 C	Benzo(a)pyrene	1.026	1.014	1.2	116	0.00
91	Dibenzo(a,h)anthracene	1.248	1.172	6.1	111	0.02
92	Benzo(g,h,i)perylene	1.232	1.161	5.8	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
Data File : BG054093.D
Acq On : 27 Jun 2022 15:48
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 28 04:54:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0