

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
 Data File : BG054095.D
 Acq On : 27 Jun 2022 17:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 04:55:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	122	0.00
2	1,4-Dioxane	40.000	34.798	13.0	105	0.00
3	Pyridine	40.000	40.480	-1.2	119	0.00
4	n-Nitrosodimethylamine	40.000	42.000	-5.0	122	0.00
5 S	2-Fluorophenol	80.000	79.418	0.7	119	0.00
6	Aniline	40.000	42.563	-6.4	127	0.00
7 S	Phenol-d6	80.000	88.001	-10.0	132	0.00
8	2-Chlorophenol	40.000	42.142	-5.4	127	0.00
9	Benzaldehyde	40.000	39.063	2.3	128	0.00
10 C	Phenol	40.000	44.582	-11.5	131	0.00
11	bis(2-Chloroethyl)ether	40.000	41.775	-4.4	128	0.00
12	1,3-Dichlorobenzene	40.000	38.125	4.7	118	0.00
13 C	1,4-Dichlorobenzene	40.000	38.815	3.0	122	0.00
14	1,2-Dichlorobenzene	40.000	39.018	2.5	119	0.00
15	Benzyl Alcohol	40.000	45.301	-13.3	133	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	45.568	-13.9	139	0.00
17	2-Methylphenol	40.000	44.906	-12.3	135	0.00
18	Hexachloroethane	40.000	41.055	-2.6	125	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	45.228	-13.1	136	0.00
20	3+4-Methylphenols	40.000	45.940	-14.8	138	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	138	0.00
22	Acetophenone	40.000	38.208	4.5	130	0.00
23 S	Nitrobenzene-d5	80.000	79.542	0.6	134	0.00
24	Nitrobenzene	40.000	40.259	-0.6	136	0.00
25	Isophorone	40.000	39.561	1.1	136	0.00
26 C	2-Nitrophenol	40.000	44.010	-10.0	145	0.00
27	2,4-Dimethylphenol	40.000	40.946	-2.4	139	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.215	-0.5	138	0.00
29 C	2,4-Dichlorophenol	40.000	40.972	-2.4	138	0.00
30	1,2,4-Trichlorobenzene	40.000	35.213	12.0	120	0.00
31	Naphthalene	40.000	38.453	3.9	132	0.00
32	Benzoic acid	40.000	45.925	-14.8	186	0.00
33	4-Chloroaniline	40.000	40.785	-2.0	137	0.00
34 C	Hexachlorobutadiene	40.000	33.411	16.5	113	0.00
35	Caprolactam	40.000	43.416	-8.5	141	0.01
36 C	4-Chloro-3-methylphenol	40.000	42.709	-6.8	142	0.00
37	2-Methylnaphthalene	40.000	39.419	1.5	134	0.00
38	1-Methylnaphthalene	40.000	39.271	1.8	135	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	140	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.295	11.8	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.684	-1.7	136	0.00
42 S	2,4,6-Tribromophenol	80.000	70.648	11.7	120	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.238	1.9	131	0.00
44	2,4,5-Trichlorophenol	40.000	39.949	0.1	134	0.00
45 S	2-Fluorobiphenyl	80.000	75.703	5.4	131	0.00
46	1,1'-Biphenyl	40.000	38.865	2.8	134	0.00
47	2-Chloronaphthalene	40.000	38.264	4.3	132	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
 Data File : BG054095.D
 Acq On : 27 Jun 2022 17:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 04:55:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.043	-12.6	149	0.00
49	Acenaphthylene	40.000	38.929	2.7	133	0.00
50	Dimethylphthalate	40.000	37.761	5.6	131	0.00
51	2,6-Dinitrotoluene	40.000	41.131	-2.8	138	0.00
52 C	Acenaphthene	40.000	39.968	0.1	137	0.00
53	3-Nitroaniline	40.000	41.834	-4.6	138	0.00
54 P	2,4-Dinitrophenol	40.000	38.937	2.7	144	0.00
55	Dibenzofuran	40.000	38.618	3.5	134	0.00
56 P	4-Nitrophenol	40.000	41.360	-3.4	137	0.00
57	2,4-Dinitrotoluene	40.000	40.496	-1.2	135	0.00
58	Fluorene	40.000	38.444	3.9	133	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.068	7.3	125	0.00
60	Diethylphthalate	40.000	38.401	4.0	134	0.00
61	4-Chlorophenyl-phenylether	40.000	36.425	8.9	127	0.00
62	4-Nitroaniline	40.000	40.762	-1.9	138	0.00
63	Azobenzene	40.000	41.235	-3.1	144	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	134	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.599	-11.5	145	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.787	0.5	131	0.00
67	4-Bromophenyl-phenylether	40.000	37.075	7.3	123	0.00
68	Hexachlorobenzene	40.000	34.631	13.4	116	0.00
69	Atrazine	40.000	37.884	5.3	122	0.00
70 C	Pentachlorophenol	40.000	38.281	4.3	123	0.00
71	Phenanthrene	40.000	39.065	2.3	130	0.00
72	Anthracene	40.000	38.768	3.1	129	0.00
73	Carbazole	40.000	38.706	3.2	129	0.00
74	Di-n-butylphthalate	40.000	40.363	-0.9	133	0.00
75 C	Fluoranthene	40.000	36.256	9.4	121	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	121	0.00
77	Benzydine	40.000	42.249	-5.6	123	0.00
78	Pyrene	40.000	40.846	-2.1	120	0.00
79 S	Terphenyl-d14	80.000	76.989	3.8	113	0.00
80	Butylbenzylphthalate	40.000	46.119	-15.3	134	0.00
81	Benzo(a)anthracene	40.000	39.038	2.4	116	0.00
82	3,3'-Dichlorobenzidine	40.000	38.922	2.7	113	0.00
83	Chrysene	40.000	38.311	4.2	115	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.082	-12.7	131	0.00
85 c	Di-n-octyl phthalate	40.000	45.731	-14.3	132	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.282	6.8	106	0.00
88	Benzo(b)fluoranthene	40.000	38.047	4.9	108	0.00
89	Benzo(k)fluoranthene	40.000	38.715	3.2	114	0.00
90 C	Benzo(a)pyrene	40.000	39.270	1.8	112	0.00
91	Dibenzo(a,h)anthracene	40.000	37.267	6.8	108	0.02
92	Benzo(g,h,i)perylene	40.000	36.605	8.5	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062722\
Data File : BG054095.D
Acq On : 27 Jun 2022 17:07
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 28 04:55:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0