

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062823\
 Data File : BG057883.D
 Acq On : 28 Jun 2023 8:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 08:51:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG062723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 01:12:33 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	38.916	2.7	111	0.00
3	Pyridine	40.000	40.304	-0.8	114	0.00
4	n-Nitrosodimethylamine	40.000	39.476	1.3	112	0.00
5 S	2-Fluorophenol	80.000	80.253	-0.3	113	0.00
6	Aniline	40.000	41.098	-2.7	120	0.00
7 S	Phenol-d6	80.000	83.038	-3.8	119	0.00
8	2-Chlorophenol	40.000	40.498	-1.2	115	0.00
9	Benzaldehyde	40.000	44.112	-10.3	117	0.00
10 C	Phenol	40.000	41.706	-4.3	119	0.00
11	bis(2-Chloroethyl)ether	40.000	40.386	-1.0	116	0.00
12	1,3-Dichlorobenzene	40.000	40.039	-0.1	113	0.00
13 C	1,4-Dichlorobenzene	40.000	39.661	0.8	113	0.00
14	1,2-Dichlorobenzene	40.000	39.614	1.0	113	0.00
15	Benzyl Alcohol	40.000	41.602	-4.0	122	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.056	2.4	114	0.00
17	2-Methylphenol	40.000	41.236	-3.1	119	0.00
18	Hexachloroethane	40.000	40.435	-1.1	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.880	-2.2	123	0.00
20	3+4-Methylphenols	40.000	42.726	-6.8	124	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	121	0.00
22	Acetophenone	40.000	39.566	1.1	120	0.00
23 S	Nitrobenzene-d5	80.000	79.971	0.0	120	0.00
24	Nitrobenzene	40.000	40.499	-1.2	121	0.00
25	Isophorone	40.000	39.869	0.3	125	0.00
26 C	2-Nitrophenol	40.000	42.438	-6.1	123	0.00
27	2,4-Dimethylphenol	40.000	40.511	-1.3	122	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.503	-1.3	125	0.00
29 C	2,4-Dichlorophenol	40.000	41.523	-3.8	124	0.00
30	1,2,4-Trichlorobenzene	40.000	39.949	0.1	120	0.00
31	Naphthalene	40.000	39.732	0.7	121	0.00
32	Benzoic acid	40.000	45.736	-14.3	141	0.00
33	4-Chloroaniline	40.000	41.148	-2.9	127	0.00
34 C	Hexachlorobutadiene	40.000	39.405	1.5	119	0.00
35	Caprolactam	40.000	39.557	1.1	123	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.563	-3.9	130	0.00
37	2-Methylnaphthalene	40.000	40.084	-0.2	125	0.00
38	1-Methylnaphthalene	40.000	40.271	-0.7	126	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	128	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.491	1.3	124	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.654	0.9	119	0.00
42 S	2,4,6-Tribromophenol	80.000	79.333	0.8	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.453	-3.6	130	0.00
44	2,4,5-Trichlorophenol	40.000	41.751	-4.4	131	0.00
45 S	2-Fluorobiphenyl	80.000	77.814	2.7	125	0.00
46	1,1'-Biphenyl	40.000	39.451	1.4	127	0.00
47	2-Chloronaphthalene	40.000	40.009	-0.0	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062823\
 Data File : BG057883.D
 Acq On : 28 Jun 2023 8:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 08:51:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG062723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 01:12:33 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.438	-3.6	127	0.00
49	Acenaphthylene	40.000	39.700	0.7	128	0.00
50	Dimethylphthalate	40.000	39.163	2.1	126	0.00
51	2,6-Dinitrotoluene	40.000	41.513	-3.8	128	0.00
52 C	Acenaphthene	40.000	39.573	1.1	126	0.00
53	3-Nitroaniline	40.000	40.469	-1.2	125	0.00
54 P	2,4-Dinitrophenol	40.000	36.712	8.2	113	0.00
55	Dibenzofuran	40.000	39.127	2.2	126	0.00
56 P	4-Nitrophenol	40.000	40.509	-1.3	119	0.00
57	2,4-Dinitrotoluene	40.000	40.619	-1.5	122	0.00
58	Fluorene	40.000	39.104	2.2	126	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.101	-0.3	127	0.00
60	Diethylphthalate	40.000	38.375	4.1	122	0.00
61	4-Chlorophenyl-phenylether	40.000	39.274	1.8	127	0.00
62	4-Nitroaniline	40.000	38.595	3.5	115	0.00
63	Azobenzene	40.000	39.451	1.4	126	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.948	0.1	113	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.189	-3.0	124	0.00
67	4-Bromophenyl-phenylether	40.000	40.881	-2.2	125	0.00
68	Hexachlorobenzene	40.000	40.335	-0.8	120	0.00
69	Atrazine	40.000	37.558	6.1	113	0.00
70 C	Pentachlorophenol	40.000	40.995	-2.5	117	0.00
71	Phenanthrene	40.000	39.836	0.4	117	0.00
72	Anthracene	40.000	39.485	1.3	116	0.00
73	Carbazole	40.000	38.772	3.1	110	0.00
74	Di-n-butylphthalate	40.000	39.017	2.5	111	0.00
75 C	Fluoranthene	40.000	38.199	4.5	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	40.042	-0.1	96	0.00
78	Pyrene	40.000	39.649	0.9	106	0.00
79 S	Terphenyl-d14	80.000	79.445	0.7	107	0.00
80	Butylbenzylphthalate	40.000	40.856	-2.1	106	0.00
81	Benzo(a)anthracene	40.000	40.338	-0.8	106	0.00
82	3,3'-Dichlorobenzidine	40.000	39.972	0.1	105	0.00
83	Chrysene	40.000	40.396	-1.0	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.167	-0.4	106	0.00
85 c	Di-n-octyl phthalate	40.000	40.805	-2.0	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.088	-0.2	106	-0.01
88	Benzo(b)fluoranthene	40.000	40.627	-1.6	108	0.00
89	Benzo(k)fluoranthene	40.000	40.038	-0.1	106	0.00
90 C	Benzo(a)pyrene	40.000	40.099	-0.2	106	-0.01
91	Dibenzo(a,h)anthracene	40.000	40.358	-0.9	108	0.00
92	Benzo(g,h,i)perylene	40.000	40.262	-0.7	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062823\
Data File : BG057883.D
Acq On : 28 Jun 2023 8:15
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 28 08:51:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG062723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 28 01:12:33 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0