

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062824\
 Data File : BG061777.D
 Acq On : 28 Jun 2024 18:22
 Operator : MA/JU
 Sample : P3005-10
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Jun 29 01:10:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 27 12:23:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.054	152	83780	20.000	ng/u1	0.00
20) Naphthalene-d8	10.868	136	392439	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.687	164	259298	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.425	188	576076	20.000	ng/u1	0.00
79) Chrysene-d12	21.691	240	441033	20.000	ng/u1	0.00
88) Perylene-d12	24.904	264	511377	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.571	96	92	0.040	ng/uL	0.12
4) Pyridine-d5	3.864	84	463	0.066	ng/u1	0.00
7) Phenol-d5	0.000	99	0	0.000	ng/u1	
9) Bis-(2-Chloroethyl)eth...	7.301	67	113	0.020	ng/u1	-0.08
11) 2-Chlorophenol-d4	0.000	132	0	0.000	ng/u1	
15) 4-Methylphenol-d8	0.000	113	0	0.000	ng/u1	
21) Nitrobenzene-d5	0.000	128	0	0.000	ng/u1	
24) 2-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
28) 2,4-Dichlorophenol-d3	0.000	165	0	0.000	ng/u1	
31) 4-Chloroaniline-d4	10.874	131	635	0.066	ng/u1	-0.12
46) Dimethylphthalate-d6	0.000	166	0	0.000	ng/u1	
49) Acenaphthylene-d8	0.000	160	0	0.000	ng/u1	
54) 4-Nitrophenol-d4	14.893	143	58	0.017	ng/u1	0.03
60) Fluorene-d10	15.686	176	398	0.023	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	0.000	200	0	0.000	ng/u1	
73) Anthracene-d10	17.425	188	576076	21.908	ng/u1	-0.11
81) Pyrene-d10	19.810	212	196	0.008	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.663	264	258	0.011	ng/u1	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062824\
Data File : BG061777.D
Acq On : 28 Jun 2024 18:22
Operator : MA/JU
Sample : P3005-10
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Jun 29 01:10:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 27 12:23:53 2024
Response via : Initial Calibration

