

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062824\
 Data File : BG061788.D
 Acq On : 29 Jun 2024 1:50
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020545

Quant Time: Jun 29 03:16:10 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 27 12:23:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.058	152	89912	20.000	ng/u1	0.00
20) Naphthalene-d8	10.866	136	421198	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.685	164	280562	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.429	188	626891	20.000	ng/u1	0.00
79) Chrysene-d12	21.689	240	485772	20.000	ng/u1	0.00
88) Perylene-d12	24.903	264	568561	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.457	96	18026	7.242	ng/uL	0.00
4) Pyridine-d5	3.875	84	137339	18.250	ng/u1	0.00
7) Phenol-d5	7.206	99	183124	18.761	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.382	67	109191	17.945	ng/u1	0.00
11) 2-Chlorophenol-d4	7.588	132	127354	19.013	ng/u1	0.00
15) 4-Methylphenol-d8	8.757	113	137836	18.267	ng/u1	0.00
21) Nitrobenzene-d5	9.221	128	62005	21.681	ng/u1	0.00
24) 2-Nitrophenol-d4	9.944	143	75789	26.165	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.479	165	141805	20.675	ng/u1	0.00
31) 4-Chloroaniline-d4	10.996	131	193547	18.639	ng/u1	0.00
46) Dimethylphthalate-d6	14.086	166	412305	19.097	ng/u1	0.00
49) Acenaphthylene-d8	14.380	160	468476	19.799	ng/u1	0.00
54) 4-Nitrophenol-d4	14.868	143	70797	19.093	ng/u1	0.00
60) Fluorene-d10	15.673	176	367311	19.433	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.784	200	56816	21.195	ng/u1	0.00
73) Anthracene-d10	17.523	188	568882	19.880	ng/u1	0.00
81) Pyrene-d10	19.809	212	589337	21.719	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.685	264	547324	20.155	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.493	88	19362	6.260	ng/uL#	71
5) Pyridine	3.892	79	142870	17.661	ng/u1	87
6) Benzaldehyde	7.194	77	106656	22.205	ng/u1#	86
8) Phenol	7.235	94	187097	18.215	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.476	93	136463	17.651	ng/u1#	95
12) 2-Chlorophenol	7.617	128	128275	19.089	ng/u1#	87
13) 2-Methylphenol	8.493	108	131162	18.251	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.581	45	302353	19.143	ng/u1#	95
16) Acetophenone	8.880	105	215223	17.322	ng/u1#	88
17) N-Nitroso-di-n-propyla...	8.863	70	113775	17.046	ng/u1	92
18) 4-Methylphenol	8.822	108	140611	17.465	ng/u1	95
19) Hexachloroethane	9.139	117	52342	19.426	ng/u1	93
22) Nitrobenzene	9.262	77	179664	21.807	ng/u1	97
23) Isophorone	9.785	82	338064	17.842	ng/u1	98
25) 2-Nitrophenol	9.973	139	74700	24.663	ng/u1#	78
26) 2,4-Dimethylphenol	10.026	107	169038	19.955	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.267	93	203938	19.027	ng/u1#	96
29) 2,4-Dichlorophenol	10.508	162	129955	20.680	ng/u1	98
30) Naphthalene	10.919	128	455187	19.314	ng/u1	97
32) 4-Chloroaniline	11.025	127	184458	17.911	ng/u1#	86
33) Hexachlorobutadiene	11.195	225	92887	21.399	ng/u1	88
34) Caprolactam	11.783	113	35033	13.482	ng/u1	85
35) 4-Chloro-3-methylphenol	12.135	107	161486	18.377	ng/u1	98
36) 2-Methylnaphthalene	12.517	142	311436	18.728	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062824\
 Data File : BG061788.D
 Acq On : 29 Jun 2024 1:50
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020545

Quant Time: Jun 29 03:16:10 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 27 12:23:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.735	142	323801	18.249	ng/ul	91
39) 1,2,4,5-Tetrachloroben...	12.882	216	170671	21.495	ng/ul	94
40) Hexachlorocyclopentadiene	12.858	237	75205	15.919	ng/ul	97
41) 2,4,6-Trichlorophenol	13.117	196	109367	21.349	ng/ul	89
42) 2,4,5-Trichlorophenol	13.187	196	120492	21.579	ng/ul	91
43) 1,1'-Biphenyl	13.522	154	421590	20.744	ng/ul	99
44) 2-Chloronaphthalene	13.563	162	320169	20.509	ng/ul	99
45) 2-Nitroaniline	13.763	65	126839	23.734	ng/ul	100
47) Dimethylphthalate	14.133	163	401230	18.714	ng/ul	96
48) 2,6-Dinitrotoluene	14.257	165	86749	24.383	ng/ul#	93
50) Acenaphthylene	14.409	152	517082	19.577	ng/ul	96
51) 3-Nitroaniline	14.586	138	89658	24.609	ng/ul#	91
52) Acenaphthene	14.750	153	349347	19.099	ng/ul	93
53) 2,4-Dinitrophenol	14.791	184	37702	20.854	ng/ul#	86
55) 4-Nitrophenol	14.885	109	71174	18.587	ng/ul	97
56) Dibenzofuran	15.085	168	492476	19.240	ng/ul	92
57) 2,4-Dinitrotoluene	15.044	165	122608	23.217	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.302	232	106857	19.517	ng/ul#	97
59) Diethylphthalate	15.496	149	417755	18.919	ng/ul	98
61) Fluorene	15.731	166	409447	19.329	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.725	204	207524	19.734	ng/ul	95
63) 4-Nitroaniline	15.743	138	91224	23.635	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.802	198	59064	19.452	ng/ul#	91
67) N-Nitrosodiphenylamine	15.937	169	339111	20.918	ng/ul	97
68) 4-Bromophenyl-phenylether	16.618	248	140316	21.739	ng/ul	92
69) Hexachlorobenzene	16.724	284	164411	20.989	ng/ul	96
70) Atrazine	16.883	200	134361	20.750	ng/ul	94
71) Pentachlorophenol	17.071	266	85049	17.824	ng/ul	93
72) Phenanthrene	17.470	178	643172	19.502	ng/ul	100
74) Anthracene	17.564	178	659081	19.542	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.487	216	171676	24.580	ng/ul	97
76) Pentachlorobenzene	14.997	250	184705	22.022	ng/ul	97
77) Carbazole	17.829	167	555439	19.396	ng/ul	98
78) Di-n-butylphthalate	18.387	149	714550	20.863	ng/ul	98
80) Fluoranthene	19.474	202	695613	22.094	ng/ul	99
82) Pyrene	19.838	202	716000	21.388	ng/ul	95
83) Butylbenzylphthalate	20.720	149	298117	24.458	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.583	252	213266	20.763	ng/ul	93
85) Benzo(a)anthracene	21.671	228	683462	20.044	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.577	149	426750	23.819	ng/ul	97
87) Chrysene	21.736	228	642897	19.925	ng/ul	99
89) Di-n-octyl phthalate	22.799	149	715812	23.870	ng/ul	100
90) Benzo(b)fluoranthene	23.881	252	686224	20.078	ng/ul	99
91) Benzo(k)fluoranthene	23.951	252	681383	19.385	ng/ul	99
93) Benzo(a)pyrene	24.756	252	642542	19.528	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	28.569	276	834872	20.362	ng/ul	97
95) Dibenzo(a,h)anthracene	28.640	278	697671	20.463	ng/ul	99
96) Benzo(g,h,i)perylene	29.709	276	456883	13.823	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 6.260 ng/uL
 RT: 3.493 min Scan# 31
 Delta R.T. 0.005 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020545


Tgt Ion: 88 Resp: 19362
 Ion Ratio Lower Upper
 88 100
 43 51.9 24.1 36.1#
 58 100.6 63.0 94.6#



#5
 Pyridine
 Concen: 17.661 ng/uL
 RT: 3.892 min Scan# 103
 Delta R.T. -0.001 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50

Tgt Ion: 79 Resp: 142870
 Ion Ratio Lower Upper
 79 100
 52 124.5 84.4 126.6
 51 55.1 43.0 64.6









#6
Benzaldehyde
Concen: 22.205 ng/ul
RT: 7.194 min Scan# 664
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion: 77 Resp: 106656
Ion Ratio Lower Upper
77 100
105 97.6 71.0 106.6
106 95.9 63.5 95.3#



#8
Phenol
Concen: 18.215 ng/ul
RT: 7.235 min Scan# 672
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion: 94 Resp: 187097
Ion Ratio Lower Upper
94 100
65 32.4 25.8 38.6
66 51.0 36.9 55.3









#10
Bis(2-Chloroethyl)ether
Concen: 17.651 ng/ul
RT: 7.476 min Scan# 712
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion: 93 Resp: 136463
Ion Ratio Lower Upper
93 100
63 88.3 72.5 108.7
95 33.3 21.8 32.8#



#12
2-Chlorophenol
Concen: 19.089 ng/ul
RT: 7.617 min Scan# 737
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:128 Resp: 128275
Ion Ratio Lower Upper
128 100
64 65.0 43.2 64.8#
130 35.0 24.1 36.1









#13
2-Methylphenol
Concen: 18.251 ng/ul
RT: 8.493 min Scan# 885
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:108 Resp: 131162
Ion Ratio Lower Upper
108 100
107 89.3 69.0 103.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.143 ng/ul
RT: 8.581 min Scan# 901
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion: 45 Resp: 302353
Ion Ratio Lower Upper
45 100
77 8.2 8.9 13.3#
79 7.3 5.8 8.8









#16
Acetophenone
Concen: 17.322 ng/ul
RT: 8.880 min Scan# 91
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:105 Resp: 215223
Ion Ratio Lower Upper
105 100
77 84.9 60.4 90.6
51 46.7 31.1 46.7#



#17
N-Nitroso-di-n-propylamine
Concen: 17.046 ng/ul
RT: 8.863 min Scan# 949
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion: 70 Resp: 113775
Ion Ratio Lower Upper
70 100
42 85.9 62.7 94.1
101 8.7 5.8 8.8
130 19.0 13.3 19.9









#18
4-Methylphenol
Concen: 17.465 ng/ul
RT: 8.822 min Scan# 941
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:108 Resp: 140611
Ion Ratio Lower Upper
108 100
107 107.4 90.2 135.2



#19
Hexachloroethane
Concen: 19.426 ng/ul
RT: 9.139 min Scan# 996
Delta R.T. -0.007 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:117 Resp: 52342
Ion Ratio Lower Upper
117 100
201 99.1 79.3 118.9
199 63.4 61.9 92.9









#22
Nitrobenzene
Concen: 21.807 ng/ul
RT: 9.262 min Scan# 1017
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion: 77 Resp: 179664
Ion Ratio Lower Upper
77 100
123 34.2 29.0 43.6
65 14.4 11.0 16.6



#23
Isophorone
Concen: 17.842 ng/ul
RT: 9.785 min Scan# 1106
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion: 82 Resp: 338064
Ion Ratio Lower Upper
82 100
95 7.8 6.4 9.6
138 15.5 11.7 17.5









#25
2-Nitrophenol
Concen: 24.663 ng/ul
RT: 9.973 min Scan# 1138
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:139 Resp: 74700
Ion Ratio Lower Upper
139 100
65 73.3 42.2 63.4#
109 40.7 27.0 40.4#



#26
2,4-Dimethylphenol
Concen: 19.955 ng/ul
RT: 10.026 min Scan# 1147
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:107 Resp: 169038
Ion Ratio Lower Upper
107 100
121 44.3 34.5 51.7
122 73.7 60.3 90.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.027 ng/ul
 RT: 10.267 min Scan# 1188
 Delta R.T. -0.007 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020545


Tgt Ion: 93 Resp: 203938
 Ion Ratio Lower Upper
 93 100
 95 32.3 27.4 41.2
 123 10.3 6.7 10.1#



#29
 2,4-Dichlorophenol
 Concen: 20.680 ng/ul
 RT: 10.508 min Scan# 1229
 Delta R.T. -0.001 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50

Tgt Ion:162 Resp: 129955
 Ion Ratio Lower Upper
 162 100
 164 65.9 51.7 77.5
 98 41.3 34.9 52.3









#30
Naphthalene
Concen: 19.314 ng/ul
RT: 10.919 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:128 Resp: 455187
Ion Ratio Lower Upper
128 100
129 10.0 9.0 13.6
127 14.2 10.5 15.7



#32
4-Chloroaniline
Concen: 17.911 ng/ul
RT: 11.025 min Scan# 1317
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:127 Resp: 184458
Ion Ratio Lower Upper
127 100
129 35.9 22.9 34.3#









#33
Hexachlorobutadiene
Concen: 21.399 ng/ul
RT: 11.195 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:225 Resp: 92887
Ion Ratio Lower Upper
225 100
223 70.7 49.4 74.0
227 66.2 45.6 68.4



#34
Caprolactam
Concen: 13.482 ng/ul
RT: 11.783 min Scan# 1446
Delta R.T. 0.011 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:113 Resp: 35033
Ion Ratio Lower Upper
113 100
55 295.4 207.4 311.2
56 194.6 167.2 250.8









#35
4-Chloro-3-methylphenol
Concen: 18.377 ng/ul
RT: 12.135 min Scan# 1506
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion	Ratio	Lower	Upper
107	100		
144	22.3	19.6	29.4
142	74.8	60.2	90.4



#36
2-Methylnaphthalene
Concen: 18.728 ng/ul
RT: 12.517 min Scan# 1571
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion	Ratio	Lower	Upper
142	100		
141	84.6	68.5	102.7





#37

1-Methylnaphthalene

Concen: 18.249 ng/ul

RT: 12.735 min Scan# 1

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:142 Resp: 323801

Ion Ratio Lower Upper

142 100

141 89.2 64.6 97.0

116 4.3 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.495 ng/ul

RT: 12.882 min Scan# 1633

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Tgt Ion:216 Resp: 170671




Ion Ratio Lower Upper


216 100

214 83.1 61.4 92.2

179 23.2 17.1 25.7

108 17.8 14.3 21.5





#40

Hexachlorocyclopentadiene

Concen: 15.919 ng/ul

RT: 12.858 min Scan# 1

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:237 Resp: 75205

Ion Ratio Lower Upper

237 100

235 66.2 51.0 76.6

272 11.3 9.1 13.7



#41

2,4,6-Trichlorophenol

Concen: 21.349 ng/ul

RT: 13.117 min Scan# 1673

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50




Tgt Ion:196 Resp: 109367


Ion Ratio Lower Upper

196 100

198 104.3 73.1 109.7

200 36.2 26.4 39.6









#42
2,4,5-Trichlorophenol
Concen: 21.579 ng/ul
RT: 13.187 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:196 Resp: 120492
Ion Ratio Lower Upper
196 100
198 102.1 72.7 109.1
200 33.7 27.0 40.6



#43
1,1'-Biphenyl
Concen: 20.744 ng/ul
RT: 13.522 min Scan# 1742
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:154 Resp: 421590
Ion Ratio Lower Upper
154 100
153 38.6 31.4 47.0
76 13.3 11.0 16.4









#44
2-Chloronaphthalene
Concen: 20.509 ng/ul
RT: 13.563 min Scan# 1749
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:162 Resp: 320169
Ion Ratio Lower Upper
162 100
164 30.7 25.1 37.7
127 41.8 33.7 50.5



#45
2-Nitroaniline
Concen: 23.734 ng/ul
RT: 13.763 min Scan# 1783
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion: 65 Resp: 126839
Ion Ratio Lower Upper
65 100
92 64.1 51.4 77.0
138 78.7 63.0 94.6









#47
Dimethylphthalate
Concen: 18.714 ng/ul
RT: 14.133 min Scan# 1846
Delta R.T. -0.007 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:163 Resp: 401230
Ion Ratio Lower Upper
163 100
194 3.6 2.5 3.7
164 11.1 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 24.383 ng/ul
RT: 14.257 min Scan# 1867
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:165 Resp: 86749
Ion Ratio Lower Upper
165 100
89 67.3 55.9 83.9
121 21.2 23.4 35.2#









#50
Acenaphthylene
Concen: 19.577 ng/ul
RT: 14.409 min Scan# 1893
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:152 Resp: 517082
Ion Ratio Lower Upper
152 100
151 21.6 15.3 22.9
153 14.4 11.0 16.4



#51
3-Nitroaniline
Concen: 24.609 ng/ul
RT: 14.586 min Scan# 1923
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:138 Resp: 89658
Ion Ratio Lower Upper
138 100
108 11.6 14.3 21.5#
92 133.0 115.0 172.4









#52
Acenaphthene
Concen: 19.099 ng/ul
RT: 14.750 min Scan# 1951
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:153 Resp: 349347
Ion Ratio Lower Upper
153 100
152 44.6 40.0 60.0
154 85.9 74.1 111.1



#53
2,4-Dinitrophenol
Concen: 20.854 ng/ul
RT: 14.791 min Scan# 1958
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:184 Resp: 37702
Ion Ratio Lower Upper
184 100
63 86.1 74.2 111.4
154 80.4 80.4 120.6#









#55
4-Nitrophenol
Concen: 18.587 ng/ul
RT: 14.885 min Scan# 1973
Delta R.T. 0.005 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion	109	Resp	71174
Ion Ratio	100		
Lower			
Upper			
109	100		
139	95.1	77.1	115.7
65	133.3	110.9	166.3



#56
Dibenzofuran
Concen: 19.240 ng/ul
RT: 15.085 min Scan# 2008
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion	168	Resp	492476
Ion Ratio <td>100</td> <td></td> <td></td>	100		
Lower			
Upper			
168	100		
139	38.5	27.0	40.4





#57

2,4-Dinitrotoluene

Concen: 23.217 ng/ul

RT: 15.044 min Scan# 2000

Delta R.T. 0.005 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:165 Resp: 122608

Ion Ratio Lower Upper

165 100

63 55.9 49.4 74.0

182 1.8 1.6 2.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.517 ng/ul

RT: 15.302 min Scan# 2045

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Tgt Ion:232 Resp: 106857




Ion Ratio Lower Upper


232 100

131 43.7 33.1 49.7

130 2.1 2.6 3.8#

166 29.6 23.4 35.0









#59
Diethylphthalate
Concen: 18.919 ng/ul
RT: 15.496 min Scan# 2078
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:149 Resp: 417755
Ion Ratio Lower Upper
149 100
177 23.8 19.6 29.4
150 13.4 9.2 13.8



#61
Fluorene
Concen: 19.329 ng/ul
RT: 15.731 min Scan# 2118
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:166 Resp: 409447
Ion Ratio Lower Upper
166 100
165 100.6 84.9 127.3
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 19.734 ng/ul
RT: 15.725 min Scan# 2116
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:204 Resp: 207524
Ion Ratio Lower Upper
204 100
206 31.1 26.4 39.6
141 59.8 51.1 76.7



#63
4-Nitroaniline
Concen: 23.635 ng/ul
RT: 15.743 min Scan# 2120
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:138 Resp: 91224
Ion Ratio Lower Upper
138 100
92 57.9 54.8 82.2
108 107.7 84.2 126.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.452 ng/ul
 RT: 15.802 min Scan# 2130
 Delta R.T. 0.005 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020545


Tgt Ion:198 Resp: 59064
 Ion Ratio Lower Upper
 198 100
 182 4.7 3.6 5.4
 77 29.5 19.5 29.3#



#67
 N-Nitrosodiphenylamine
 Concen: 20.918 ng/ul
 RT: 15.937 min Scan# 2153
 Delta R.T. -0.001 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50

Tgt Ion:169 Resp: 339111
 Ion Ratio Lower Upper
 169 100
 168 69.4 54.6 81.8
 167 37.5 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 21.739 ng/ul
RT: 16.618 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:248 Resp: 140316
Ion Ratio Lower Upper
248 100
250 93.1 82.1 123.1
141 66.1 57.6 86.4



#69
Hexachlorobenzene
Concen: 20.989 ng/ul
RT: 16.724 min Scan# 2287
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:284 Resp: 164411
Ion Ratio Lower Upper
284 100
142 29.7 25.2 37.8
249 31.0 26.9 40.3





#70

Atrazine

Concen: 20.750 ng/ul

RT: 16.883 min Scan# 21

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:200 Resp: 134361

Ion Ratio Lower Upper

200 100

173 24.0 18.2 27.2

215 40.8 36.6 54.8



#71

Pentachlorophenol

Concen: 17.824 ng/ul

RT: 17.071 min Scan# 2346

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50




Tgt Ion:266 Resp: 85049


Ion Ratio Lower Upper

266 100

264 67.5 52.5 78.7

268 68.0 47.5 71.3









#72
Phenanthrene
Concen: 19.502 ng/u1
RT: 17.470 min Scan# 2414
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:178 Resp: 643172
Ion Ratio Lower Upper
178 100
179 14.9 12.0 18.0
176 20.3 16.2 24.2



#74
Anthracene
Concen: 19.542 ng/u1
RT: 17.564 min Scan# 2430
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:178 Resp: 659081
Ion Ratio Lower Upper
178 100
179 15.2 11.4 17.2
176 19.4 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 24.580 ng/ul

RT: 13.487 min Scan# 1735

Delta R.T. 0.005 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:216 Resp: 171676

Ion Ratio Lower Upper

216 100

214 75.7 61.9 92.9

218 45.6 39.6 59.4



#76

Pentachlorobenzene

Concen: 22.022 ng/ul

RT: 14.997 min Scan# 1993

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50




Tgt Ion:250 Resp: 184705


Ion Ratio Lower Upper

250 100

248 61.3 45.7 68.5

252 60.2 48.3 72.5









#77
 Carbazole
 Concen: 19.396 ng/ul
 RT: 17.829 min Scan# 2475
 Delta R.T. -0.001 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020545


Tgt Ion:167 Resp: 555439
 Ion Ratio Lower Upper
 167 100
 166 22.4 17.5 26.3
 139 12.8 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 20.863 ng/ul
 RT: 18.387 min Scan# 2570
 Delta R.T. -0.007 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50

Tgt Ion:149 Resp: 714550
 Ion Ratio Lower Upper
 149 100
 150 9.3 6.7 10.1
 104 5.6 4.6 6.8









#80
Fluoranthene
Concen: 22.094 ng/u1
RT: 19.474 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50




Instrument :
BNA_G
ClientSampleId :
SSTD020545


Tgt Ion:202 Resp: 695613
Ion Ratio Lower Upper
202 100
101 9.4 7.8 11.6
100 9.1 7.7 11.5



#82
Pyrene
Concen: 21.388 ng/u1
RT: 19.838 min Scan# 2817
Delta R.T. -0.001 min
Lab File: BG061788.D
Acq: 29 Jun 2024 1:50

Tgt Ion:202 Resp: 716000
Ion Ratio Lower Upper
202 100
101 10.6 10.1 15.1
100 10.4 9.9 14.9





#83

Butylbenzylphthalate

Concen: 24.458 ng/ul

RT: 20.720 min Scan# 2967

Delta R.T. -0.007 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:149 Resp: 298117

Ion Ratio Lower Upper

149 100

91 80.5 68.1 102.1

206 19.2 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 20.763 ng/ul

RT: 21.583 min Scan# 3114

Delta R.T. -0.007 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50




Tgt Ion:252 Resp: 213266


Ion Ratio Lower Upper

252 100

254 60.6 53.9 80.9

126 9.9 8.1 12.1









#85
 Benzo(a)anthracene
 Concen: 20.044 ng/ul
 RT: 21.671 min Scan# 3129
 Delta R.T. -0.001 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020545


Tgt Ion:228 Resp: 683462
 Ion Ratio Lower Upper
 228 100
 229 20.5 15.8 23.8
 226 26.8 21.4 32.2



#86
 Bis(2-ethylhexyl)phthalate
 Concen: 23.819 ng/ul
 RT: 21.577 min Scan# 3113
 Delta R.T. -0.013 min
 Lab File: BG061788.D
 Acq: 29 Jun 2024 1:50

Tgt Ion:149 Resp: 426750
 Ion Ratio Lower Upper
 149 100
 167 27.0 20.5 30.7
 279 4.3 2.9 4.3





#87

Chrysene

Concen: 19.925 ng/ul

RT: 21.736 min Scan# 3140

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:228 Resp: 642897

Ion Ratio Lower Upper

228 100

226 30.6 24.2 36.2

229 19.8 16.5 24.7



#89

Di-n-octyl phthalate

Concen: 23.870 ng/ul




RT: 22.799 min Scan# 3321


Delta R.T. -0.007 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Tgt Ion:149 Resp: 715812





#90

Benzo(b)fluoranthene

Concen: 20.078 ng/ul

RT: 23.881 min Scan# 3505

Delta R.T. -0.007 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:252 Resp: 686224

Ion Ratio Lower Upper

252 100

253 22.5 18.6 27.8

125 8.7 7.0 10.6



#91

Benzo(k)fluoranthene

Concen: 19.385 ng/ul

RT: 23.951 min Scan# 3517

Delta R.T. -0.007 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50




Tgt Ion:252 Resp: 681383


Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 8.2 6.2 9.2





#93

Benzo(a)pyrene

Concen: 19.528 ng/ul

RT: 24.756 min Scan# 3

Delta R.T. -0.007 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:252 Resp: 642542

Ion Ratio Lower Upper

252 100

253 23.3 16.8 25.2

125 8.7 6.5 9.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.362 ng/ul

RT: 28.569 min Scan# 4303

Delta R.T. 0.005 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50




Tgt Ion:276 Resp: 834872


Ion Ratio Lower Upper

276 100

138 14.3 12.2 18.4

277 22.6 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 20.463 ng/ul

RT: 28.640 min Scan# 41

Delta R.T. 0.005 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Instrument :

BNA_G

ClientSampleId :

SSTD020545





Tgt Ion:278 Resp: 697671

Ion Ratio Lower Upper

278 100

139 11.7 9.1 13.7

279 23.6 18.6 28.0



#96

Benzo(g,h,i)perylene

Concen: 13.823 ng/ul

RT: 29.709 min Scan# 4497

Delta R.T. -0.001 min

Lab File: BG061788.D

Acq: 29 Jun 2024 1:50

Tgt Ion:276 Resp: 456883

Ion Ratio Lower Upper

276 100

138 17.2 12.2 18.4

277 25.2 18.6 27.8

