

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062922\
 Data File : BG054153.D
 Acq On : 29 Jun 2022 20:15
 Operator : CG/JU
 Sample : N3525-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MH-725

Quant Time: Jun 29 21:00:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.070	152	23227	20.000	ng	0.00
21) Naphthalene-d8	10.879	136	95964	20.000	ng	0.00
39) Acenaphthene-d10	14.692	164	72799	20.000	ng	0.00
64) Phenanthrene-d10	17.436	188	183226	20.000	ng	0.00
76) Chrysene-d12	21.713	240	192501	20.000	ng	0.00
86) Perylene-d12	24.968	264	195373	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.638	112	77921	64.140	ng	0.00
7) Phenol-d6	7.230	99	61916	37.569	ng	0.00
23) Nitrobenzene-d5	9.228	82	155952	91.325	ng	0.00
42) 2,4,6-Tribromophenol	16.178	330	160633	137.335	ng	0.00
45) 2-Fluorobiphenyl	13.317	172	435504	86.990	ng	0.00
79) Terphenyl-d14	20.044	244	884694	92.563	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062922\
Data File : BG054153.D
Acq On : 29 Jun 2022 20:15
Operator : CG/JU
Sample : N3525-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MH-725

Quant Time: Jun 29 21:00:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

