

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062922\
 Data File : BG054155.D
 Acq On : 29 Jun 2022 21:35
 Operator : CG/JU
 Sample : N3528-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CONCRETE-BLOCK

Quant Time: Jun 29 22:07:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.065	152	21616	20.000	ng	0.00
21) Naphthalene-d8	10.880	136	88863	20.000	ng	0.00
39) Acenaphthene-d10	14.693	164	67547	20.000	ng	0.00
64) Phenanthrene-d10	17.437	188	169383	20.000	ng	# 0.00
76) Chrysene-d12	21.714	240	183798	20.000	ng	0.00
86) Perylene-d12	24.969	264	187686	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.639	112	49649	43.914	ng	0.00
7) Phenol-d6	7.231	99	141082	91.986	ng	0.00
23) Nitrobenzene-d5	9.229	82	118703	75.067	ng	0.00
42) 2,4,6-Tribromophenol	16.179	330	16149	14.880	ng	0.00
45) 2-Fluorobiphenyl	13.318	172	344490	74.161	ng	0.00
79) Terphenyl-d14	20.045	244	704199	77.167	ng	0.00
Target Compounds						Qvalue
71) Phenanthrene	17.478	178	48399	5.562	ng	98
75) Fluoranthene	19.487	202	40049	3.506	ng	98
78) Pyrene	19.846	202	30271	2.727	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062922\
Data File : BG054155.D
Acq On : 29 Jun 2022 21:35
Operator : CG/JU
Sample : N3528-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
CONCRETE-BLOCK

Quant Time: Jun 29 22:07:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

