

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062922\
 Data File : BG054147.D
 Acq On : 29 Jun 2022 16:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 29 17:20:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.457	0.422	7.7	88	0.00
3	Pyridine	1.220	1.180	3.3	89	0.00
4	n-Nitrosodimethylamine	0.537	0.563	-4.8	95	0.00
5 S	2-Fluorophenol	1.046	1.046	0.0	94	0.00
6	Aniline	1.682	1.793	-6.6	100	0.00
7 S	Phenol-d6	1.419	1.531	-7.9	102	0.00
8	2-Chlorophenol	1.117	1.206	-8.0	102	0.00
9	Benzaldehyde	0.838	0.791	5.6	98	0.00
10 C	Phenol	1.438	1.554	-8.1	100	0.00
11	bis(2-Chloroethyl)ether	1.104	1.178	-6.7	103	0.00
12	1,3-Dichlorobenzene	1.400	1.396	0.3	97	0.00
13 C	1,4-Dichlorobenzene	1.406	1.412	-0.4	99	0.00
14	1,2-Dichlorobenzene	1.350	1.355	-0.4	96	0.00
15	Benzyl Alcohol	1.061	1.220	-15.0	106	0.00
16	2,2'-oxybis(1-Chloropropane	1.650	1.777	-7.7	103	0.00
17	2-Methylphenol	0.999	1.102	-10.3	105	0.00
18	Hexachloroethane	0.466	0.490	-5.2	101	0.00
19 P	n-Nitroso-di-n-propylamine	0.953	1.071	-12.4	106	0.00
20	3+4-Methylphenols	1.401	1.605	-14.6	108	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.486	0.474	2.5	104	0.00
23 S	Nitrobenzene-d5	0.356	0.362	-1.7	107	0.00
24	Nitrobenzene	0.350	0.360	-2.9	108	0.00
25	Isophorone	0.665	0.665	0.0	107	0.00
26 C	2-Nitrophenol	0.155	0.178	-14.8	118	0.00
27	2,4-Dimethylphenol	0.251	0.258	-2.8	109	0.00
28	bis(2-Chloroethoxy)methane	0.358	0.364	-1.7	109	0.00
29 C	2,4-Dichlorophenol	0.336	0.367	-9.2	115	0.00
30	1,2,4-Trichlorobenzene	0.416	0.393	5.5	100	0.00
31	Naphthalene	1.010	0.994	1.6	105	0.00
32	Benzoic acid	0.087	0.150	-72.4#	176#	-0.02
33	4-Chloroaniline	0.419	0.440	-5.0	110	0.00
34 C	Hexachlorobutadiene	0.320	0.298	6.9	98	0.00
35	Caprolactam	0.092	0.108	-17.4	117	0.01
36 C	4-Chloro-3-methylphenol	0.327	0.372	-13.8	118	0.00
37	2-Methylnaphthalene	0.747	0.775	-3.7	110	0.00
38	1-Methylnaphthalene	0.727	0.752	-3.4	111	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	0.771	0.709	8.0	106	0.00
41 P	Hexachlorocyclopentadiene	0.359	0.365	-1.7	114	0.00
42 S	2,4,6-Tribromophenol	0.321	0.327	-1.9	116	0.00
43 C	2,4,6-Trichlorophenol	0.433	0.450	-3.9	117	0.00
44	2,4,5-Trichlorophenol	0.483	0.503	-4.1	118	0.00
45 S	2-Fluorobiphenyl	1.375	1.310	4.7	111	0.00
46	1,1'-Biphenyl	1.393	1.338	3.9	111	0.00
47	2-Chloronaphthalene	1.127	1.089	3.4	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062922\
 Data File : BG054147.D
 Acq On : 29 Jun 2022 16:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 29 17:20:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.300	0.333	-11.0	123	0.00
49	Acenaphthylene	1.753	1.724	1.7	113	0.00
50	Dimethylphthalate	1.523	1.497	1.7	114	0.00
51	2,6-Dinitrotoluene	0.308	0.329	-6.8	120	0.00
52 C	Acenaphthene	1.118	1.132	-1.3	117	0.00
53	3-Nitroaniline	0.298	0.311	-4.4	116	0.00
54 P	2,4-Dinitrophenol	0.140	0.167	-19.3	136	0.00
55	Dibenzofuran	1.801	1.782	1.1	115	0.00
56 P	4-Nitrophenol	0.212	0.217	-2.4	114	0.00
57	2,4-Dinitrotoluene	0.435	0.478	-9.9	123	0.00
58	Fluorene	1.478	1.473	0.3	116	0.00
59	2,3,4,6-Tetrachlorophenol	0.473	0.480	-1.5	115	0.00
60	Diethylphthalate	1.475	1.502	-1.8	119	0.00
61	4-Chlorophenyl-phenylether	0.885	0.872	1.5	115	0.00
62	4-Nitroaniline	0.311	0.322	-3.5	118	0.00
63	Azobenzene	1.187	1.225	-3.2	121	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	0.100	0.114	-14.0	129	0.00
66 c	n-Nitrosodiphenylamine	0.521	0.531	-1.9	117	0.00
67	4-Bromophenyl-phenylether	0.250	0.245	2.0	114	0.00
68	Hexachlorobenzene	0.286	0.272	4.9	111	0.00
69	Atrazine	0.216	0.215	0.5	112	0.00
70 C	Pentachlorophenol	0.133	0.138	-3.8	117	0.00
71	Phenanthrene	1.028	1.010	1.8	114	0.00
72	Anthracene	1.003	0.993	1.0	115	0.00
73	Carbazole	0.861	0.846	1.7	114	0.00
74	Di-n-butylphthalate	0.996	1.032	-3.6	119	0.00
75 C	Fluoranthene	1.349	1.237	8.3	107	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
77	Benzidine	0.414	0.391	5.6	96	0.00
78	Pyrene	1.208	1.265	-4.7	108	0.00
79 S	Terphenyl-d14	0.993	1.005	-1.2	104	0.00
80	Butylbenzylphthalate	0.387	0.438	-13.2	115	0.00
81	Benzo(a)anthracene	1.295	1.277	1.4	103	0.00
82	3,3'-Dichlorobenzidine	0.387	0.387	0.0	102	0.00
83	Chrysene	1.234	1.207	2.2	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.554	0.617	-11.4	113	0.00
85 c	Di-n-octyl phthalate	0.952	1.061	-11.4	113	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.01
87	Indeno(1,2,3-cd)pyrene	1.509	1.459	3.3	98	0.02
88	Benzo(b)fluoranthene	1.216	1.190	2.1	99	0.00
89	Benzo(k)fluoranthene	1.204	1.165	3.2	101	0.00
90 C	Benzo(a)pyrene	1.026	1.008	1.8	100	0.00
91	Dibenzo(a,h)anthracene	1.248	1.201	3.8	98	0.02
92	Benzo(g,h,i)perylene	1.232	1.161	5.8	96	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062922\
Data File : BG054147.D
Acq On : 29 Jun 2022 16:16
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 29 17:20:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0