

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070123\
 Data File : BG057937.D
 Acq On : 1 Jul 2023 11:10
 Operator : CG/JU
 Sample : SSTD00586
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005435

Quant Time: Jul 02 07:09:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 07:09:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.926	152	43883	20.000	ng/u1	0.00
20) Naphthalene-d8	10.735	136	198953	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.565	164	143114	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.309	188	344388	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	292145	20.000	ng/u1	0.00
88) Perylene-d12	24.712	264	339930	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.367	96	2392	1.909	ng/uL	0.00
4) Pyridine-d5	3.778	84	18169	4.776	ng/u1	0.00
7) Phenol-d5	7.092	99	18410	4.108	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.256	67	12085	4.475	ng/u1	0.00
11) 2-Chlorophenol-d4	7.462	132	13863	4.485	ng/u1	0.00
15) 4-Methylphenol-d8	8.631	113	14257	4.216	ng/u1	0.00
21) Nitrobenzene-d5	9.089	128	6802	4.195	ng/u1	0.00
24) 2-Nitrophenol-d4	9.812	143	6449	3.784	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.353	165	14001	4.283	ng/u1	0.00
31) 4-Chloroaniline-d4	10.870	131	21058	4.248	ng/u1	0.00
46) Dimethylphthalate-d6	13.966	166	52022	4.607	ng/u1	0.00
49) Acenaphthylene-d8	14.260	160	56435	4.580	ng/u1	0.00
54) 4-Nitrophenol-d4	14.759	143	7187	3.603	ng/u1	0.00
60) Fluorene-d10	15.552	176	44600	4.651	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.664	200	5651	3.022	ng/u1	0.00
73) Anthracene-d10	17.403	188	75759	4.741	ng/u1	0.00
81) Pyrene-d10	19.695	212	88456	4.808	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.495	264	80014	4.620	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.402	88	3092	1.997	ng/uL	93
5) Pyridine	3.796	79	17838	4.576	ng/u1	94
6) Benzaldehyde	7.068	77	5141	3.435	ng/u1#	82
8) Phenol	7.115	94	19617	4.314	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.344	93	15631	4.451	ng/u1	97
12) 2-Chlorophenol	7.491	128	14585	4.534	ng/u1	94
13) 2-Methylphenol	8.373	108	15111	4.231	ng/u1	85
14) 2,2'-oxybis(1-Chloropr...	8.461	45	24274	4.491	ng/u1	96
16) Acetophenone	8.749	105	26065	4.646	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.731	70	13617	4.540	ng/u1	93
18) 4-Methylphenol	8.696	108	16534	4.281	ng/u1	98
19) Hexachloroethane	9.013	117	5643	4.576	ng/u1	91
22) Nitrobenzene	9.136	77	18665	4.434	ng/u1	91
23) Isophorone	9.648	82	41401	4.609	ng/u1	97
25) 2-Nitrophenol	9.841	139	7479	4.090	ng/u1	97
26) 2,4-Dimethylphenol	9.900	107	18399	4.474	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.135	93	23315	4.606	ng/u1	94
29) 2,4-Dichlorophenol	10.382	162	13908	4.375	ng/u1	97
30) Naphthalene	10.782	128	51611	4.616	ng/u1	100
32) 4-Chloroaniline	10.893	127	21934	4.328	ng/u1	97
33) Hexachlorobutadiene	11.064	225	10331	4.591	ng/u1	93
34) Caprolactam	11.633	113	5212	4.101	ng/u1	96
35) 4-Chloro-3-methylphenol	12.015	107	16993	4.352	ng/u1	99
36) 2-Methylnaphthalene	12.391	142	34757	4.612	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070123\
 Data File : BG057937.D
 Acq On : 1 Jul 2023 11:10
 Operator : CG/JU
 Sample : SST00586
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST005435


Quant Time: Jul 02 07:09:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 07:09:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.609	142	34937	4.623	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.756	216	19742	4.527	ng/ul	98
40) Hexachlorocyclopentadiene	12.738	237	9429	3.432	ng/ul	95
41) 2,4,6-Trichlorophenol	12.997	196	12414	4.200	ng/ul	94
42) 2,4,5-Trichlorophenol	13.073	196	13343	4.174	ng/ul	95
43) 1,1'-Biphenyl	13.396	154	47916	4.646	ng/ul	98
44) 2-Chloronaphthalene	13.437	162	38406	4.640	ng/ul	94
45) 2-Nitroaniline	13.643	65	11851	4.118	ng/ul	89
47) Dimethylphthalate	14.013	163	53729	4.736	ng/ul	99
48) 2,6-Dinitrotoluene	14.136	165	9734	4.180	ng/ul	95
50) Acenaphthylene	14.283	152	61540	4.625	ng/ul	99
51) 3-Nitroaniline	14.465	138	8476	4.307	ng/ul	99
52) Acenaphthene	14.630	153	42479	4.652	ng/ul	100
53) 2,4-Dinitrophenol	14.671	184	2629	2.166	ng/ul#	81
55) 4-Nitrophenol	14.777	109	5340	3.613	ng/ul	85
56) Dibenzofuran	14.959	168	60654	4.673	ng/ul	95
57) 2,4-Dinitrotoluene	14.918	165	13789	4.179	ng/ul#	100
58) 2,3,4,6-Tetrachlorophenol	15.188	232	12604	4.331	ng/ul#	97
59) Diethylphthalate	15.376	149	56318	4.756	ng/ul	97
61) Fluorene	15.611	166	50773	4.768	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.605	204	26557	4.713	ng/ul	97
63) 4-Nitroaniline	15.623	138	5812	3.317	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.682	198	6371	3.331	ng/ul#	97
67) N-Nitrosodiphenylamine	15.817	169	43752	4.774	ng/ul	96
68) 4-Bromophenyl-phenylether	16.492	248	17143	4.535	ng/ul	92
69) Hexachlorobenzene	16.610	284	19595	4.630	ng/ul	98
70) Atrazine	16.757	200	17545	4.681	ng/ul	91
71) Pentachlorophenol	16.957	266	9066	3.568	ng/ul	95
72) Phenanthrene	17.350	178	84772	4.839	ng/ul	98
74) Anthracene	17.438	178	86342	4.871	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.367	216	20692	4.530	ng/ul	93
76) Pentachlorobenzene	14.883	250	22928	4.700	ng/ul	92
77) Carbazole	17.709	167	78882	4.920	ng/ul	99
78) Di-n-butylphthalate	18.267	149	94894	4.646	ng/ul	99
80) Fluoranthene	19.360	202	104341	4.894	ng/ul	98
82) Pyrene	19.718	202	105991	4.939	ng/ul	98
83) Butylbenzylphthalate	20.605	149	40279	4.628	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.457	252	34283	4.847	ng/ul	98
85) Benzo(a)anthracene	21.540	228	100767	4.804	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.451	149	59997	4.811	ng/ul	99
87) Chrysene	21.604	228	92131	4.685	ng/ul	99
89) Di-n-octyl phthalate	22.632	149	103598	4.960	ng/ul	100
90) Benzo(b)fluoranthene	23.708	252	97014	4.609	ng/ul	97
91) Benzo(k)fluoranthene	23.778	252	100244	4.859	ng/ul	98
93) Benzo(a)pyrene	24.565	252	86621	4.650	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	28.302	276	113535	4.589	ng/ul	95
95) Dibenzo(a,h)anthracene	28.361	278	92692	4.544	ng/ul	99
96) Benzo(g,h,i)perylene	29.424	276	94771	4.690	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 02 07:09:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Jul 02 07:09:00 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.997 ng/uL
RT: 3.402 min Scan# 51
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion: 88	Resp: 3092
Ion Ratio	Lower Upper
88	100
43	29.0 20.6 30.8
58	75.6 65.5 98.3



#5
Pyridine
Concen: 4.576 ng/uL
RT: 3.796 min Scan# 120
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion: 79	Resp: 17838
Ion Ratio	Lower Upper
79	100
52	78.0 68.6 102.8
51	36.5 29.4 44.0









#6
Benzaldehyde
Concen: 3.435 ng/ul
RT: 7.068 min Scan# 61
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion: 77 Resp: 5141
Ion Ratio Lower Upper
77 100
105 79.6 74.7 112.1
106 74.8 75.9 113.9#



#8
Phenol
Concen: 4.314 ng/ul
RT: 7.115 min Scan# 685
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion: 94 Resp: 19617
Ion Ratio Lower Upper
94 100
65 30.5 22.2 33.4
66 34.2 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 4.451 ng/ul
RT: 7.344 min Scan# 71
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion: 93 Resp: 15631
Ion Ratio Lower Upper
93 100
63 90.1 69.4 104.0
95 33.0 27.0 40.6



#12
2-Chlorophenol
Concen: 4.534 ng/ul
RT: 7.491 min Scan# 749
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:128 Resp: 14585
Ion Ratio Lower Upper
128 100
64 50.1 44.3 66.5
130 31.5 27.3 40.9









#13
2-Methylphenol
Concen: 4.231 ng/ul
RT: 8.373 min Scan# 899
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:108 Resp: 15111
Ion Ratio Lower Upper
108 100
107 79.2 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.491 ng/ul
RT: 8.461 min Scan# 914
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion: 45 Resp: 24274
Ion Ratio Lower Upper
45 100
77 12.9 9.4 14.2
79 12.4 8.3 12.5









#16
Acetophenone
Concen: 4.646 ng/ul
RT: 8.749 min Scan# 963
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:105 Resp: 26065
Ion Ratio Lower Upper
105 100
77 77.9 61.2 91.8
51 26.9 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 4.540 ng/ul
RT: 8.731 min Scan# 960
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion: 70 Resp: 13617
Ion Ratio Lower Upper
70 100
42 52.9 48.0 72.0
101 8.3 7.4 11.2
130 15.8 13.4 20.0









#18
4-Methylphenol
Concen: 4.281 ng/ul
RT: 8.696 min Scan# 91
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:108 Resp: 16534
Ion Ratio Lower Upper
108 100
107 116.5 91.1 136.7



#19
Hexachloroethane
Concen: 4.576 ng/ul
RT: 9.013 min Scan# 1008
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:117 Resp: 5643
Ion Ratio Lower Upper
117 100
201 110.1 94.0 141.0
199 64.7 59.8 89.6









#22
Nitrobenzene
Concen: 4.434 ng/ul
RT: 9.136 min Scan# 1029
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion: 77 Resp: 18665
Ion Ratio Lower Upper
77 100
123 34.9 33.3 49.9
65 10.6 10.4 15.6



#23
Isophorone
Concen: 4.609 ng/ul
RT: 9.648 min Scan# 1116
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10


Tgt Ion: 82 Resp: 41401
Ion Ratio Lower Upper
82 100
95 6.6 4.8 7.2
138 15.4 11.4 17.0








#25
2-Nitrophenol
Concen: 4.090 ng/ul
RT: 9.841 min Scan# 1149
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Instrument :
BNA_G
ClientSampleId :
SSTD005435







Tgt Ion:139 Resp: 7479
Ion Ratio Lower Upper
139 100
65 54.2 42.6 64.0
109 29.6 20.7 31.1



#26
2,4-Dimethylphenol
Concen: 4.474 ng/ul
RT: 9.900 min Scan# 1159
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:107 Resp: 18399
Ion Ratio Lower Upper
107 100
121 50.7 40.6 60.8
122 75.5 67.1 100.7





#27

Bis(2-Chloroethoxy)methane

Concen: 4.606 ng/ul

RT: 10.135 min Scan# 1199

Delta R.T. -0.004 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion: 93 Resp: 23315

Ion Ratio Lower Upper

93 100

95 37.5 26.4 39.6

123 9.8 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 4.375 ng/ul

RT: 10.382 min Scan# 1241

Delta R.T. 0.002 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10




Tgt Ion:162 Resp: 13908


Ion Ratio Lower Upper

162 100

164 66.7 51.1 76.7

98 34.5 27.9 41.9









#30
Naphthalene
Concen: 4.616 ng/ul
RT: 10.782 min Scan# 1309
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:128 Resp: 51611
Ion Ratio Lower Upper
128 100
129 10.5 8.6 12.8
127 12.9 10.2 15.2



#32
4-Chloroaniline
Concen: 4.328 ng/ul
RT: 10.893 min Scan# 1328
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:127 Resp: 21934
Ion Ratio Lower Upper
127 100
129 30.9 26.0 39.0





#33

Hexachlorobutadiene

Concen: 4.591 ng/ul

RT: 11.064 min Scan# 11

Delta R.T. -0.004 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:225 Resp: 10331

Ion Ratio Lower Upper

225 100

223 56.3 53.2 79.8

227 63.6 51.4 77.2



#34

Caprolactam

Concen: 4.101 ng/ul

RT: 11.633 min Scan# 1454

Delta R.T. -0.021 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10




Tgt Ion:113 Resp: 5212


Ion Ratio Lower Upper

113 100

55 222.3 183.4 275.2

56 157.0 123.0 184.6





#35

4-Chloro-3-methylphenol

Concen: 4.352 ng/ul

RT: 12.015 min Scan# 1519

Delta R.T. 0.002 min

Lab File: BG057937.D


Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435






Tgt Ion:107 Resp: 16993

Ion Ratio Lower Upper

107 100

144 24.8 19.0 28.4

142 72.7 58.4 87.6



#36

2-Methylnaphthalene

Concen: 4.612 ng/ul

RT: 12.391 min Scan# 1583

Delta R.T. 0.002 min

Lab File: BG057937.D




Acq: 1 Jul 2023 11:10


Tgt Ion:142 Resp: 34757

Ion Ratio Lower Upper

142 100

141 88.1 69.4 104.2









#37
1-Methylnaphthalene
Concen: 4.623 ng/ul
RT: 12.609 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:142 Resp: 34937
Ion Ratio Lower Upper
142 100
141 93.5 73.8 110.8
116 4.0 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.527 ng/ul
RT: 12.756 min Scan# 1645
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:216 Resp: 19742
Ion Ratio Lower Upper
216 100
214 81.5 64.1 96.1
179 19.4 16.0 24.0
108 14.1 12.1 18.1





#40

Hexachlorocyclopentadiene

Concen: 3.432 ng/ul

RT: 12.738 min Scan# 1

Delta R.T. 0.002 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:237 Resp: 9429

Ion Ratio Lower Upper

237 100

235 66.3 49.2 73.8

272 13.3 10.4 15.6



#41

2,4,6-Trichlorophenol

Concen: 4.200 ng/ul

RT: 12.997 min Scan# 1686

Delta R.T. -0.004 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10




Tgt Ion:196 Resp: 12414


Ion Ratio Lower Upper

196 100

198 98.2 73.5 110.3

200 33.6 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 4.174 ng/ul
RT: 13.073 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:196 Resp: 13343
Ion Ratio Lower Upper
196 100
198 102.2 77.2 115.8
200 31.3 23.8 35.6



#43
1,1'-Biphenyl
Concen: 4.646 ng/ul
RT: 13.396 min Scan# 1754
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:154 Resp: 47916
Ion Ratio Lower Upper
154 100
153 42.0 32.6 49.0
76 14.4 11.6 17.4









#44
2-Chloronaphthalene
Concen: 4.640 ng/ul
RT: 13.437 min Scan# 1761
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:162 Resp: 38406
Ion Ratio Lower Upper
162 100
164 28.8 27.7 41.5
127 37.0 30.9 46.3



#45
2-Nitroaniline
Concen: 4.118 ng/ul
RT: 13.643 min Scan# 1796
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion: 65 Resp: 11851
Ion Ratio Lower Upper
65 100
92 53.7 53.2 79.8
138 81.7 71.1 106.7









#47
Dimethylphthalate
Concen: 4.736 ng/ul
RT: 14.013 min Scan# 1859
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:163 Resp: 53729
Ion Ratio Lower Upper
163 100
194 3.3 3.0 4.4
164 10.5 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 4.180 ng/ul
RT: 14.136 min Scan# 1880
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:165 Resp: 9734
Ion Ratio Lower Upper
165 100
89 56.7 42.7 64.1
121 20.9 14.8 22.2









#50
Acenaphthylene
Concen: 4.625 ng/ul
RT: 14.283 min Scan# 1906
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:152 Resp: 61540
Ion Ratio Lower Upper
152 100
151 20.4 15.6 23.4
153 13.0 10.5 15.7



#51
3-Nitroaniline
Concen: 4.307 ng/ul
RT: 14.465 min Scan# 1936
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:138 Resp: 8476
Ion Ratio Lower Upper
138 100
108 11.6 8.2 12.2
92 125.0 100.2 150.4









#52
Acenaphthene
Concen: 4.652 ng/ul
RT: 14.630 min Scan# 1964
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:153 Resp: 42479
Ion Ratio Lower Upper
153 100
152 45.3 36.5 54.7
154 87.7 70.5 105.7



#53
2,4-Dinitrophenol
Concen: 2.166 ng/ul
RT: 14.671 min Scan# 1971
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:184 Resp: 2629
Ion Ratio Lower Upper
184 100
63 82.0 51.1 76.7#
154 64.8 43.3 64.9









#55
4-Nitrophenol
Concen: 3.613 ng/ul
RT: 14.777 min Scan# 1989
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:109 Resp: 5340
Ion Ratio Lower Upper
109 100
139 140.1 119.0 178.4
65 137.8 135.5 203.3



#56
Dibenzofuran
Concen: 4.673 ng/ul
RT: 14.959 min Scan# 2020
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:168 Resp: 60654
Ion Ratio Lower Upper
168 100
139 37.2 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 4.179 ng/ul

RT: 14.918 min Scan# 2013

Delta R.T. -0.004 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:165 Resp: 13789

Ion Ratio Lower Upper

165 100

63 44.7 35.8 53.8

182 2.5 1.6 2.4#



#58

2,3,4,6-Tetrachlorophenol

Concen: 4.331 ng/ul

RT: 15.188 min Scan# 2059

Delta R.T. 0.002 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Tgt Ion:232 Resp: 12604




Ion Ratio Lower Upper


232 100

131 36.4 30.5 45.7

130 2.1 1.4 2.0#

166 29.7 22.6 34.0









#59
Diethylphthalate
Concen: 4.756 ng/ul
RT: 15.376 min Scan# 2092
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:149 Resp: 56318
Ion Ratio Lower Upper
149 100
177 21.0 17.8 26.8
150 11.7 10.1 15.1



#61
Fluorene
Concen: 4.768 ng/ul
RT: 15.611 min Scan# 2131
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:166 Resp: 50773
Ion Ratio Lower Upper
166 100
165 102.9 79.1 118.7
167 14.3 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 4.713 ng/ul
RT: 15.605 min Scan# 2130
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:204 Resp: 26557
Ion Ratio Lower Upper
204 100
206 30.6 25.8 38.6
141 50.2 41.8 62.6



#63
4-Nitroaniline
Concen: 3.317 ng/ul
RT: 15.623 min Scan# 2133
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:138 Resp: 5812
Ion Ratio Lower Upper
138 100
92 51.8 43.8 65.8
108 81.9 58.8 88.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 3.331 ng/ul
 RT: 15.682 min Scan# 2143
 Delta R.T. -0.004 min
 Lab File: BG057937.D
 Acq: 1 Jul 2023 11:10




Instrument :
 BNA_G
 ClientSampleId :
 SST005435


Tgt Ion:198 Resp: 6371
 Ion Ratio Lower Upper
 198 100
 182 6.8 3.1 4.7#
 77 21.5 17.9 26.9



#67
 N-Nitrosodiphenylamine
 Concen: 4.774 ng/ul
 RT: 15.817 min Scan# 2166
 Delta R.T. 0.002 min
 Lab File: BG057937.D
 Acq: 1 Jul 2023 11:10

Tgt Ion:169 Resp: 43752
 Ion Ratio Lower Upper
 169 100
 168 72.2 54.2 81.2
 167 38.7 30.8 46.2









#68
4-Bromophenyl-phenylether
Concen: 4.535 ng/ul
RT: 16.492 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:248 Resp: 17143
Ion Ratio Lower Upper
248 100
250 101.1 78.3 117.5
141 72.7 49.1 73.7



#69
Hexachlorobenzene
Concen: 4.630 ng/ul
RT: 16.610 min Scan# 2301
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:284 Resp: 19595
Ion Ratio Lower Upper
284 100
142 27.0 21.9 32.9
249 33.0 25.0 37.6





#70

Atrazine

Concen: 4.681 ng/ul

RT: 16.757 min Scan# 21

Delta R.T. -0.003 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:200 Resp: 17545

Ion Ratio Lower Upper

200 100

173 27.7 20.0 30.0

215 51.5 35.8 53.8



#71

Pentachlorophenol

Concen: 3.568 ng/ul

RT: 16.957 min Scan# 2360

Delta R.T. 0.002 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10




Tgt Ion:266 Resp: 9066


Ion Ratio Lower Upper

266 100

264 60.7 51.5 77.3

268 64.7 49.0 73.4









#72
Phenanthrene
Concen: 4.839 ng/ul
RT: 17.350 min Scan# 2427
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:178 Resp: 84772
Ion Ratio Lower Upper
178 100
179 16.4 12.8 19.2
176 18.6 15.7 23.5



#74
Anthracene
Concen: 4.871 ng/ul
RT: 17.438 min Scan# 2442
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:178 Resp: 86342
Ion Ratio Lower Upper
178 100
179 14.4 12.6 18.8
176 18.9 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.530 ng/uL
RT: 13.367 min Scan# 1749
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:216 Resp: 20692
Ion Ratio Lower Upper
216 100
214 74.9 64.8 97.2
218 43.5 39.1 58.7



#76
Pentachlorobenzene
Concen: 4.700 ng/uL
RT: 14.883 min Scan# 2007
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:250 Resp: 22928
Ion Ratio Lower Upper
250 100
248 61.6 52.8 79.2
252 56.4 52.4 78.6









#77
 Carbazole
 Concen: 4.920 ng/ul
 RT: 17.709 min Scan# 2488
 Delta R.T. -0.003 min
 Lab File: BG057937.D
 Acq: 1 Jul 2023 11:10




Instrument :
 BNA_G
 ClientSampleId :
 SST005435


Tgt Ion:167 Resp: 78882
 Ion Ratio Lower Upper
 167 100
 166 20.9 17.5 26.3
 139 12.0 9.5 14.3



#78
 Di-n-butylphthalate
 Concen: 4.646 ng/ul
 RT: 18.267 min Scan# 2583
 Delta R.T. 0.002 min
 Lab File: BG057937.D
 Acq: 1 Jul 2023 11:10

Tgt Ion:149 Resp: 94894
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.2
 104 4.8 3.9 5.9









#80
Fluoranthene
Concen: 4.894 ng/ul
RT: 19.360 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:202 Resp: 104341
Ion Ratio Lower Upper
202 100
101 10.9 9.1 13.7
100 7.8 7.1 10.7



#82
Pyrene
Concen: 4.939 ng/ul
RT: 19.718 min Scan# 2830
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:202 Resp: 105991
Ion Ratio Lower Upper
202 100
101 13.6 10.2 15.2
100 11.5 8.9 13.3









#83
Butylbenzylphthalate
Concen: 4.628 ng/ul
RT: 20.605 min Scan# 2982
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10




Instrument :
BNA_G
ClientSampleId :
SSTD005435


Tgt Ion:149 Resp: 40279
Ion Ratio Lower Upper
149 100
91 72.4 56.5 84.7
206 22.9 17.4 26.0



#84
3,3'-Dichlorobenzidine
Concen: 4.847 ng/ul
RT: 21.457 min Scan# 3126
Delta R.T. -0.004 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:252 Resp: 34283
Ion Ratio Lower Upper
252 100
254 61.7 50.7 76.1
126 7.7 6.6 10.0





#85

Benzo(a)anthracene

Concen: 4.804 ng/ul

RT: 21.540 min Scan# 3141

Delta R.T. -0.003 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:228 Resp: 100767

Ion Ratio Lower Upper

228 100

229 21.0 16.0 24.0

226 27.9 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.811 ng/ul

RT: 21.451 min Scan# 3125

Delta R.T. 0.002 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10




Tgt Ion:149 Resp: 59997


Ion Ratio Lower Upper

149 100

167 26.3 20.9 31.3

279 4.6 3.1 4.7





#87

Chrysene

Concen: 4.685 ng/ul

RT: 21.604 min Scan# 3151

Delta R.T. -0.003 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:228 Resp: 92131

Ion Ratio Lower Upper

228 100

226 31.4 24.4 36.6

229 19.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 4.960 ng/ul




RT: 22.632 min Scan# 3326


Delta R.T. -0.004 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Tgt Ion:149 Resp: 103598





#90

Benzo(b)fluoranthene

Concen: 4.609 ng/ul

RT: 23.708 min Scan# 3509

Delta R.T. -0.003 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:252 Resp: 97014

Ion Ratio Lower Upper

252 100

253 24.0 18.1 27.1

125 10.5 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 4.859 ng/ul

RT: 23.778 min Scan# 3521

Delta R.T. 0.002 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10




Tgt Ion:252 Resp: 100244


Ion Ratio Lower Upper

252 100

253 23.2 17.7 26.5

125 8.9 7.4 11.2





#93

Benzo(a)pyrene

Concen: 4.650 ng/ul

RT: 24.565 min Scan# 3

Delta R.T. -0.004 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10

Instrument :

BNA_G

ClientSampleId :

SSTD005435





Tgt Ion:252 Resp: 86621

Ion Ratio Lower Upper

252 100

253 23.5 17.8 26.6

125 9.6 8.3 12.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.589 ng/ul

RT: 28.302 min Scan# 4291

Delta R.T. -0.003 min

Lab File: BG057937.D

Acq: 1 Jul 2023 11:10




Tgt Ion:276 Resp: 113535


Ion Ratio Lower Upper

276 100

138 18.2 14.0 21.0

277 20.8 19.4 29.2









#95
Dibenzo(a,h)anthracene
Concen: 4.544 ng/ul
RT: 28.361 min Scan# 41
Delta R.T. -0.009 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Instrument :
BNA_G
ClientSampleId :
SSTD005435

Tgt Ion:278 Resp: 92692

Ion	Ratio	Lower	Upper
278	100		
139	15.2	11.3	16.9
279	23.2	18.9	28.3



#96
Benzo(g,h,i)perylene
Concen: 4.690 ng/ul
RT: 29.424 min Scan# 4482
Delta R.T. -0.003 min
Lab File: BG057937.D
Acq: 1 Jul 2023 11:10

Tgt Ion:276 Resp: 94771

Ion	Ratio	Lower	Upper
276	100		
138	17.6	14.2	21.2
277	20.6	18.6	27.8

