

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070123\
 Data File : BG057942.D
 Acq On : 1 Jul 2023 14:38
 Operator : CG/JU
 Sample : SSTD16091
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160440

Quant Time: Jul 02 07:11:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 07:09:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.929	152	43558	20.000	ng/u1	0.00
20) Naphthalene-d8	10.738	136	196149	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.569	164	139279	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.312	188	329105	20.000	ng/u1	0.00
79) Chrysene-d12	21.572	240	265546	20.000	ng/u1	0.01
88) Perylene-d12	24.727	264	331434	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	72826	58.546	ng/uL	0.00
4) Pyridine-d5	3.775	84	563653	149.256	ng/u1	0.00
7) Phenol-d5	7.107	99	670741	150.784	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	7.265	67	385196	143.691	ng/u1	0.00
11) 2-Chlorophenol-d4	7.471	132	462119	150.624	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	496765	148.000	ng/u1	0.02
21) Nitrobenzene-d5	9.104	128	245472	153.545	ng/u1	0.01
24) 2-Nitrophenol-d4	9.821	143	277064	164.898	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.362	165	492994	152.950	ng/u1	0.00
31) 4-Chloroaniline-d4	10.879	131	686296	140.436	ng/u1	0.01
46) Dimethylphthalate-d6	13.981	166	1462326	133.068	ng/u1	0.01
49) Acenaphthylene-d8	14.269	160	1603206	133.691	ng/u1	0.01
54) 4-Nitrophenol-d4	14.792	143	293127	150.976	ng/u1	0.03
60) Fluorene-d10	15.567	176	1235380	132.370	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.697	200	291848	163.313	ng/u1	0.03
73) Anthracene-d10	17.418	188	1946556	127.465	ng/u1	0.01
81) Pyrene-d10	19.704	212	2175542	130.083	ng/u1	0.01
92) Benzo(a)pyrene-d12	24.528	264	2393059	141.720	ng/u1	0.03
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	80882	52.633	ng/uL	100
5) Pyridine	3.799	79	567384	146.636	ng/u1	98
6) Benzaldehyde	7.071	77	208317	140.227	ng/u1	96
8) Phenol	7.130	94	667930	147.966	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.359	93	496009	142.308	ng/u1	99
12) 2-Chlorophenol	7.500	128	468075	146.611	ng/u1	95
13) 2-Methylphenol	8.382	108	531436	149.903	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.464	45	767792	143.106	ng/u1	98
16) Acetophenone	8.764	105	765776	137.510	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.770	70	400916	134.670	ng/u1	96
18) 4-Methylphenol	8.728	108	563039	146.877	ng/u1	95
19) Hexachloroethane	9.016	117	181464	148.264	ng/u1	99
22) Nitrobenzene	9.146	77	609726	146.915	ng/u1	96
23) Isophorone	9.680	82	1246491	140.756	ng/u1	98
25) 2-Nitrophenol	9.856	139	281807	156.305	ng/u1	93
26) 2,4-Dimethylphenol	9.915	107	583657	143.963	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.150	93	697216	139.715	ng/u1	100
29) 2,4-Dichlorophenol	10.391	162	457509	145.987	ng/u1	92
30) Naphthalene	10.791	128	1498178	135.912	ng/u1	96
32) 4-Chloroaniline	10.908	127	683122	136.710	ng/u1	98
33) Hexachlorobutadiene	11.073	225	320161	144.302	ng/u1	97
34) Caprolactam	11.713	113	101141	80.726	ng/u1	94
35) 4-Chloro-3-methylphenol	12.036	107	576442	149.738	ng/u1	99
36) 2-Methylnaphthalene	12.395	142	1005631	135.334	ng/u1	96


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070123\
 Data File : BG057942.D
 Acq On : 1 Jul 2023 14:38
 Operator : CG/JU
 Sample : SSTD16091
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160440

Quant Time: Jul 02 07:11:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 07:09:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.618	142	1012983	135.944	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.765	216	589318	138.845	ng/ul	99
40) Hexachlorocyclopentadiene	12.741	237	425267	159.069	ng/ul	99
41) 2,4,6-Trichlorophenol	13.006	196	435666	151.445	ng/ul	97
42) 2,4,5-Trichlorophenol	13.082	196	476866	153.277	ng/ul	99
43) 1,1'-Biphenyl	13.405	154	1276961	127.225	ng/ul	97
44) 2-Chloronaphthalene	13.452	162	1077447	133.754	ng/ul	97
45) 2-Nitroaniline	13.658	65	435037	155.330	ng/ul	98
47) Dimethylphthalate	14.034	163	1431609	129.667	ng/ul	98
48) 2,6-Dinitrotoluene	14.157	165	345148	152.300	ng/ul	95
50) Acenaphthylene	14.298	152	1631724	126.005	ng/ul	95
51) 3-Nitroaniline	14.486	138	293565	153.285	ng/ul	99
52) Acenaphthene	14.639	153	1167822	131.421	ng/ul	97
53) 2,4-Dinitrophenol	14.686	184	221761	187.698	ng/ul	95
55) 4-Nitrophenol	14.809	109	221386	153.922	ng/ul	86
56) Dibenzofuran	14.974	168	1600116	126.675	ng/ul	95
57) 2,4-Dinitrotoluene	14.945	165	476274	148.331	ng/ul#	95
58) 2,3,4,6-Tetrachlorophenol	15.197	232	429817	151.761	ng/ul#	99
59) Diethylphthalate	15.397	149	1479689	128.408	ng/ul	96
61) Fluorene	15.626	166	1253157	120.910	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.609	204	702745	128.143	ng/ul	98
63) 4-Nitroaniline	15.661	138	279468	163.891	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.708	198	292425	160.012	ng/ul	98
67) N-Nitrosodiphenylamine	15.826	169	1149075	131.195	ng/ul	98
68) 4-Bromophenyl-phenylether	16.502	248	520315	144.030	ng/ul	99
69) Hexachlorobenzene	16.619	284	576391	142.508	ng/ul	98
70) Atrazine	16.784	200	483646	135.036	ng/ul	95
71) Pentachlorophenol	16.966	266	390488	160.815	ng/ul	98
72) Phenanthrene	17.359	178	2074825	123.940	ng/ul	95
74) Anthracene	17.453	178	2030978	119.887	ng/ul	93
75) 1,2,3,4-Tetrachloroben...	13.370	216	636866	145.913	ng/ul	98
76) Pentachlorobenzene	14.892	250	667081	143.104	ng/ul	99
77) Carbazole	17.724	167	1968949	128.512	ng/ul	95
78) Di-n-butylphthalate	18.276	149	2307059	118.210	ng/ul	96
80) Fluoranthene	19.369	202	2408718	124.295	ng/ul	98
82) Pyrene	19.733	202	2340425	119.978	ng/ul	95
83) Butylbenzylphthalate	20.614	149	1071052	135.401	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.472	252	788683	122.683	ng/ul	95
85) Benzo(a)anthracene	21.555	228	2519662	132.151	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.455	149	1440581	127.085	ng/ul	95
87) Chrysene	21.619	228	2382056	133.272	ng/ul	94
89) Di-n-octyl phthalate	22.641	149	2628762	129.088	ng/ul	100
90) Benzo(b)fluoranthene	23.734	252	2840616	138.421	ng/ul	98
91) Benzo(k)fluoranthene	23.811	252	2633957	130.947	ng/ul	97
93) Benzo(a)pyrene	24.610	252	2494997	137.362	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	28.370	276	3509681	145.510	ng/ul	99
95) Dibenzo(a,h)anthracene	28.441	278	2849253	143.255	ng/ul	97
96) Benzo(g,h,i)perylene	29.522	276	2886417	146.512	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 52.633 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. -0.000 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:	88	Resp:	80882
Ion	Ratio	Lower	Upper
88	100		
43	26.2	20.6	30.8
58	81.7	65.5	98.3



#5
Pyridine
Concen: 146.636 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:	79	Resp:	567384
Ion	Ratio	Lower	Upper
79	100		
52	87.3	68.6	102.8
51	39.0	29.4	44.0









#6
Benzaldehyde
Concen: 140.227 ng/ul
RT: 7.071 min Scan# 61
Delta R.T. -0.000 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion: 77 Resp: 208317
Ion Ratio Lower Upper
77 100
105 101.1 74.7 112.1
106 94.8 75.9 113.9



#8
Phenol
Concen: 147.966 ng/ul
RT: 7.130 min Scan# 688
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion: 94 Resp: 667930
Ion Ratio Lower Upper
94 100
65 30.5 22.2 33.4
66 37.2 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 142.308 ng/ul
RT: 7.359 min Scan# 71
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion: 93 Resp: 496009
Ion Ratio Lower Upper
93 100
63 86.5 69.4 104.0
95 32.7 27.0 40.6



#12
2-Chlorophenol
Concen: 146.611 ng/ul
RT: 7.500 min Scan# 751
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:128 Resp: 468075
Ion Ratio Lower Upper
128 100
64 59.9 44.3 66.5
130 33.0 27.3 40.9









#13
2-Methylphenol
Concen: 149.903 ng/ul
RT: 8.382 min Scan# 901
Delta R.T. 0.011 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:108 Resp: 531436
Ion Ratio Lower Upper
108 100
107 88.3 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 143.106 ng/ul
RT: 8.464 min Scan# 915
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion: 45 Resp: 767792
Ion Ratio Lower Upper
45 100
77 13.2 9.4 14.2
79 10.2 8.3 12.5









#16
Acetophenone
Concen: 137.510 ng/ul
RT: 8.764 min Scan# 966
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:105 Resp: 765776
Ion Ratio Lower Upper
105 100
77 78.7 61.2 91.8
51 31.1 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 134.670 ng/ul
RT: 8.770 min Scan# 967
Delta R.T. 0.041 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion: 70 Resp: 400916
Ion Ratio Lower Upper
70 100
42 62.9 48.0 72.0
101 9.2 7.4 11.2
130 18.8 13.4 20.0









#18
4-Methylphenol
Concen: 146.877 ng/ul
RT: 8.728 min Scan# 90
Delta R.T. 0.029 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:108 Resp: 563039
Ion Ratio Lower Upper
108 100
107 108.6 91.1 136.7



#19
Hexachloroethane
Concen: 148.264 ng/ul
RT: 9.016 min Scan# 1009
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:117 Resp: 181464
Ion Ratio Lower Upper
117 100
201 117.8 94.0 141.0
199 73.6 59.8 89.6









#22
Nitrobenzene
Concen: 146.915 ng/ul
RT: 9.146 min Scan# 1029
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion: 77 Resp: 609726
Ion Ratio Lower Upper
77 100
123 43.9 33.3 49.9
65 14.5 10.4 15.6



#23
Isophorone
Concen: 140.756 ng/ul
RT: 9.680 min Scan# 1122
Delta R.T. 0.029 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion: 82 Resp: 1246491
Ion Ratio Lower Upper
82 100
95 6.7 4.8 7.2
138 15.3 11.4 17.0









#25
2-Nitrophenol
Concen: 156.305 ng/ul
RT: 9.856 min Scan# 1150
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:139 Resp: 281807
Ion Ratio Lower Upper
139 100
65 57.8 42.6 64.0
109 29.8 20.7 31.1



#26
2,4-Dimethylphenol
Concen: 143.963 ng/ul
RT: 9.915 min Scan# 1162
Delta R.T. 0.011 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:107 Resp: 583657
Ion Ratio Lower Upper
107 100
121 49.8 40.6 60.8
122 83.2 67.1 100.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 139.715 ng/ul
 RT: 10.150 min Scan# 1202
 Delta R.T. 0.011 min
 Lab File: BG057942.D
 Acq: 1 Jul 2023 14:38




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160440


Tgt Ion: 93 Resp: 697216
 Ion Ratio Lower Upper
 93 100
 95 33.4 26.4 39.6
 123 9.9 7.9 11.9



#29
 2,4-Dichlorophenol
 Concen: 145.987 ng/ul
 RT: 10.391 min Scan# 1243
 Delta R.T. 0.012 min
 Lab File: BG057942.D
 Acq: 1 Jul 2023 14:38

Tgt Ion:162 Resp: 457509
 Ion Ratio Lower Upper
 162 100
 164 70.2 51.1 76.7
 98 39.0 27.9 41.9









#30
Naphthalene
Concen: 135.912 ng/ul
RT: 10.791 min Scan# 1311
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:128 Resp: 1498178
Ion Ratio Lower Upper
128 100
129 12.2 8.6 12.8
127 14.6 10.2 15.2



#32
4-Chloroaniline
Concen: 136.710 ng/ul
RT: 10.908 min Scan# 1331
Delta R.T. 0.017 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:127 Resp: 683122
Ion Ratio Lower Upper
127 100
129 33.4 26.0 39.0









#33
Hexachlorobutadiene
Concen: 144.302 ng/ul
RT: 11.073 min Scan# 11
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:225 Resp: 320161
Ion Ratio Lower Upper
225 100
223 64.0 53.2 79.8
227 65.9 51.4 77.2



#34
Caprolactam
Concen: 80.726 ng/ul
RT: 11.713 min Scan# 1468
Delta R.T. 0.059 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:113 Resp: 101141
Ion Ratio Lower Upper
113 100
55 229.9 183.4 275.2
56 171.2 123.0 184.6









#35
 4-Chloro-3-methylphenol
 Concen: 149.738 ng/ul
 RT: 12.036 min Scan# 1519
 Delta R.T. 0.023 min
 Lab File: BG057942.D
 Acq: 1 Jul 2023 14:38

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160440

Tgt Ion:107 Resp: 576442




Ion	Ratio	Lower	Upper
107	100		
144	23.0	19.0	28.4
142	73.5	58.4	87.6




#36
 2-Methylnaphthalene
 Concen: 135.334 ng/ul
 RT: 12.395 min Scan# 1584
 Delta R.T. 0.006 min
 Lab File: BG057942.D
 Acq: 1 Jul 2023 14:38

Tgt Ion:142 Resp: 1005631

Ion	Ratio	Lower	Upper
142	100		
141	90.1	69.4	104.2





#37

1-Methylnaphthalene

Concen: 135.944 ng/ul

RT: 12.618 min Scan# 1

Delta R.T. 0.006 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD160440





Tgt Ion:142 Resp: 1012983

Ion Ratio Lower Upper

142 100

141 92.4 73.8 110.8

116 3.8 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 138.845 ng/ul

RT: 12.765 min Scan# 1647

Delta R.T. 0.006 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Tgt Ion:216 Resp: 589318




Ion Ratio Lower Upper


216 100

214 80.8 64.1 96.1

179 20.5 16.0 24.0


108 15.9 12.1 18.1








#40
Hexachlorocyclopentadiene
Concen: 159.069 ng/ul
RT: 12.741 min Scan# 1643
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Instrument :
BNA_G
ClientSampleId :
SSTD160440







Tgt Ion:237 Resp: 425267
Ion Ratio Lower Upper
237 100
235 62.6 49.2 73.8
272 13.1 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 151.445 ng/ul
RT: 13.006 min Scan# 1688
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:196 Resp: 435666
Ion Ratio Lower Upper
196 100
198 96.0 73.5 110.3
200 31.2 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 153.277 ng/ul
RT: 13.082 min Scan# 1701
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:196 Resp: 476866
Ion Ratio Lower Upper
196 100
198 96.7 77.2 115.8
200 30.9 23.8 35.6



#43
1,1'-Biphenyl
Concen: 127.225 ng/ul
RT: 13.405 min Scan# 1756
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:154 Resp: 1276961
Ion Ratio Lower Upper
154 100
153 42.4 32.6 49.0
76 15.7 11.6 17.4









#44
2-Chloronaphthalene
Concen: 133.754 ng/ul
RT: 13.452 min Scan# 1764
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:162 Resp: 1077447
Ion Ratio Lower Upper
162 100
164 36.7 27.7 41.5
127 40.1 30.9 46.3



#45
2-Nitroaniline
Concen: 155.330 ng/ul
RT: 13.658 min Scan# 1799
Delta R.T. 0.017 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion: 65 Resp: 435037
Ion Ratio Lower Upper
65 100
92 64.9 53.2 79.8
138 91.1 71.1 106.7









#47
Dimethylphthalate
Concen: 129.667 ng/ul
RT: 14.034 min Scan# 1863
Delta R.T. 0.017 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:163 Resp: 1431609
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.4
164 11.5 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 152.300 ng/ul
RT: 14.157 min Scan# 1884
Delta R.T. 0.017 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38


Tgt Ion:165 Resp: 345148
Ion Ratio Lower Upper
165 100
89 58.0 42.7 64.1
121 19.1 14.8 22.2








#50
Acenaphthylene
Concen: 126.005 ng/ul
RT: 14.298 min Scan# 1908
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Instrument :
BNA_G
ClientSampleId :
SSTD160440







Tgt Ion:152 Resp: 1631724
Ion Ratio Lower Upper
152 100
151 22.2 15.6 23.4
153 14.5 10.5 15.7



#51
3-Nitroaniline
Concen: 153.285 ng/ul
RT: 14.486 min Scan# 1940
Delta R.T. 0.023 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:138 Resp: 293565
Ion Ratio Lower Upper
138 100
108 11.6 8.2 12.2
92 126.2 100.2 150.4





#52

Acenaphthene

Concen: 131.421 ng/ul

RT: 14.639 min Scan# 1966

Delta R.T. 0.012 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD160440





Tgt Ion:153 Resp: 1167822

Ion Ratio Lower Upper

153 100

152 49.4 36.5 54.7

154 87.3 70.5 105.7



#53

2,4-Dinitrophenol

Concen: 187.698 ng/ul

RT: 14.686 min Scan# 1974

Delta R.T. 0.012 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38




Tgt Ion:184 Resp: 221761


Ion Ratio Lower Upper

184 100

63 68.6 51.1 76.7

154 57.4 43.3 64.9









#55
4-Nitrophenol
Concen: 153.922 ng/ul
RT: 14.809 min Scan# 1989
Delta R.T. 0.035 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:109 Resp: 221386
Ion Ratio Lower Upper
109 100
139 129.2 119.0 178.4
65 152.5 135.5 203.3



#56
Dibenzofuran
Concen: 126.675 ng/ul
RT: 14.974 min Scan# 2023
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:168 Resp: 1600116
Ion Ratio Lower Upper
168 100
139 37.4 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 148.331 ng/ul

RT: 14.945 min Scan# 2018

Delta R.T. 0.023 min

Lab File: BG057942.D


Acq: 1 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD160440






Tgt Ion:165 Resp: 476274

Ion Ratio Lower Upper

165 100

63 48.4 35.8 53.8

182 2.4 1.6 2.4#



#58

2,3,4,6-Tetrachlorophenol

Concen: 151.761 ng/ul

RT: 15.197 min Scan# 2061

Delta R.T. 0.012 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Tgt Ion:232 Resp: 429817




Ion Ratio Lower Upper


232 100

131 37.6 30.5 45.7

130 2.1 1.4 2.0#

166 28.3 22.6 34.0









#59
Diethylphthalate
Concen: 128.408 ng/ul
RT: 15.397 min Scan# 2095
Delta R.T. 0.017 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:149 Resp: 1479689
Ion Ratio Lower Upper
149 100
177 24.5 17.8 26.8
150 13.3 10.1 15.1



#61
Fluorene
Concen: 120.910 ng/ul
RT: 15.626 min Scan# 2134
Delta R.T. 0.017 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:166 Resp: 1253157
Ion Ratio Lower Upper
166 100
165 100.5 79.1 118.7
167 14.5 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 128.143 ng/ul
RT: 15.609 min Scan# 2110
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:204 Resp: 702745
Ion Ratio Lower Upper
204 100
206 34.4 25.8 38.6
141 52.3 41.8 62.6



#63
4-Nitroaniline
Concen: 163.891 ng/ul
RT: 15.661 min Scan# 2140
Delta R.T. 0.035 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:138 Resp: 279468
Ion Ratio Lower Upper
138 100
92 54.8 43.8 65.8
108 79.1 58.8 88.2









#66
4,6-Dinitro-2-methylphenol
Concen: 160.012 ng/ul
RT: 15.708 min Scan# 2148
Delta R.T. 0.023 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:198 Resp: 292425
Ion Ratio Lower Upper
198 100
182 3.3 3.1 4.7
77 23.6 17.9 26.9



#67
N-Nitrosodiphenylamine
Concen: 131.195 ng/ul
RT: 15.826 min Scan# 2168
Delta R.T. 0.011 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:169 Resp: 1149075
Ion Ratio Lower Upper
169 100
168 70.0 54.2 81.2
167 39.7 30.8 46.2









#68
4-Bromophenyl-phenylether
Concen: 144.030 ng/ul
RT: 16.502 min Scan# 21
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:248 Resp: 520315
Ion Ratio Lower Upper
248 100
250 97.0 78.3 117.5
141 61.5 49.1 73.7



#69
Hexachlorobenzene
Concen: 142.508 ng/ul
RT: 16.619 min Scan# 2303
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:284 Resp: 576391
Ion Ratio Lower Upper
284 100
142 28.3 21.9 32.9
249 32.6 25.0 37.6





#70

Atrazine

Concen: 135.036 ng/ul

RT: 16.784 min Scan# 21

Delta R.T. 0.023 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD160440





Tgt Ion:200 Resp: 483646

Ion Ratio Lower Upper

200 100

173 26.4 20.0 30.0

215 48.5 35.8 53.8



#71

Pentachlorophenol

Concen: 160.815 ng/ul

RT: 16.966 min Scan# 2362

Delta R.T. 0.012 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38




Tgt Ion:266 Resp: 390488


Ion Ratio Lower Upper

266 100

264 62.6 51.5 77.3

268 62.8 49.0 73.4





#72

Phenanthrene

Concen: 123.940 ng/ul

RT: 17.359 min Scan# 2429

Delta R.T. 0.011 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD160440





Tgt Ion:178 Resp: 2074825

Ion Ratio Lower Upper

178 100

179 17.8 12.8 19.2

176 21.8 15.7 23.5



#74

Anthracene

Concen: 119.887 ng/ul

RT: 17.453 min Scan# 2445

Delta R.T. 0.011 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38




Tgt Ion:178 Resp: 2030978


Ion Ratio Lower Upper

178 100

179 18.7 12.6 18.8

176 22.3 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 145.913 ng/uL
RT: 13.370 min Scan# 1749
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:216 Resp: 636866
Ion Ratio Lower Upper
216 100
214 78.2 64.8 97.2
218 48.0 39.1 58.7



#76
Pentachlorobenzene
Concen: 143.104 ng/uL
RT: 14.892 min Scan# 2009
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:250 Resp: 667081
Ion Ratio Lower Upper
250 100
248 65.2 52.8 79.2
252 65.0 52.4 78.6









#77
 Carbazole
 Concen: 128.512 ng/ul
 RT: 17.724 min Scan# 2489
 Delta R.T. 0.012 min
 Lab File: BG057942.D
 Acq: 1 Jul 2023 14:38




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160440


Tgt Ion:167 Resp: 1968949
 Ion Ratio Lower Upper
 167 100
 166 24.8 17.5 26.3
 139 13.5 9.5 14.3



#78
 Di-n-butylphthalate
 Concen: 118.210 ng/ul
 RT: 18.276 min Scan# 2585
 Delta R.T. 0.012 min
 Lab File: BG057942.D
 Acq: 1 Jul 2023 14:38

Tgt Ion:149 Resp: 2307059
 Ion Ratio Lower Upper
 149 100
 150 11.0 7.4 11.2
 104 5.8 3.9 5.9









#80
Fluoranthene
Concen: 124.295 ng/u1
RT: 19.369 min Scan# 2771
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:202 Resp: 2408718
Ion Ratio Lower Upper
202 100
101 12.0 9.1 13.7
100 9.7 7.1 10.7



#82
Pyrene
Concen: 119.978 ng/u1
RT: 19.733 min Scan# 2833
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:202 Resp: 2340425
Ion Ratio Lower Upper
202 100
101 14.8 10.2 15.2
100 12.5 8.9 13.3





#83

Butylbenzylphthalate

Concen: 135.401 ng/ul

RT: 20.614 min Scan# 2983

Delta R.T. 0.006 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD160440





Tgt Ion:149 Resp: 1071052

Ion Ratio Lower Upper

149 100

91 75.4 56.5 84.7

206 24.1 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 122.683 ng/ul

RT: 21.472 min Scan# 3129

Delta R.T. 0.012 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38




Tgt Ion:252 Resp: 788683


Ion Ratio Lower Upper

252 100

254 67.2 50.7 76.1

126 9.4 6.6 10.0









#85
Benzo(a)anthracene
Concen: 132.151 ng/ul
RT: 21.555 min Scan# 3143
Delta R.T. 0.012 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:228 Resp: 2519662
Ion Ratio Lower Upper
228 100
229 22.5 16.0 24.0
226 30.6 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 127.085 ng/ul
RT: 21.455 min Scan# 3126
Delta R.T. 0.006 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:149 Resp: 1440581
Ion Ratio Lower Upper
149 100
167 29.0 20.9 31.3
279 4.4 3.1 4.7





#87

Chrysene

Concen: 133.272 ng/ul

RT: 21.619 min Scan# 3154

Delta R.T. 0.012 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Instrument :

BNA_G

ClientSampleId :

SSTD160440





Tgt Ion:228 Resp: 2382056

Ion Ratio Lower Upper

228 100

226 33.6 24.4 36.6

229 22.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 129.088 ng/ul




RT: 22.641 min Scan# 3328


Delta R.T. 0.006 min

Lab File: BG057942.D

Acq: 1 Jul 2023 14:38

Tgt Ion:149 Resp: 2628762









#90
Benzo(b)fluoranthene
Concen: 138.421 ng/ul
RT: 23.734 min Scan# 3514
Delta R.T. 0.023 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:252 Resp: 2840616
Ion Ratio Lower Upper
252 100
253 23.7 18.1 27.1
125 9.3 7.5 11.3



#91
Benzo(k)fluoranthene
Concen: 130.947 ng/ul
RT: 23.811 min Scan# 3527
Delta R.T. 0.035 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:252 Resp: 2633957
Ion Ratio Lower Upper
252 100
253 24.3 17.7 26.5
125 9.3 7.4 11.2









#93
Benzo(a)pyrene
Concen: 137.362 ng/ul
RT: 24.610 min Scan# 30
Delta R.T. 0.041 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38




Instrument :
BNA_G
ClientSampleId :
SSTD160440


Tgt Ion:252 Resp: 2494997
Ion Ratio Lower Upper
252 100
253 23.5 17.8 26.6
125 9.8 8.3 12.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 145.510 ng/ul
RT: 28.370 min Scan# 4303
Delta R.T. 0.064 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:276 Resp: 3509681
Ion Ratio Lower Upper
276 100
138 16.7 14.0 21.0
277 24.9 19.4 29.2









#95
Dibenzo(a,h)anthracene
Concen: 143.255 ng/ul
RT: 28.441 min Scan# 41
Delta R.T. 0.070 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Instrument :
BNA_G
ClientSampleId :
SSTD160440

Tgt Ion:278 Resp: 2849253

Ion	Ratio	Lower	Upper
278	100		
139	13.1	11.3	16.9
279	25.6	18.9	28.3



#96
Benzo(g,h,i)perylene
Concen: 146.512 ng/ul
RT: 29.522 min Scan# 4499
Delta R.T. 0.094 min
Lab File: BG057942.D
Acq: 1 Jul 2023 14:38

Tgt Ion:276 Resp: 2886417

Ion	Ratio	Lower	Upper
276	100		
138	17.9	14.2	21.2
277	24.7	18.6	27.8

