

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070123\
 Data File : BG057943.D
 Acq On : 1 Jul 2023 15:20
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV441

Quant Time: Jul 02 07:50:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 07:50:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.929	152	39859	20.000	ng/u1	0.00
20) Naphthalene-d8	10.732	136	177705	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.563	164	123124	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.307	188	297095	20.000	ng/u1	0.00
79) Chrysene-d12	21.560	240	259965	20.000	ng/u1	0.00
88) Perylene-d12	24.716	264	318140	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	10174	8.938	ng/uL	0.00
4) Pyridine-d5	3.775	84	75436	21.829	ng/u1	0.00
7) Phenol-d5	7.089	99	89317	21.942	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.254	67	55009	22.425	ng/u1	0.00
11) 2-Chlorophenol-d4	7.459	132	61560	21.927	ng/u1	0.00
15) 4-Methylphenol-d8	8.634	113	67334	21.922	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	32802	22.647	ng/u1	0.00
24) 2-Nitrophenol-d4	9.810	143	35527	23.339	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.350	165	67764	23.206	ng/u1	0.00
31) 4-Chloroaniline-d4	10.867	131	101049	22.824	ng/u1	0.00
46) Dimethylphthalate-d6	13.969	166	219162	22.560	ng/u1	0.00
49) Acenaphthylene-d8	14.257	160	247138	23.313	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	38053	22.171	ng/u1	0.00
60) Fluorene-d10	15.556	176	190547	23.096	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.667	200	35529	22.023	ng/u1	0.00
73) Anthracene-d10	17.407	188	317387	23.022	ng/u1	0.00
81) Pyrene-d10	19.692	212	372064	22.725	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.498	264	356061	21.968	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	10989	7.815	ng/uL	92
5) Pyridine	3.793	79	79449	22.438	ng/u1	98
6) Benzaldehyde	7.072	77	17624	12.964	ng/u1	94
8) Phenol	7.119	94	92140	22.306	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.348	93	70979	22.254	ng/u1	94
12) 2-Chlorophenol	7.495	128	64524	22.086	ng/u1	98
13) 2-Methylphenol	8.370	108	71230	21.957	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.458	45	108006	21.999	ng/u1	98
16) Acetophenone	8.752	105	116353	22.832	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.734	70	61403	22.540	ng/u1	93
18) 4-Methylphenol	8.699	108	78075	22.257	ng/u1	97
19) Hexachloroethane	9.010	117	24961	22.287	ng/u1	95
22) Nitrobenzene	9.134	77	84287	22.417	ng/u1	98
23) Isophorone	9.651	82	182295	22.722	ng/u1	99
25) 2-Nitrophenol	9.845	139	37709	23.086	ng/u1	98
26) 2,4-Dimethylphenol	9.904	107	81854	22.285	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.139	93	103025	22.788	ng/u1	99
29) 2,4-Dichlorophenol	10.379	162	64448	22.699	ng/u1	93
30) Naphthalene	10.785	128	226353	22.666	ng/u1	97
32) 4-Chloroaniline	10.891	127	105777	23.366	ng/u1	98
33) Hexachlorobutadiene	11.067	225	45041	22.408	ng/u1	99
34) Caprolactam	11.654	113	24433	21.405	ng/u1	94
35) 4-Chloro-3-methylphenol	12.013	107	78279	22.444	ng/u1	98
36) 2-Methylnaphthalene	12.389	142	151449	22.497	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070123\
 Data File : BG057943.D
 Acq On : 1 Jul 2023 15:20
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SICV441


Quant Time: Jul 02 07:50:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 07:50:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.612	142	153134	22.684	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.759	216	86887	23.157	ng/ul	99
40) Hexachlorocyclopentadiene	12.736	237	53827	22.775	ng/ul	95
41) 2,4,6-Trichlorophenol	12.994	196	59451	23.378	ng/ul	99
42) 2,4,5-Trichlorophenol	13.070	196	63866	23.222	ng/ul	94
43) 1,1'-Biphenyl	13.399	154	204529	23.051	ng/ul	99
44) 2-Chloronaphthalene	13.441	162	164957	23.165	ng/ul	99
45) 2-Nitroaniline	13.646	65	58252	23.528	ng/ul	95
47) Dimethylphthalate	14.016	163	222625	22.810	ng/ul	100
48) 2,6-Dinitrotoluene	14.134	165	46384	23.153	ng/ul	93
50) Acenaphthylene	14.287	152	265157	23.163	ng/ul	98
51) 3-Nitroaniline	14.469	138	33084	19.541	ng/ul	91
52) Acenaphthene	14.627	153	179046	22.793	ng/ul	98
53) 2,4-Dinitrophenol	14.674	184	22723	21.756	ng/ul	97
55) 4-Nitrophenol	14.774	109	29615	23.292	ng/ul	95
56) Dibenzofuran	14.962	168	260201	23.302	ng/ul	97
57) 2,4-Dinitrotoluene	14.921	165	67898	23.921	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.186	232	57845	23.104	ng/ul#	97
59) Diethylphthalate	15.379	149	232802	22.853	ng/ul	100
61) Fluorene	15.609	166	212516	23.195	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.603	204	112477	23.201	ng/ul	97
63) 4-Nitroaniline	15.626	138	26148	17.241	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.685	198	38276	23.201	ng/ul	95
67) N-Nitrosodiphenylamine	15.814	169	181725	22.984	ng/ul	97
68) 4-Bromophenyl-phenylether	16.496	248	75277	23.083	ng/ul	98
69) Hexachlorobenzene	16.613	284	85128	23.315	ng/ul	98
70) Atrazine	16.760	200	73728	22.803	ng/ul	97
71) Pentachlorophenol	16.954	266	48655	22.197	ng/ul	97
72) Phenanthrene	17.348	178	347607	23.002	ng/ul	100
74) Anthracene	17.442	178	351213	22.966	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.364	216	92325	23.432	ng/uL	100
76) Pentachlorobenzene	14.880	250	100347	23.846	ng/uL	94
77) Carbazole	17.706	167	331556	23.972	ng/ul	100
78) Di-n-butylphthalate	18.264	149	404946	22.984	ng/ul	99
80) Fluoranthene	19.357	202	429970	22.664	ng/ul	99
82) Pyrene	19.721	202	434156	22.734	ng/ul	99
83) Butylbenzylphthalate	20.609	149	177590	22.933	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.461	252	151404	24.057	ng/ul	99
85) Benzo(a)anthracene	21.543	228	424589	22.747	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.449	149	252633	22.765	ng/ul	98
87) Chrysene	21.602	228	400896	22.911	ng/ul	99
89) Di-n-octyl phthalate	22.636	149	445716	22.802	ng/ul	100
90) Benzo(b)fluoranthene	23.711	252	433038	21.983	ng/ul	99
91) Benzo(k)fluoranthene	23.776	252	437235	22.645	ng/ul	99
93) Benzo(a)pyrene	24.569	252	389292	22.328	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.306	276	515918	22.284	ng/ul	98
95) Dibenzo(a,h)anthracene	28.370	278	424969	22.259	ng/ul	97
96) Benzo(g,h,i)perylene	29.428	276	416091	22.003	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 02 07:50:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Jul 02 07:50:15 2023
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 7.815 ng/uL

RT: 3.405 min Scan# 54

Delta R.T. 0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Instrument :

BNA_G

ClientSampleId :

SICV441





Tgt Ion: 88 Resp: 10989

Ion Ratio Lower Upper

88 100

43 27.0 20.6 30.8

58 90.2 65.5 98.3



#5

Pyridine

Concen: 22.438 ng/uL

RT: 3.793 min Scan# 120

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20




Tgt Ion: 79 Resp: 79449


Ion Ratio Lower Upper

79 100

52 86.6 68.6 102.8

51 38.6 29.4 44.0









#6
Benzaldehyde
Concen: 12.964 ng/ul
RT: 7.072 min Scan# 61
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion: 77 Resp: 17624
Ion Ratio Lower Upper
77 100
105 95.8 74.7 112.1
106 86.0 75.9 113.9



#8
Phenol
Concen: 22.306 ng/ul
RT: 7.119 min Scan# 686
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion: 94 Resp: 92140
Ion Ratio Lower Upper
94 100
65 26.1 22.2 33.4
66 34.3 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 22.254 ng/ul
RT: 7.348 min Scan# 71
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Instrument :
BNA_G
ClientSampleId :
SICV441

Tgt Ion: 93 Resp: 70979




Ion	Ratio	Lower	Upper
93	100		
63	80.7	69.4	104.0
95	30.6	27.0	40.6




#12
2-Chlorophenol
Concen: 22.086 ng/ul
RT: 7.495 min Scan# 750
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:128 Resp: 64524

Ion	Ratio	Lower	Upper
128	100		
64	53.8	44.3	66.5
130	33.6	27.3	40.9









#13
2-Methylphenol
Concen: 21.957 ng/ul
RT: 8.370 min Scan# 899
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:108 Resp: 71230
Ion Ratio Lower Upper
108 100
107 89.9 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.999 ng/ul
RT: 8.458 min Scan# 914
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion: 45 Resp: 108006
Ion Ratio Lower Upper
45 100
77 12.9 9.4 14.2
79 10.3 8.3 12.5









#16
Acetophenone
Concen: 22.832 ng/ul
RT: 8.752 min Scan# 964
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:105 Resp: 116353
Ion Ratio Lower Upper
105 100
77 77.5 61.2 91.8
51 28.2 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 22.540 ng/ul
RT: 8.734 min Scan# 961
Delta R.T. 0.006 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion: 70 Resp: 61403
Ion Ratio Lower Upper
70 100
42 53.0 48.0 72.0
101 7.9 7.4 11.2
130 17.8 13.4 20.0









#18
4-Methylphenol
Concen: 22.257 ng/ul
RT: 8.699 min Scan# 91
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:108 Resp: 78075
Ion Ratio Lower Upper
108 100
107 116.9 91.1 136.7



#19
Hexachloroethane
Concen: 22.287 ng/ul
RT: 9.010 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:117 Resp: 24961
Ion Ratio Lower Upper
117 100
201 116.1 94.0 141.0
199 65.6 59.8 89.6









#22
Nitrobenzene
Concen: 22.417 ng/ul
RT: 9.134 min Scan# 1029
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion: 77 Resp: 84287
Ion Ratio Lower Upper
77 100
123 43.0 33.3 49.9
65 14.0 10.4 15.6



#23
Isophorone
Concen: 22.722 ng/ul
RT: 9.651 min Scan# 1117
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion: 82 Resp: 182295
Ion Ratio Lower Upper
82 100
95 6.4 4.8 7.2
138 14.6 11.4 17.0









#25
2-Nitrophenol
Concen: 23.086 ng/ul
RT: 9.845 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:139 Resp: 37709
Ion Ratio Lower Upper
139 100
65 54.9 42.6 64.0
109 26.5 20.7 31.1



#26
2,4-Dimethylphenol
Concen: 22.285 ng/ul
RT: 9.904 min Scan# 1160
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:107 Resp: 81854
Ion Ratio Lower Upper
107 100
121 48.5 40.6 60.8
122 82.3 67.1 100.7





#27

Bis(2-Chloroethoxy)methane

Concen: 22.788 ng/ul

RT: 10.139 min Scan# 1200

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Instrument :

BNA_G

ClientSampleId :

SICV441





Tgt Ion: 93 Resp: 103025

Ion Ratio Lower Upper

93 100

95 32.8 26.4 39.6

123 10.2 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 22.699 ng/ul

RT: 10.379 min Scan# 1241

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20




Tgt Ion:162 Resp: 64448


Ion Ratio Lower Upper

162 100

164 70.2 51.1 76.7

98 37.7 27.9 41.9









#30
Naphthalene
Concen: 22.666 ng/ul
RT: 10.785 min Scan# 11
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:128 Resp: 226353
Ion Ratio Lower Upper
128 100
129 11.7 8.6 12.8
127 13.7 10.2 15.2



#32
4-Chloroaniline
Concen: 23.366 ng/ul
RT: 10.891 min Scan# 1328
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:127 Resp: 105777
Ion Ratio Lower Upper
127 100
129 31.6 26.0 39.0









#33
Hexachlorobutadiene
Concen: 22.408 ng/ul
RT: 11.067 min Scan# 11
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:225 Resp: 45041
Ion Ratio Lower Upper
225 100
223 65.3 53.2 79.8
227 63.4 51.4 77.2



#34
Caprolactam
Concen: 21.405 ng/ul
RT: 11.654 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:113 Resp: 24433
Ion Ratio Lower Upper
113 100
55 216.4 183.4 275.2
56 150.7 123.0 184.6









#35
 4-Chloro-3-methylphenol
 Concen: 22.444 ng/ul
 RT: 12.013 min Scan# 1519
 Delta R.T. -0.000 min
 Lab File: BG057943.D
 Acq: 1 Jul 2023 15:20




Instrument :
 BNA_G
 ClientSampleId :
 SICV441


Tgt Ion:107 Resp: 78279
 Ion Ratio Lower Upper
 107 100
 144 24.9 19.0 28.4
 142 74.1 58.4 87.6



#36
 2-Methylnaphthalene
 Concen: 22.497 ng/ul
 RT: 12.389 min Scan# 1583
 Delta R.T. -0.000 min
 Lab File: BG057943.D
 Acq: 1 Jul 2023 15:20

Tgt Ion:142 Resp: 151449
 Ion Ratio Lower Upper
 142 100
 141 90.1 69.4 104.2





#37

1-Methylnaphthalene

Concen: 22.684 ng/ul

RT: 12.612 min Scan# 1

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Instrument :

BNA_G

ClientSampleId :

SICV441





Tgt Ion:142 Resp: 153134

Ion Ratio Lower Upper

142 100

141 94.2 73.8 110.8

116 3.7 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 23.157 ng/ul

RT: 12.759 min Scan# 1646

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Tgt Ion:216 Resp: 86887




Ion Ratio Lower Upper


216 100

214 81.3 64.1 96.1

179 19.8 16.0 24.0

108 14.5 12.1 18.1









#40
Hexachlorocyclopentadiene
Concen: 22.775 ng/ul
RT: 12.736 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:237 Resp: 53827
Ion Ratio Lower Upper
237 100
235 66.0 49.2 73.8
272 13.0 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 23.378 ng/ul
RT: 12.994 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:196 Resp: 59451
Ion Ratio Lower Upper
196 100
198 93.0 73.5 110.3
200 30.8 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 23.222 ng/ul
RT: 13.070 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:196 Resp: 63866
Ion Ratio Lower Upper
196 100
198 89.3 77.2 115.8
200 30.3 23.8 35.6



#43
1,1'-Biphenyl
Concen: 23.051 ng/ul
RT: 13.399 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:154 Resp: 204529
Ion Ratio Lower Upper
154 100
153 41.4 32.6 49.0
76 13.7 11.6 17.4









#44
2-Chloronaphthalene
Concen: 23.165 ng/ul
RT: 13.441 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:162 Resp: 164957
Ion Ratio Lower Upper
162 100
164 34.2 27.7 41.5
127 39.6 30.9 46.3



#45
2-Nitroaniline
Concen: 23.528 ng/ul
RT: 13.646 min Scan# 1797
Delta R.T. 0.006 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion: 65 Resp: 58252
Ion Ratio Lower Upper
65 100
92 63.8 53.2 79.8
138 95.0 71.1 106.7









#47
Dimethylphthalate
Concen: 22.810 ng/ul
RT: 14.016 min Scan# 1860
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:163 Resp: 222625
Ion Ratio Lower Upper
163 100
194 3.6 3.0 4.4
164 10.8 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 23.153 ng/ul
RT: 14.134 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:165 Resp: 46384
Ion Ratio Lower Upper
165 100
89 59.4 42.7 64.1
121 19.8 14.8 22.2









#50
Acenaphthylene
Concen: 23.163 ng/ul
RT: 14.287 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:152 Resp: 265157
Ion Ratio Lower Upper
152 100
151 20.5 15.6 23.4
153 14.0 10.5 15.7



#51
3-Nitroaniline
Concen: 19.541 ng/ul
RT: 14.469 min Scan# 1937
Delta R.T. 0.006 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:138 Resp: 33084
Ion Ratio Lower Upper
138 100
108 10.7 8.2 12.2
92 114.1 100.2 150.4









#52
Acenaphthene
Concen: 22.793 ng/ul
RT: 14.627 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:153 Resp: 179046
Ion Ratio Lower Upper
153 100
152 48.7 36.5 54.7
154 88.3 70.5 105.7



#53
2,4-Dinitrophenol
Concen: 21.756 ng/ul
RT: 14.674 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:184 Resp: 22723
Ion Ratio Lower Upper
184 100
63 64.0 51.1 76.7
154 48.8 43.3 64.9









#55
4-Nitrophenol
Concen: 23.292 ng/ul
RT: 14.774 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:109 Resp: 29615
Ion Ratio Lower Upper
109 100
139 145.5 119.0 178.4
65 158.6 135.5 203.3



#56
Dibenzofuran
Concen: 23.302 ng/ul
RT: 14.962 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:168 Resp: 260201
Ion Ratio Lower Upper
168 100
139 33.0 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 23.921 ng/ul
RT: 14.921 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:165 Resp: 67898
Ion Ratio Lower Upper
165 100
63 43.3 35.8 53.8
182 2.1 1.6 2.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.104 ng/ul
RT: 15.186 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:232 Resp: 57845
Ion Ratio Lower Upper
232 100
131 36.8 30.5 45.7
130 2.4 1.4 2.0#
166 30.3 22.6 34.0









#59
Diethylphthalate
Concen: 22.853 ng/ul
RT: 15.379 min Scan# 2092
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:149 Resp: 232802
Ion Ratio Lower Upper
149 100
177 22.4 17.8 26.8
150 12.2 10.1 15.1



#61
Fluorene
Concen: 23.195 ng/ul
RT: 15.609 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:166 Resp: 212516
Ion Ratio Lower Upper
166 100
165 100.6 79.1 118.7
167 14.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 23.201 ng/ul
RT: 15.603 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:204 Resp: 112477
Ion Ratio Lower Upper
204 100
206 33.6 25.8 38.6
141 50.3 41.8 62.6



#63
4-Nitroaniline
Concen: 17.241 ng/ul
RT: 15.626 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:138 Resp: 26148
Ion Ratio Lower Upper
138 100
92 50.9 43.8 65.8
108 75.7 58.8 88.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 23.201 ng/ul
 RT: 15.685 min Scan# 2144
 Delta R.T. -0.000 min
 Lab File: BG057943.D
 Acq: 1 Jul 2023 15:20




Instrument :
 BNA_G
 ClientSampleId :
 SICV441


Tgt Ion:198 Resp: 38276
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.1 4.7
 77 19.9 17.9 26.9



#67
 N-Nitrosodiphenylamine
 Concen: 22.984 ng/ul
 RT: 15.814 min Scan# 2166
 Delta R.T. -0.000 min
 Lab File: BG057943.D
 Acq: 1 Jul 2023 15:20

Tgt Ion:169 Resp: 181725
 Ion Ratio Lower Upper
 169 100
 168 70.6 54.2 81.2
 167 37.6 30.8 46.2









#68
4-Bromophenyl-phenylether
Concen: 23.083 ng/ul
RT: 16.496 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:248 Resp: 75277
Ion Ratio Lower Upper
248 100
250 95.9 78.3 117.5
141 59.0 49.1 73.7



#69
Hexachlorobenzene
Concen: 23.315 ng/ul
RT: 16.613 min Scan# 2302
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:284 Resp: 85128
Ion Ratio Lower Upper
284 100
142 27.3 21.9 32.9
249 33.3 25.0 37.6









#70
Atrazine
Concen: 22.803 ng/ul
RT: 16.760 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:200 Resp: 73728
Ion Ratio Lower Upper
200 100
173 27.1 20.0 30.0
215 46.8 35.8 53.8



#71
Pentachlorophenol
Concen: 22.197 ng/ul
RT: 16.954 min Scan# 2360
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:266 Resp: 48655
Ion Ratio Lower Upper
266 100
264 66.1 51.5 77.3
268 64.7 49.0 73.4









#72
Phenanthrene
Concen: 23.002 ng/ul
RT: 17.348 min Scan# 2427
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:178 Resp: 347607
Ion Ratio Lower Upper
178 100
179 16.0 12.8 19.2
176 19.3 15.7 23.5



#74
Anthracene
Concen: 22.966 ng/ul
RT: 17.442 min Scan# 2443
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:178 Resp: 351213
Ion Ratio Lower Upper
178 100
179 15.6 12.6 18.8
176 19.4 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.432 ng/uL
RT: 13.364 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:216 Resp: 92325
Ion Ratio Lower Upper
216 100
214 81.0 64.8 97.2
218 48.2 39.1 58.7



#76
Pentachlorobenzene
Concen: 23.846 ng/uL
RT: 14.880 min Scan# 2007
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:250 Resp: 100347
Ion Ratio Lower Upper
250 100
248 61.4 52.8 79.2
252 60.3 52.4 78.6









#77
Carbazole
Concen: 23.972 ng/ul
RT: 17.706 min Scan# 2489
Delta R.T. -0.006 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:167 Resp: 331556
Ion Ratio Lower Upper
167 100
166 21.9 17.5 26.3
139 12.0 9.5 14.3



#78
Di-n-butylphthalate
Concen: 22.984 ng/ul
RT: 18.264 min Scan# 2583
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:149 Resp: 404946
Ion Ratio Lower Upper
149 100
150 9.6 7.4 11.2
104 4.8 3.9 5.9









#80
Fluoranthene
Concen: 22.664 ng/ul
RT: 19.357 min Scan# 2769
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:202 Resp: 429970
Ion Ratio Lower Upper
202 100
101 11.8 9.1 13.7
100 8.7 7.1 10.7



#82
Pyrene
Concen: 22.734 ng/ul
RT: 19.721 min Scan# 2831
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:202 Resp: 434156
Ion Ratio Lower Upper
202 100
101 13.6 10.2 15.2
100 11.0 8.9 13.3





#83

Butylbenzylphthalate

Concen: 22.933 ng/ul

RT: 20.609 min Scan# 2982

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Instrument :

BNA_G

ClientSampleId :

SICV441





Tgt Ion:149 Resp: 177590

Ion Ratio Lower Upper

149 100

91 68.7 56.5 84.7

206 20.9 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 24.057 ng/ul

RT: 21.461 min Scan# 3127

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20




Tgt Ion:252 Resp: 151404


Ion Ratio Lower Upper

252 100

254 63.5 50.7 76.1

126 9.9 6.6 10.0





#85

Benzo(a)anthracene

Concen: 22.747 ng/ul

RT: 21.543 min Scan# 3141

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Instrument :

BNA_G

ClientSampleId :

SICV441





Tgt Ion:228 Resp: 424589

Ion Ratio Lower Upper

228 100

229 20.6 16.0 24.0

226 28.2 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.765 ng/ul

RT: 21.449 min Scan# 3125

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20




Tgt Ion:149 Resp: 252633


Ion Ratio Lower Upper

149 100

167 27.0 20.9 31.3

279 3.7 3.1 4.7





#87

Chrysene

Concen: 22.911 ng/ul

RT: 21.602 min Scan# 3151

Delta R.T. -0.006 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Instrument :

BNA_G

ClientSampleId :

SICV441





Tgt Ion:228 Resp: 400896

Ion Ratio Lower Upper

228 100

226 30.7 24.4 36.6

229 19.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 22.802 ng/ul




RT: 22.636 min Scan# 3327


Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Tgt Ion:149 Resp: 445716









#90
Benzo(b)fluoranthene
Concen: 21.983 ng/ul
RT: 23.711 min Scan# 3510
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20




Instrument :
BNA_G
ClientSampleId :
SICV441


Tgt Ion:252 Resp: 433038
Ion Ratio Lower Upper
252 100
253 22.5 18.1 27.1
125 8.8 7.5 11.3



#91
Benzo(k)fluoranthene
Concen: 22.645 ng/ul
RT: 23.776 min Scan# 3521
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:252 Resp: 437235
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 8.9 7.4 11.2





#93

Benzo(a)pyrene

Concen: 22.328 ng/ul

RT: 24.569 min Scan# 3

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20

Instrument :

BNA_G

ClientSampleId :

SICV441





Tgt Ion:252 Resp: 389292

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 10.5 8.3 12.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.284 ng/ul

RT: 28.306 min Scan# 4292

Delta R.T. -0.000 min

Lab File: BG057943.D

Acq: 1 Jul 2023 15:20




Tgt Ion:276 Resp: 515918


Ion Ratio Lower Upper

276 100

138 16.8 14.0 21.0

277 25.1 19.4 29.2









#95
Dibenzo(a,h)anthracene
Concen: 22.259 ng/ul
RT: 28.370 min Scan# 41
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Instrument :
BNA_G
ClientSampleId :
SICV441

Tgt Ion:278 Resp: 424969
Ion Ratio Lower Upper
278 100
139 13.3 11.3 16.9
279 22.1 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 22.003 ng/ul
RT: 29.428 min Scan# 4483
Delta R.T. -0.000 min
Lab File: BG057943.D
Acq: 1 Jul 2023 15:20

Tgt Ion:276 Resp: 416091
Ion Ratio Lower Upper
276 100
138 18.3 14.2 21.2
277 23.3 18.6 27.8

