


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG070218\
 Data File : BG035569.D
 Acq On : 3 Jul 2018 3:27
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG060718.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 08 17:16:28 2018



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.1	13295	PASS
68	69	0.00	2	1.9	332	PASS
69	198	0.00	100	32.1	17668	PASS
70	69	0.00	2	0.7	117	PASS
127	198	10	80	36.2	19910	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	55064	PASS
199	198	5	9	6.8	3756	PASS
275	198	10	60	30.7	16913	PASS
365	198	1	100	3.8	2086	PASS
441	443	0.01	100	80.3	6006	PASS
442	198	50	100	70.8	38978	PASS
443	442	15	24	19.2	7476	PASS