

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070323\
 Data File : BG057946.D
 Acq On : 3 Jul 2023 8:43
 Operator : CG/JU
 Sample : PB153469BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK469

Quant Time: Jul 04 02:14:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 07:50:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.930	152	29190	20.000	ng/u1	0.00
20) Naphthalene-d8	10.732	136	136826	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.563	164	98871	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.307	188	259126	20.000	ng/u1	0.00
79) Chrysene-d12	21.555	240	244224	20.000	ng/u1	0.00
88) Perylene-d12	24.710	264	285186	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	5508	6.607	ng/uL	0.00
4) Pyridine-d5	3.776	84	86059	34.005	ng/u1	0.00
7) Phenol-d5	7.089	99	97610	32.744	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.260	67	64149	35.709	ng/u1	0.00
11) 2-Chlorophenol-d4	7.459	132	74215	36.097	ng/u1	0.00
15) 4-Methylphenol-d8	8.635	113	76593	34.051	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	38498	34.521	ng/u1	0.00
24) 2-Nitrophenol-d4	9.810	143	39991	34.120	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.350	165	71650	31.867	ng/u1	0.00
31) 4-Chloroaniline-d4	10.867	131	112025	32.862	ng/u1	0.00
46) Dimethylphthalate-d6	13.969	166	288982	37.044	ng/u1	0.00
49) Acenaphthylene-d8	14.257	160	299440	35.175	ng/u1	0.00
54) 4-Nitrophenol-d4	14.757	143	43926	31.871	ng/u1	0.00
60) Fluorene-d10	15.556	176	238716	36.032	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.668	200	36825	26.172	ng/u1	0.00
73) Anthracene-d10	17.407	188	424480	35.302	ng/u1	0.00
81) Pyrene-d10	19.692	212	549500	35.725	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.498	264	546849	37.637	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070323\
Data File : BG057946.D
Acq On : 3 Jul 2023 8:43
Operator : CG/JU
Sample : PB153469BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK469

Quant Time: Jul 04 02:14:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Jul 02 07:50:15 2023
Response via : Initial Calibration

