


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070323\
Data File : BG057944.D
Acq On : 3 Jul 2023 6:32
Operator : CG/JU
Sample : DFTPP512
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP512

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 04 05:09:13 2023



AutoFind: Scans 2492, 2493, 2494; Background Corrected with Scan 2487

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	37.8	42693	PASS
68	69	0.00	2	1.0	376	PASS
69	198	0.00	100	32.6	36805	PASS
70	69	0.00	2	0.6	210	PASS
127	198	10	80	38.9	43925	PASS
197	198	0.00	2	0.3	286	PASS
198	198	100	100	100.0	112989	PASS
199	198	5	9	6.8	7636	PASS
275	198	10	60	27.7	31272	PASS
365	198	1	100	3.4	3833	PASS
441	198	0.01	100	13.8	15625	PASS
442	198	50	100	80.7	91152	PASS
443	442	15	24	19.8	18083	PASS

m/z	Abundance
43.90	1434.0
206.90	409.0

Instrument :

BNA_G

ClientSampleId :

DFTPP512

m/z	Abundance
38.00	329.0
39.00	2672.0
40.00	244.0
41.00	219.0
43.90	1369.0
49.00	258.0
50.00	8932.0
51.00	36008.0
52.00	1753.0
52.90	182.0
55.00	208.0
55.90	1203.0
56.90	2157.0
61.00	420.0
61.90	591.0
63.00	1523.0
64.00	276.0
65.00	882.0
68.00	513.0
68.90	31624.0
70.00	230.0
72.90	366.0
74.00	3481.0
75.00	5384.0
76.00	1673.0
77.00	39200.0
78.00	2727.0
79.00	2315.0
79.90	2018.0
80.90	2544.0
82.00	722.0
83.00	502.0
84.90	448.0
85.90	635.0
86.90	277.0
90.90	410.0
91.90	574.0
92.90	3549.0
94.00	272.0
95.80	322.0
97.90	2947.0
98.90	2368.0
100.00	364.0
101.00	1554.0
103.00	258.0
103.90	897.0
105.00	727.0
106.00	265.0
107.00	11031.0
108.00	1404.0
109.90	21144.0
111.00	3213.0
112.00	284.0
115.90	608.0
117.00	7168.0
117.90	502.0
121.90	581.0
123.00	856.0
123.90	427.0
124.90	526.0
127.00	35016.0
127.90	2968.0
128.90	12595.0
130.00	903.0
130.90	389.0
131.90	156.0
133.90	449.0
135.00	1148.0
135.90	567.0
136.90	732.0
137.80	205.0
140.00	213.0
140.90	1647.0
141.90	674.0
142.90	398.0
145.90	440.0
146.90	831.0
148.00	1524.0
149.00	482.0
151.10	294.0
151.70	152.0
153.00	508.0
153.90	355.0
155.00	1121.0
156.00	1594.0
157.00	247.0
157.90	244.0
158.90	431.0
160.00	458.0
161.00	842.0
161.80	339.0
164.90	972.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

166.00	541.0
166.90	3102.0
167.90	1993.0
169.00	326.0
170.10	159.0
170.90	248.0
172.00	241.0
173.00	461.0
174.00	783.0
175.00	1433.0
176.00	606.0
177.00	743.0
177.90	234.0
178.90	2693.0
180.00	1979.0
180.90	936.0
182.00	156.0
184.00	252.0
185.00	1558.0
186.00	10158.0
187.00	3187.0
188.00	376.0
188.90	681.0
189.90	194.0
191.00	422.0
192.00	990.0
192.90	1095.0
193.90	175.0
194.90	179.0
196.00	2172.0
197.90	81944.0
198.90	6013.0
200.00	456.0
201.20	196.0
201.90	190.0
202.90	555.0
204.00	2869.0
205.00	4937.0
206.00	21800.0
207.00	3355.0
207.90	721.0
209.00	257.0
209.90	285.0
211.00	709.0
216.00	388.0
216.90	5098.0
218.00	687.0
221.00	3305.0
222.00	712.0
222.90	1099.0
224.00	11672.0
225.00	3500.0
226.00	442.0
227.00	4541.0
228.00	590.0
228.90	1016.0
230.90	309.0
234.00	332.0
235.00	319.0
235.90	265.0
236.90	395.0
238.90	271.0
240.90	164.0
242.00	566.0
243.00	440.0
244.00	8707.0
245.00	1477.0
246.00	1619.0
246.80	255.0
249.00	404.0
252.80	203.0
255.00	46720.0
255.90	6960.0
257.00	691.0
258.00	2654.0
258.90	323.0
264.90	980.0
265.90	263.0
271.70	157.0
272.90	1558.0
274.00	3561.0
275.00	20944.0
276.00	2312.0
276.90	1671.0
278.00	336.0
282.90	169.0
285.00	266.0
293.00	511.0
294.00	156.0
296.00	5496.0
296.90	830.0
302.80	629.0
303.70	159.0
314.00	389.0
314.90	611.0
315.90	392.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

323.00	1756.0
323.90	433.0
326.90	220.0
327.80	247.0
332.90	261.0
333.90	1228.0
335.00	316.0
341.00	383.0
346.10	407.0
352.00	762.0
353.00	475.0
353.90	531.0
365.00	2261.0
366.00	394.0
370.80	175.0
372.00	948.0
372.90	256.0
383.10	269.0
401.90	323.0
403.10	519.0
403.80	307.0
420.80	412.0
421.90	368.0
423.00	2729.0
423.90	1035.0
441.00	9090.0
442.00	50928.0
443.00	10058.0
444.00	853.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

m/z	Abundance
37.00	336.0
38.00	523.0
39.00	3235.0
40.00	218.0
40.90	207.0
44.00	1415.0
50.00	11172.0
51.00	50680.0
52.00	2551.0
54.90	402.0
56.00	1353.0
57.00	3606.0
58.00	223.0
61.00	554.0
62.00	823.0
62.90	1760.0
64.00	216.0
64.90	672.0
68.90	42392.0
70.00	183.0
73.00	381.0
74.00	4145.0
75.00	7072.0
76.00	2311.0
77.00	52120.0
78.00	3480.0
79.00	2737.0
79.90	2484.0
81.00	3377.0
82.00	964.0
82.90	858.0
85.00	723.0
85.90	1046.0
87.00	497.0
87.90	237.0
90.90	762.0
91.90	921.0
93.00	4925.0
93.90	427.0
95.00	233.0
96.00	301.0
97.90	3821.0
98.90	3132.0
100.00	280.0
100.90	1927.0
103.00	703.0
103.90	1088.0
104.90	1150.0
106.00	386.0
107.00	15223.0
107.90	2707.0
109.90	29480.0
111.00	4060.0
111.90	436.0
115.90	670.0
117.00	10729.0
117.90	544.0
118.90	158.0
122.00	1058.0
122.90	1679.0
124.00	710.0
124.90	737.0
127.00	52464.0
128.00	4133.0
129.00	19016.0
129.90	1268.0
131.00	210.0
132.00	283.0
133.90	576.0
135.00	1849.0
136.00	571.0
136.90	872.0
137.80	318.0
138.90	150.0
140.00	500.0
141.00	2284.0
142.00	961.0
142.90	701.0
144.10	198.0
145.90	486.0
147.00	1203.0
147.90	2846.0
149.00	589.0
149.90	194.0
150.80	341.0
151.90	175.0
152.90	827.0
153.90	602.0
155.00	1801.0
156.00	2542.0
157.00	533.0
157.90	551.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

159.00	498.0
159.90	781.0
161.00	1233.0
161.90	391.0
165.00	800.0
166.00	745.0
167.00	5178.0
168.00	2797.0
168.80	398.0
170.20	183.0
170.90	217.0
172.00	485.0
172.80	552.0
174.00	1393.0
175.00	2264.0
175.90	650.0
177.00	1044.0
178.00	451.0
178.90	4075.0
180.00	3252.0
181.00	1508.0
182.00	238.0
184.00	503.0
184.90	2389.0
186.00	15623.0
187.00	4671.0
188.00	560.0
189.00	1079.0
191.00	579.0
191.90	1536.0
192.90	1565.0
194.00	444.0
195.00	377.0
195.90	3454.0
196.70	859.0
197.90	131648.0
198.90	8859.0
199.90	893.0
201.40	601.0
203.00	899.0
204.00	4715.0
205.00	7466.0
206.00	34528.0
207.00	5341.0
208.00	1249.0
208.90	309.0
210.00	534.0
211.00	1211.0
214.90	513.0
215.90	675.0
216.90	8430.0
218.00	1190.0
220.90	4723.0
221.90	1171.0
223.00	2052.0
224.00	19544.0
225.00	5119.0
226.00	492.0
226.90	8173.0
227.90	1118.0
228.90	1445.0
230.00	226.0
230.90	715.0
234.00	435.0
234.90	486.0
235.90	461.0
237.00	776.0
239.00	333.0
239.90	299.0
241.00	647.0
242.00	969.0
243.00	1049.0
244.00	15139.0
245.00	2073.0
245.90	2834.0
246.90	766.0
248.90	705.0
249.80	197.0
251.00	162.0
252.10	182.0
253.10	404.0
255.00	80840.0
256.00	12079.0
257.00	867.0
257.90	4264.0
259.00	701.0
260.80	214.0
263.80	178.0
265.00	1711.0
266.00	261.0
269.80	158.0
271.00	208.0
271.90	262.0
272.90	2419.0
274.00	6163.0
275.00	36608.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

276.00	4642.0
277.00	2817.0
278.00	502.0
280.00	160.0
282.00	155.0
283.00	298.0
284.00	331.0
284.90	454.0
290.00	165.0
292.10	225.0
293.00	574.0
293.90	340.0
296.00	10217.0
296.90	1447.0
302.00	172.0
302.90	1495.0
303.90	336.0
310.00	173.0
313.90	430.0
314.90	1195.0
316.00	531.0
320.90	366.0
321.90	198.0
323.00	3371.0
324.00	736.0
326.80	665.0
328.00	404.0
331.90	239.0
332.90	416.0
334.00	2236.0
335.00	659.0
340.90	374.0
345.90	597.0
352.00	1282.0
352.90	930.0
354.00	1079.0
354.90	233.0
364.90	4395.0
365.80	639.0
371.00	264.0
372.00	1891.0
373.00	436.0
383.00	438.0
390.00	215.0
390.90	231.0
392.10	162.0
402.00	815.0
402.90	1033.0
404.00	323.0
420.90	725.0
422.00	935.0
423.00	6295.0
423.90	1220.0
425.00	230.0
441.00	17512.0
442.00	100440.0
443.00	20528.0
444.00	1742.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

m/z	Abundance
37.10	219.0
38.00	314.0
39.00	2735.0
40.00	258.0
43.90	1276.0
50.00	9709.0
51.00	41392.0
52.00	2070.0
54.90	249.0
55.90	1217.0
56.90	2788.0
57.90	152.0
61.00	544.0
62.00	495.0
63.00	1738.0
63.90	326.0
65.00	760.0
68.00	615.0
68.90	36400.0
70.00	216.0
72.90	357.0
74.00	3783.0
75.00	5533.0
76.00	1848.0
77.00	45792.0
78.00	2878.0
78.90	2461.0
79.90	1934.0
81.00	2897.0
81.90	928.0
82.90	723.0
84.90	713.0
85.90	797.0
87.00	355.0
91.00	702.0
92.00	983.0
93.00	4233.0
94.00	293.0
95.90	195.0
98.00	3226.0
99.00	2760.0
99.90	277.0
100.90	1463.0
101.90	153.0
103.00	517.0
104.00	1076.0
104.90	1162.0
106.00	374.0
106.90	13132.0
108.00	2162.0
109.90	25152.0
111.00	3809.0
111.90	336.0
112.90	171.0
116.00	765.0
117.00	8641.0
118.00	752.0
119.90	178.0
120.90	167.0
121.90	654.0
123.00	1207.0
123.90	678.0
125.00	642.0
127.00	44296.0
128.00	3363.0
129.00	16848.0
130.00	1511.0
130.90	401.0
133.90	443.0
134.90	1519.0
135.90	592.0
137.00	656.0
137.80	289.0
139.80	245.0
140.90	2373.0
142.00	730.0
143.00	526.0
145.90	365.0
147.00	887.0
147.90	2243.0
148.90	564.0
149.90	198.0
150.80	256.0
151.20	198.0
153.00	899.0
154.00	403.0
155.00	1418.0
156.00	2300.0
156.90	465.0
157.90	414.0
158.90	431.0
159.90	784.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

160.90	1033.0
161.90	430.0
162.90	195.0
165.00	1153.0
165.90	963.0
167.00	4742.0
168.00	2767.0
168.90	494.0
169.90	192.0
171.00	221.0
171.90	378.0
172.90	557.0
174.00	980.0
175.00	2336.0
176.00	492.0
176.90	888.0
178.00	552.0
179.00	3854.0
180.00	2614.0
180.90	1149.0
182.10	307.0
183.00	184.0
183.90	305.0
184.90	2099.0
186.00	15276.0
187.00	4629.0
188.00	473.0
188.90	951.0
189.90	220.0
191.00	539.0
191.90	1299.0
193.00	1485.0
194.00	313.0
196.00	3445.0
197.90	125376.0
198.90	8037.0
200.00	544.0
201.50	572.0
203.00	921.0
204.00	4453.0
205.00	7486.0
206.00	32504.0
207.00	4999.0
208.00	1211.0
209.00	365.0
210.10	599.0
211.00	1056.0
214.90	290.0
216.00	728.0
216.90	8295.0
217.90	1015.0
221.00	5246.0
222.90	2298.0
224.00	18696.0
225.00	4613.0
226.00	525.0
227.00	7635.0
228.00	1175.0
228.90	1682.0
229.90	260.0
231.00	710.0
233.00	181.0
233.80	399.0
234.90	673.0
235.90	515.0
237.00	830.0
238.90	299.0
240.00	189.0
240.90	509.0
242.00	896.0
242.90	828.0
244.00	16350.0
245.00	1995.0
246.00	2624.0
246.90	715.0
247.90	165.0
249.00	643.0
251.10	166.0
252.00	223.0
253.00	479.0
255.00	81648.0
256.00	11629.0
257.00	845.0
257.90	4355.0
259.00	737.0
259.90	158.0
260.80	180.0
265.00	1946.0
265.90	323.0
271.00	210.0
272.00	320.0
272.90	2443.0
274.00	7293.0
275.00	36264.0
276.00	5235.0
276.90	2997.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512

277.80	473.0
281.90	154.0
282.90	420.0
284.00	460.0
285.00	638.0
288.90	215.0
291.90	158.0
293.00	920.0
294.00	306.0
295.90	10721.0
297.00	1434.0
300.90	220.0
302.00	282.0
303.00	1140.0
304.00	301.0
313.90	505.0
314.90	915.0
316.00	815.0
316.90	273.0
320.90	390.0
322.00	202.0
323.00	3537.0
324.10	510.0
326.90	522.0
328.00	412.0
331.90	338.0
333.00	368.0
334.00	2519.0
335.00	572.0
341.00	534.0
345.90	753.0
346.90	152.0
352.00	1314.0
352.90	676.0
354.00	1372.0
354.90	304.0
364.90	4843.0
365.90	730.0
371.10	256.0
372.00	2249.0
372.90	660.0
382.90	756.0
384.00	236.0
390.00	253.0
391.00	211.0
402.00	993.0
402.90	1231.0
403.90	468.0
420.90	1076.0
422.00	946.0
423.00	6879.0
424.00	1485.0
441.00	20272.0
442.00	122088.0
443.00	23664.0
444.00	2173.0

Instrument :
BNA_G
ClientSampleId :
DFTPP512