

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070523\
 Data File : BG058072.D
 Acq On : 7 Jul 2023 12:44
 Operator : CG/JU
 Sample : PB153594BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK594

Quant Time: Jul 07 16:22:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 05 23:20:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.918	152	33606	20.000	ng/ul	0.00
20) Naphthalene-d8	10.720	136	154785	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.557	164	118716	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.301	188	316103	20.000	ng/ul	0.00
79) Chrysene-d12	21.555	240	296335	20.000	ng/ul	0.00
88) Perylene-d12	24.710	264	367542	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	6869	7.157	ng/uL	-0.01
4) Pyridine-d5	3.764	84	96121	32.990	ng/ul	-0.01
7) Phenol-d5	7.077	99	106328	30.981	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.242	67	71770	34.701	ng/ul	0.00
11) 2-Chlorophenol-d4	7.448	132	80589	34.046	ng/ul	-0.01
15) 4-Methylphenol-d8	8.623	113	85852	33.152	ng/ul	-0.01
21) Nitrobenzene-d5	9.075	128	43876	34.779	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.798	143	47224	35.617	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.344	165	81040	31.861	ng/ul	0.00
31) 4-Chloroaniline-d4	10.855	131	130094	33.735	ng/ul	0.00
46) Dimethylphthalate-d6	13.958	166	341321	36.439	ng/ul	0.00
49) Acenaphthylene-d8	14.251	160	353574	34.591	ng/ul	0.00
54) 4-Nitrophenol-d4	14.757	143	52389	31.657	ng/ul	-0.01
60) Fluorene-d10	15.544	176	286106	35.966	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.662	200	51236	29.850	ng/ul	-0.01
73) Anthracene-d10	17.401	188	516705	35.227	ng/ul	0.00
81) Pyrene-d10	19.686	212	638768	34.226	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.492	264	663197	35.417	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070523\
Data File : BG058072.D
Acq On : 7 Jul 2023 12:44
Operator : CG/JU
Sample : PB153594BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK594

Quant Time: Jul 07 16:22:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 05 23:20:13 2023
Response via : Initial Calibration

