




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070722\
Data File : BG054220.D
Acq On : 7 Jul 2022 10:50
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 07 13:41:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration



TIC: BG054220.D\data.ms

(70) Pentachlorophenol (C)




17.090min (-0.005) 28483.77 ng

response 69563

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	70.20	60.61
264.00	66.20	68.48
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070722\
Data File : BG054220.D
Acq On : 7 Jul 2022 10:50
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 07 13:41:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration



TIC: BG054220.D\data.ms

(77) Benzidine

19.658min (-0.005) 0.00 ng

response 433027

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.43
183.00	11.10	12.47
0.00	0.00	0.00