

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070823\
 Data File : BG058119.D
 Acq On : 9 Jul 2023 9:51
 Operator : CG/JU
 Sample : 03349-09
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 YBW02

Quant Time: Jul 10 00:40:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 08 10:01:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.914	152	53097	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	236443	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.553	164	175362	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.297	188	450327	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	409817	20.000	ng/u1	0.00
88) Perylene-d12	24.706	264	495151	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.354	96	4385	2.892	ng/uL	0.00
4) Pyridine-d5	3.766	84	70126	15.233	ng/u1	0.00
7) Phenol-d5	7.079	99	87981	16.225	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.238	67	54888	16.797	ng/u1	0.00
11) 2-Chlorophenol-d4	7.444	132	61609	16.473	ng/u1	0.00
15) 4-Methylphenol-d8	8.619	113	68803	16.816	ng/u1	0.00
21) Nitrobenzene-d5	9.077	128	34376	17.838	ng/u1	0.00
24) 2-Nitrophenol-d4	9.800	143	38334	18.927	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	68709	17.684	ng/u1	0.00
31) 4-Chloroaniline-d4	10.851	131	101954	17.307	ng/u1	0.00
46) Dimethylphthalate-d6	13.954	166	268656	19.417	ng/u1	0.00
49) Acenaphthylene-d8	14.248	160	297940	19.733	ng/u1	0.00
54) 4-Nitrophenol-d4	14.753	143	33823	13.836	ng/u1	0.00
60) Fluorene-d10	15.546	176	240671	20.481	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	32653	13.353	ng/u1	0.00
73) Anthracene-d10	17.397	188	417631	19.986	ng/u1	0.00
81) Pyrene-d10	19.682	212	493874	19.135	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.488	264	512016	20.297	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070823\
Data File : BG058119.D
Acq On : 9 Jul 2023 9:51
Operator : CG/JU
Sample : 03349-09
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
YBW02

Quant Time: Jul 10 00:40:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 08 10:01:01 2023
Response via : Initial Calibration

