

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070823\
 Data File : BG058126.D
 Acq On : 9 Jul 2023 14:42
 Operator : CG/JU
 Sample : 03349-16
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 YBW09

Quant Time: Jul 10 00:41:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 08 10:01:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.914	152	69163	20.000	ng/u1	0.00
20) Naphthalene-d8	10.716	136	320696	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.553	164	243291	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.297	188	598915	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	543591	20.000	ng/u1	0.00
88) Perylene-d12	24.706	264	669862	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.360	96	6361	3.221	ng/uL	0.00
4) Pyridine-d5	3.766	84	90988	15.174	ng/u1	0.00
7) Phenol-d5	7.079	99	135660	19.206	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.238	67	85313	20.043	ng/u1	0.00
11) 2-Chlorophenol-d4	7.450	132	94035	19.303	ng/u1	0.00
15) 4-Methylphenol-d8	8.619	113	109850	20.611	ng/u1	0.00
21) Nitrobenzene-d5	9.077	128	53215	20.359	ng/u1	0.00
24) 2-Nitrophenol-d4	9.800	143	60482	22.017	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	108345	20.559	ng/u1	0.00
31) 4-Chloroaniline-d4	10.852	131	155777	19.497	ng/u1	0.00
46) Dimethylphthalate-d6	13.954	166	410662	21.393	ng/u1	0.00
49) Acenaphthylene-d8	14.248	160	462622	22.085	ng/u1	0.00
54) 4-Nitrophenol-d4	14.753	143	63627	18.761	ng/u1	0.00
60) Fluorene-d10	15.546	176	372278	22.836	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.664	200	60262	18.530	ng/u1	0.00
73) Anthracene-d10	17.397	188	620686	22.334	ng/u1	0.00
81) Pyrene-d10	19.688	212	740901	21.641	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.488	264	782022	22.914	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070823\
Data File : BG058126.D
Acq On : 9 Jul 2023 14:42
Operator : CG/JU
Sample : 03349-16
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
YBW09

Quant Time: Jul 10 00:41:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 08 10:01:01 2023
Response via : Initial Calibration

