


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070823\
Data File : BG058083.D
Acq On : 8 Jul 2023 7:27
Operator : CG/JU
Sample : DFTPP514
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP514

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070123.MA.M
Title : SVOA CALIBRATION
Last Update : Sat Jul 08 10:01:01 2023



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	41912	PASS
68	69	0.00	2	0.9	313	PASS
69	198	0.00	100	31.5	34755	PASS
70	69	0.00	2	0.6	208	PASS
127	198	10	80	37.7	41597	PASS
197	198	0.00	2	0.2	176	PASS
198	198	100	100	100.0	110288	PASS
199	198	5	9	6.6	7328	PASS
275	198	10	60	27.8	30629	PASS
365	198	1	100	3.5	3901	PASS
441	198	0.01	100	14.6	16147	PASS
442	198	50	100	86.7	95653	PASS
443	442	15	24	20.0	19136	PASS

m/z	Abundance
44.00	1187.0
207.00	172.0

Instrument :

BNA_G

ClientSampleId :

DFTPP514

m/z	Abundance
38.00	320.0
39.00	3049.0
40.00	231.0
43.90	798.0
48.90	381.0
50.00	9742.0
51.00	42160.0
52.00	2072.0
54.90	273.0
55.90	1150.0
57.00	3075.0
57.90	172.0
60.90	646.0
62.00	601.0
63.00	1611.0
63.90	301.0
65.10	909.0
67.00	185.0
68.90	34952.0
69.90	209.0
72.90	254.0
74.00	3499.0
75.00	5936.0
76.00	1709.0
77.00	42416.0
78.00	3010.0
79.00	2463.0
80.00	1932.0
80.90	2741.0
82.00	772.0
83.00	697.0
84.00	200.0
84.90	510.0
85.90	724.0
87.00	321.0
88.00	178.0
90.90	539.0
92.00	714.0
93.00	4268.0
93.90	255.0
98.00	2863.0
98.90	2756.0
99.90	178.0
100.90	1712.0
103.00	398.0
103.90	971.0
104.90	778.0
107.00	11553.0
108.00	2157.0
108.90	316.0
110.00	22680.0
110.90	3308.0
111.90	329.0
115.90	570.0
116.90	7889.0
117.90	595.0
119.90	252.0
121.80	511.0
122.90	1124.0
124.00	528.0
125.00	482.0
127.00	38128.0
128.00	3258.0
129.00	14958.0
129.90	1321.0
130.90	348.0
133.90	417.0
134.90	1204.0
135.90	519.0
136.90	604.0
137.80	186.0
139.90	170.0
140.90	2069.0
141.90	550.0
143.00	389.0
145.00	165.0
145.90	259.0
147.00	978.0
147.90	2204.0
148.90	413.0
150.90	208.0
152.00	187.0
152.90	729.0
154.00	494.0
155.00	770.0
156.00	2046.0
157.00	321.0
158.00	410.0
158.90	352.0
160.00	686.0
161.00	812.0
161.90	253.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514

163.70	173.0
164.90	783.0
166.00	511.0
167.00	3909.0
168.00	2750.0
169.00	537.0
171.00	173.0
171.90	389.0
173.00	498.0
173.90	920.0
174.90	1514.0
175.90	571.0
176.90	696.0
178.00	296.0
178.90	2871.0
180.00	2459.0
181.00	1113.0
182.00	214.0
183.90	209.0
185.00	1762.0
186.00	12019.0
187.00	3460.0
188.00	551.0
189.00	744.0
190.90	287.0
192.00	755.0
193.00	1169.0
193.90	211.0
194.90	309.0
196.00	2635.0
196.70	529.0
197.90	95840.0
198.90	6598.0
199.90	527.0
201.30	405.0
203.00	814.0
204.00	3689.0
204.90	5135.0
206.00	25256.0
207.00	3710.0
208.00	906.0
208.90	205.0
210.10	361.0
211.00	958.0
211.90	291.0
214.90	281.0
216.00	510.0
216.90	5993.0
218.00	982.0
219.80	237.0
221.00	3696.0
222.90	1245.0
224.00	14456.0
225.00	3924.0
226.00	401.0
227.00	5811.0
227.90	886.0
228.90	1165.0
229.90	187.0
230.90	564.0
234.00	373.0
235.00	483.0
236.00	243.0
236.90	468.0
239.00	287.0
240.00	209.0
240.90	340.0
242.00	787.0
242.90	866.0
244.00	10972.0
245.00	1434.0
245.90	2173.0
246.90	441.0
248.80	338.0
251.90	169.0
252.90	355.0
255.00	55168.0
256.00	8041.0
256.90	607.0
257.90	2809.0
259.00	484.0
260.00	194.0
264.90	1155.0
265.90	258.0
273.00	1774.0
273.90	4203.0
275.00	24640.0
276.00	3861.0
276.90	2035.0
278.00	354.0
282.90	322.0
284.90	387.0
293.00	683.0
293.90	170.0
296.00	6587.0
297.00	1016.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514

302.00	166.0
303.00	777.0
304.00	300.0
307.80	150.0
314.10	310.0
314.90	674.0
316.00	576.0
320.90	300.0
321.90	165.0
323.00	2248.0
324.10	271.0
326.90	254.0
332.90	224.0
334.00	1416.0
335.00	521.0
340.90	234.0
345.90	247.0
352.00	672.0
352.90	493.0
354.00	660.0
364.90	2940.0
366.00	358.0
370.90	168.0
372.00	1223.0
373.00	336.0
382.80	345.0
389.90	260.0
401.90	574.0
402.90	631.0
403.90	398.0
421.00	760.0
421.90	780.0
423.00	3601.0
424.00	1055.0
424.90	157.0
441.00	10504.0
442.00	65008.0
443.00	13393.0
444.00	1359.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514

m/z	Abundance
38.00	538.0
39.00	2887.0
40.00	270.0
43.90	1079.0
48.90	397.0
50.00	11772.0
51.00	48568.0
52.00	2807.0
54.90	234.0
56.00	1220.0
57.00	3644.0
61.00	549.0
62.00	594.0
63.00	2042.0
64.00	256.0
65.00	1090.0
68.00	939.0
68.90	40160.0
70.00	262.0
72.90	265.0
74.00	4253.0
75.00	6943.0
76.00	2286.0
77.00	50736.0
78.00	3116.0
79.00	2954.0
80.00	2326.0
81.00	3356.0
81.90	870.0
83.00	822.0
83.90	216.0
84.90	798.0
86.00	1034.0
87.00	503.0
87.80	236.0
90.90	640.0
92.00	793.0
93.00	4867.0
94.00	290.0
96.00	232.0
97.90	4336.0
98.90	3177.0
100.00	254.0
101.00	2310.0
103.00	599.0
103.90	1436.0
105.00	1029.0
105.90	426.0
107.00	15003.0
108.00	2245.0
109.90	28896.0
110.90	4073.0
112.00	418.0
116.00	745.0
117.00	10357.0
117.90	903.0
120.00	224.0
121.90	998.0
122.90	1268.0
123.90	827.0
124.90	653.0
127.00	50352.0
128.00	3766.0
129.00	18168.0
129.90	1447.0
130.90	340.0
131.90	185.0
133.90	444.0
135.00	1160.0
136.00	687.0
137.00	718.0
138.00	262.0
141.00	2485.0
142.00	893.0
143.00	640.0
146.00	326.0
147.00	1266.0
148.00	3166.0
148.90	661.0
150.00	196.0
150.90	164.0
151.90	181.0
153.00	1034.0
153.90	627.0
155.00	1209.0
156.00	2142.0
157.10	347.0
157.90	501.0
158.90	474.0
159.90	776.0
161.00	1284.0
161.90	218.0

Instrument :

BNA_G

ClientSampleId :

DFTPP514

165.00	814.0
165.90	1060.0
167.00	5244.0
168.00	2914.0
169.00	376.0
169.90	218.0
171.00	380.0
171.90	606.0
173.00	554.0
174.00	1291.0
175.00	2249.0
176.00	446.0
176.90	939.0
178.00	287.0
178.90	4361.0
180.00	3205.0
181.00	1293.0
181.90	230.0
182.90	173.0
183.90	418.0
184.90	2044.0
186.00	16992.0
187.00	4596.0
187.90	595.0
189.00	995.0
189.90	265.0
191.00	543.0
192.00	1488.0
193.00	1515.0
193.90	404.0
196.00	3512.0
197.90	131136.0
198.90	8543.0
200.00	772.0
201.40	663.0
202.90	977.0
204.00	4599.0
205.00	7660.0
206.00	34272.0
207.00	5068.0
208.00	1075.0
208.90	273.0
209.90	441.0
210.90	1210.0
211.90	419.0
215.00	525.0
216.00	1015.0
216.90	9353.0
217.90	1263.0
221.00	5124.0
222.00	1306.0
222.90	1974.0
224.00	20048.0
225.00	4905.0
225.90	632.0
227.00	7617.0
227.90	1319.0
228.90	1476.0
229.90	201.0
230.90	619.0
231.90	172.0
234.00	520.0
235.00	659.0
236.00	318.0
236.90	685.0
239.00	283.0
239.90	219.0
240.90	318.0
242.00	882.0
242.90	1146.0
244.00	15663.0
245.00	2223.0
246.00	3125.0
246.90	856.0
248.90	668.0
249.80	240.0
252.90	347.0
255.00	81528.0
256.00	12446.0
256.90	822.0
257.90	4256.0
258.90	858.0
264.90	1778.0
266.00	275.0
271.00	255.0
271.80	204.0
273.00	2927.0
274.00	6516.0
275.00	36168.0
276.00	5053.0
277.00	3035.0
277.90	435.0
278.90	180.0
283.00	329.0
284.00	201.0
285.00	651.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514

291.80	159.0
293.00	705.0
294.00	218.0
296.00	10271.0
297.00	1359.0
303.00	1364.0
304.10	269.0
307.80	189.0
309.70	181.0
313.00	167.0
314.10	807.0
314.90	1124.0
316.00	616.0
320.90	244.0
321.90	187.0
323.00	3380.0
323.90	528.0
326.90	681.0
327.80	206.0
332.00	243.0
332.90	309.0
334.00	2386.0
334.90	675.0
341.00	365.0
345.90	794.0
351.90	1134.0
352.90	770.0
354.00	1281.0
354.90	203.0
364.90	4084.0
366.00	637.0
371.00	290.0
372.00	1749.0
373.00	547.0
383.00	515.0
383.90	198.0
389.90	241.0
390.90	185.0
392.00	181.0
401.90	956.0
402.90	1290.0
403.90	380.0
421.00	781.0
422.00	903.0
423.00	6521.0
424.00	1636.0
425.00	154.0
441.00	19080.0
442.00	108224.0
443.00	21656.0
444.00	2168.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514

m/z	Abundance
38.00	400.0
39.00	2050.0
40.00	193.0
40.90	221.0
44.00	956.0
49.00	318.0
50.00	7700.0
51.00	35008.0
52.00	1829.0
55.00	238.0
55.90	1107.0
56.90	1945.0
61.00	404.0
62.00	536.0
63.00	1303.0
64.00	245.0
65.00	507.0
68.90	29152.0
70.10	154.0
73.00	234.0
74.00	3434.0
75.00	4507.0
76.00	1782.0
77.00	35784.0
78.00	2314.0
78.90	2039.0
79.90	1394.0
80.90	1813.0
82.00	678.0
83.00	628.0
85.00	698.0
86.00	634.0
87.00	324.0
87.90	230.0
90.90	444.0
92.00	497.0
93.00	3260.0
93.90	183.0
96.00	259.0
97.90	2588.0
98.90	2360.0
99.90	172.0
101.00	1215.0
102.90	390.0
104.00	822.0
104.90	663.0
106.00	317.0
107.00	10145.0
108.00	1623.0
109.90	21056.0
111.00	3068.0
112.00	258.0
115.90	857.0
116.90	7770.0
117.90	510.0
119.90	198.0
121.90	615.0
122.90	1146.0
123.90	532.0
124.90	576.0
127.00	36312.0
128.00	2681.0
128.90	13096.0
129.90	1230.0
130.90	224.0
132.00	179.0
134.00	447.0
134.90	1105.0
135.90	334.0
137.10	280.0
137.70	162.0
140.00	201.0
141.00	2105.0
141.90	553.0
142.80	322.0
144.00	183.0
145.90	317.0
147.00	1285.0
148.00	2466.0
149.00	517.0
150.80	172.0
151.40	205.0
151.60	204.0
153.00	726.0
154.00	503.0
155.00	1154.0
156.00	1647.0
156.90	455.0
157.90	363.0
159.00	351.0
159.90	491.0
161.00	1062.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514

162.00	166.0
163.00	159.0
164.90	825.0
165.90	552.0
167.00	3836.0
167.90	1900.0
168.90	292.0
169.80	168.0
171.00	272.0
171.90	387.0
173.00	458.0
174.00	956.0
175.00	1927.0
176.00	712.0
177.00	714.0
177.90	224.0
178.90	3158.0
180.00	2259.0
180.90	1114.0
184.00	291.0
185.00	1768.0
186.00	12950.0
187.00	3634.0
187.90	488.0
188.90	696.0
191.00	439.0
192.00	1256.0
193.00	1139.0
193.90	196.0
195.00	154.0
196.00	2918.0
197.90	103888.0
198.90	6843.0
200.00	563.0
201.50	406.0
203.00	643.0
203.90	3574.0
205.00	6772.0
206.00	27648.0
207.00	3692.0
208.00	901.0
208.90	280.0
210.00	458.0
210.90	1115.0
212.00	238.0
214.80	281.0
215.90	477.0
216.90	7495.0
217.90	937.0
218.80	179.0
221.00	3783.0
221.80	920.0
222.90	1988.0
224.00	16488.0
225.00	4214.0
226.00	457.0
227.00	6276.0
228.00	921.0
228.90	1118.0
230.00	223.0
230.90	676.0
234.00	453.0
234.90	260.0
235.90	399.0
237.00	651.0
238.90	260.0
241.00	388.0
242.00	809.0
243.00	844.0
244.00	13778.0
245.00	1651.0
245.90	2400.0
246.90	560.0
249.00	494.0
251.00	167.0
251.90	261.0
253.00	345.0
255.00	69152.0
256.00	9952.0
257.00	854.0
257.90	4316.0
258.90	738.0
259.90	159.0
260.90	163.0
264.90	1516.0
265.90	289.0
270.80	153.0
272.10	260.0
273.00	2159.0
273.90	5392.0
275.00	31080.0
276.00	4018.0
276.90	2657.0
277.90	542.0
282.90	449.0
284.00	162.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514

285.00	736.0
292.90	644.0
293.90	202.0
294.90	174.0
296.00	8482.0
296.90	1282.0
301.90	171.0
303.00	1210.0
303.90	270.0
309.90	186.0
314.00	487.0
314.90	855.0
316.00	557.0
316.80	216.0
321.00	276.0
323.00	3313.0
323.90	552.0
326.80	492.0
327.90	445.0
332.00	245.0
332.90	268.0
334.00	2565.0
334.90	673.0
341.00	433.0
345.90	683.0
352.00	1087.0
353.00	642.0
354.00	1140.0
355.00	213.0
364.90	4680.0
366.00	584.0
371.00	304.0
372.00	1641.0
373.00	487.0
382.80	457.0
383.90	183.0
391.00	187.0
401.90	752.0
403.00	968.0
404.00	569.0
420.90	1015.0
421.90	1034.0
423.00	6644.0
424.00	1649.0
441.00	18856.0
442.00	113728.0
443.00	22360.0
443.90	2114.0
445.00	159.0

Instrument :
BNA_G
ClientSampleId :
DFTPP514