



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070921\
 Data File : BG049222.D
 Acq On : 8 Jul 2021 18:22
 Operator : CG/JU
 Sample : DFTPP151
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP151

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG070921.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jul 09 03:15:38 2021



AutoFind: Scans 2528, 2529, 2530; Background Corrected with Scan 2523

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.6	25613	PASS
68	69	0.00	2	0.5	113	PASS
69	198	0.00	100	43.2	24792	PASS
70	69	0.00	2	0.6	148	PASS
127	198	10	80	48.6	27915	PASS
197	198	0.00	2	0.3	200	PASS
198	198	100	100	100.0	57445	PASS
199	198	5	9	6.0	3452	PASS
275	198	10	60	31.5	18106	PASS
365	198	1	100	4.9	2797	PASS
441	198	0.01	100	15.9	9153	PASS
442	198	50	100	100.0	57432	PASS
443	442	15	24	20.8	11967	PASS

m/z	Abundance
37.10	269.0
38.10	305.0
39.10	1260.0
40.10	280.0
44.20	297.0
50.10	5270.0
51.10	21304.0
52.10	1443.0
55.10	253.0
56.20	582.0
57.10	1286.0
61.20	153.0
62.10	292.0
63.10	826.0
65.10	535.0
68.00	340.0
69.10	19800.0
73.10	151.0
74.10	1744.0
75.10	3368.0
76.20	1209.0
77.10	21432.0
78.20	1578.0
79.10	1664.0
80.10	1268.0
81.00	1683.0
82.10	289.0
83.00	587.0
85.10	325.0
86.10	455.0
87.10	248.0
91.00	341.0
92.10	575.0
93.10	2988.0
93.80	224.0
98.10	2169.0
99.00	1884.0
100.10	401.0
101.10	1392.0
103.00	457.0
104.10	573.0
105.10	480.0
106.10	158.0
107.10	7179.0
108.10	1322.0
109.10	347.0
110.10	10074.0
111.10	1792.0
112.00	349.0
115.20	230.0
116.00	483.0
117.10	6498.0
118.00	440.0
122.10	591.0
123.10	984.0
124.00	512.0
125.10	347.0
127.10	22696.0
128.00	1428.0
129.10	9122.0
130.00	562.0
131.00	213.0
134.10	351.0
135.00	692.0
136.10	268.0
137.20	621.0
140.10	196.0
141.10	1590.0
142.00	295.0
142.90	244.0
146.10	190.0
147.10	665.0
148.10	1797.0
149.00	398.0
151.10	314.0
153.00	308.0
154.10	152.0
155.10	904.0
156.20	1374.0
157.10	196.0
158.00	400.0
159.10	202.0
160.20	433.0
161.10	895.0
162.20	182.0
165.00	295.0
166.10	375.0
167.10	2593.0
168.10	1004.0
169.20	191.0
171.90	225.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

173.00	211.0
174.20	489.0
175.10	1306.0
176.10	259.0
177.10	755.0
179.10	2073.0
180.10	1623.0
181.10	729.0
185.10	938.0
186.20	6642.0
187.10	1951.0
188.00	338.0
189.10	376.0
191.10	182.0
192.10	640.0
193.20	626.0
195.20	236.0
196.20	1054.0
198.10	40952.0
199.10	3085.0
200.30	169.0
203.00	327.0
204.20	1989.0
205.10	3535.0
206.10	12918.0
207.10	1786.0
208.10	434.0
209.10	197.0
210.10	198.0
211.10	618.0
216.00	366.0
217.20	3289.0
218.10	571.0
221.10	1718.0
222.10	571.0
223.10	633.0
224.20	7156.0
225.10	2106.0
227.10	3685.0
228.10	922.0
229.10	828.0
231.20	237.0
234.30	211.0
235.00	317.0
237.10	172.0
241.10	203.0
242.20	351.0
243.20	363.0
244.20	4862.0
245.30	545.0
246.20	1159.0
247.10	197.0
249.20	174.0
253.10	252.0
254.20	610.0
255.10	29072.0
256.10	4151.0
257.10	347.0
258.10	1629.0
259.10	348.0
265.10	748.0
273.20	1087.0
274.10	2031.0
275.10	11565.0
276.10	2076.0
277.10	1276.0
278.10	233.0
285.20	217.0
291.10	169.0
293.20	322.0
296.20	3810.0
297.10	722.0
302.10	177.0
303.10	279.0
314.20	336.0
315.10	470.0
316.20	204.0
321.10	260.0
323.20	1508.0
324.00	420.0
326.80	184.0
328.00	179.0
334.10	1077.0
335.00	270.0
352.00	405.0
353.00	284.0
354.20	644.0
365.10	1852.0
366.10	359.0
372.30	640.0
383.10	329.0
402.00	223.0
403.20	277.0
404.10	166.0
421.40	184.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

422.10	310.0
423.20	1603.0
424.40	361.0
441.20	4925.0
442.20	30912.0
443.20	6414.0
444.10	847.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

Scan 2523 (17.861 min): BG049222.D\data.ms
DFTPP151

m/z	Abundance
40.30	215.0
44.20	252.0
58.10	204.0
73.00	175.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

m/z	Abundance
37.20	161.0
38.10	433.0
39.10	2108.0
40.10	503.0
43.20	177.0
43.90	332.0
49.20	250.0
50.10	7715.0
51.10	29680.0
52.10	1226.0
55.20	289.0
56.10	1198.0
57.10	2199.0
58.10	285.0
61.10	345.0
62.10	493.0
63.10	1388.0
64.00	334.0
65.20	445.0
69.10	29272.0
70.10	197.0
73.20	401.0
74.10	3059.0
75.10	5825.0
76.10	1745.0
77.10	28960.0
78.20	2102.0
79.10	2680.0
80.10	1701.0
81.10	2357.0
82.10	576.0
83.10	447.0
85.00	196.0
86.10	650.0
87.20	207.0
89.10	151.0
91.00	666.0
92.20	474.0
93.10	4283.0
94.10	326.0
96.20	203.0
98.10	2874.0
99.10	2196.0
100.10	218.0
101.00	1068.0
103.00	166.0
104.10	1007.0
105.10	669.0
106.10	460.0
107.10	11124.0
108.10	2200.0
109.10	495.0
110.10	16384.0
111.00	2222.0
112.00	356.0
115.10	179.0
116.20	1016.0
117.10	8698.0
118.10	468.0
122.00	627.0
123.10	1067.0
124.10	498.0
125.00	396.0
127.10	30144.0
128.10	2334.0
129.10	13712.0
130.00	1469.0
131.10	305.0
134.00	551.0
135.20	1156.0
136.10	742.0
137.10	864.0
138.40	163.0
139.00	155.0
140.20	198.0
141.10	2499.0
142.10	486.0
143.10	574.0
144.20	225.0
146.10	239.0
147.20	1042.0
148.10	2212.0
149.00	406.0
150.20	237.0
152.00	453.0
153.10	886.0
154.10	559.0
155.10	1314.0
156.20	1459.0
157.20	354.0
158.10	452.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

m/z	Abundance
38.00	411.0
39.10	1664.0
40.00	248.0
44.10	282.0
49.00	185.0
50.10	6389.0
51.10	25856.0
52.10	1155.0
56.00	876.0
57.10	1701.0
58.00	245.0
61.10	678.0
62.10	367.0
63.10	921.0
65.10	549.0
69.10	25304.0
70.00	247.0
73.20	329.0
74.20	2613.0
75.10	4402.0
76.10	1255.0
77.10	27552.0
78.10	1812.0
79.10	1798.0
80.10	1754.0
81.00	1946.0
82.10	354.0
83.20	547.0
83.90	157.0
85.10	379.0
86.00	431.0
87.00	152.0
88.00	164.0
91.10	695.0
92.10	579.0
93.00	3631.0
94.00	306.0
96.30	180.0
98.10	3819.0
99.10	1894.0
100.10	162.0
101.10	1206.0
103.10	440.0
104.10	742.0
105.00	706.0
106.00	177.0
107.10	10203.0
108.10	1646.0
109.00	354.0
110.10	15368.0
111.10	2147.0
112.00	623.0
115.00	188.0
116.10	656.0
117.10	8179.0
118.00	570.0
122.10	493.0
123.10	1298.0
124.20	541.0
125.00	422.0
127.10	30904.0
128.10	2299.0
129.10	13345.0
130.10	824.0
131.10	346.0
132.00	201.0
134.10	430.0
135.10	1176.0
136.00	389.0
137.00	462.0
140.10	253.0
141.10	1721.0
142.10	468.0
143.10	283.0
146.00	442.0
147.00	605.0
148.10	1765.0
149.00	360.0
150.90	314.0
151.90	179.0
153.00	488.0
154.00	715.0
155.10	1178.0
156.20	1634.0
157.00	306.0
158.10	304.0
159.00	188.0
160.00	797.0
161.10	864.0
162.00	469.0
164.90	620.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

166.10	504.0
167.10	4070.0
168.20	1061.0
169.00	218.0
171.10	276.0
171.90	381.0
172.90	466.0
174.10	729.0
175.20	1445.0
176.10	224.0
177.10	717.0
178.00	297.0
179.10	3413.0
180.10	2261.0
181.10	1111.0
184.10	188.0
185.10	1402.0
186.20	10508.0
187.10	3363.0
188.10	232.0
189.20	1137.0
190.00	222.0
191.10	239.0
192.20	682.0
193.20	950.0
194.20	317.0
196.10	1373.0
196.90	600.0
198.10	65200.0
199.10	3116.0
200.10	419.0
201.50	221.0
202.10	238.0
203.20	697.0
204.20	3645.0
205.20	5877.0
206.20	20792.0
207.20	3072.0
208.20	974.0
209.20	292.0
210.40	277.0
211.10	705.0
212.00	258.0
214.90	190.0
216.00	475.0
217.10	6286.0
218.10	712.0
220.30	225.0
221.20	2118.0
221.90	816.0
223.20	1300.0
224.20	12514.0
225.20	2497.0
226.30	540.0
227.10	6514.0
228.10	925.0
229.10	1260.0
230.10	188.0
231.10	410.0
233.20	163.0
234.20	424.0
235.30	554.0
236.10	160.0
237.00	348.0
238.30	192.0
239.10	485.0
240.00	232.0
241.10	297.0
242.10	843.0
243.20	868.0
244.20	9147.0
245.20	1072.0
246.20	1999.0
247.00	346.0
248.10	259.0
249.10	245.0
252.10	165.0
253.40	215.0
255.10	54136.0
256.10	7831.0
257.10	642.0
258.20	2671.0
259.10	534.0
265.10	1712.0
266.10	187.0
271.20	267.0
272.10	196.0
273.10	1541.0
274.20	3482.0
275.20	22544.0
276.10	3181.0
277.20	1992.0
278.00	390.0
278.30	408.0
281.10	156.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

283.20	269.0
284.40	171.0
285.20	406.0
291.00	222.0
293.00	418.0
294.10	160.0
295.00	233.0
296.20	8264.0
297.20	1159.0
301.30	164.0
302.00	165.0
303.20	942.0
304.00	170.0
309.20	166.0
314.00	300.0
315.10	734.0
316.30	460.0
321.30	229.0
323.20	1969.0
324.10	282.0
327.20	575.0
328.20	329.0
329.10	161.0
333.10	386.0
334.10	1671.0
335.10	465.0
341.10	421.0
346.00	334.0
346.20	339.0
352.20	653.0
353.00	617.0
354.20	991.0
355.00	214.0
365.10	3515.0
366.00	448.0
367.50	161.0
370.10	204.0
371.20	332.0
372.20	1346.0
373.10	399.0
383.00	379.0
402.10	667.0
403.20	765.0
404.10	160.0
421.10	1122.0
422.20	718.0
423.10	4071.0
424.20	718.0
441.20	11363.0
442.20	77976.0
443.20	16416.0
444.20	1586.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

159.00	360.0
160.10	603.0
161.10	958.0
162.10	489.0
164.20	162.0
165.10	989.0
166.10	360.0
167.10	3066.0
168.20	1693.0
169.10	352.0
170.20	221.0
171.10	155.0
172.00	507.0
173.00	306.0
174.10	943.0
175.10	1617.0
176.30	287.0
177.20	663.0
178.00	537.0
179.10	2936.0
180.10	2169.0
181.10	1199.0
184.20	395.0
185.20	1717.0
186.10	10253.0
187.10	3125.0
188.00	305.0
189.10	914.0
191.10	298.0
192.10	1026.0
193.20	1056.0
194.00	289.0
195.10	230.0
196.10	1529.0
198.10	66184.0
199.10	4155.0
200.00	206.0
201.60	195.0
203.20	641.0
204.10	3110.0
205.20	4868.0
206.10	21568.0
207.20	2949.0
208.10	1126.0
209.10	335.0
210.10	496.0
211.20	713.0
215.20	204.0
216.20	493.0
217.10	6923.0
218.20	872.0
221.20	2601.0
222.00	937.0
223.20	1530.0
224.20	11395.0
225.10	3408.0
226.20	640.0
227.10	5327.0
228.00	732.0
229.10	959.0
230.10	310.0
231.20	418.0
233.10	187.0
234.10	375.0
235.00	297.0
236.10	212.0
237.00	515.0
239.10	241.0
240.20	181.0
241.20	360.0
242.20	803.0
243.20	790.0
244.20	9495.0
245.20	1073.0
246.10	1740.0
247.20	327.0
249.10	332.0
251.00	157.0
253.10	354.0
255.10	50816.0
256.20	7112.0
257.00	438.0
258.10	3130.0
259.10	456.0
261.20	181.0
264.30	177.0
265.10	1595.0
273.20	1340.0
274.10	3487.0
275.10	20208.0
276.10	3482.0
277.10	2027.0
278.20	380.0
283.20	209.0
285.30	206.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151

290.20	151.0
293.10	593.0
294.10	291.0
294.80	272.0
296.10	7650.0
297.20	1255.0
302.00	313.0
303.20	697.0
304.10	312.0
314.10	444.0
315.10	1184.0
316.30	391.0
317.30	282.0
321.40	224.0
322.00	215.0
323.10	1931.0
324.00	228.0
325.20	192.0
327.10	549.0
328.20	310.0
333.30	403.0
334.20	1952.0
335.10	571.0
341.10	464.0
342.10	193.0
346.10	560.0
347.10	155.0
352.30	665.0
353.40	269.0
354.10	348.0
354.30	371.0
365.10	3023.0
366.20	326.0
372.20	1120.0
373.10	199.0
383.10	463.0
390.10	158.0
402.10	490.0
403.20	519.0
404.20	221.0
421.30	400.0
422.10	418.0
423.20	3323.0
424.20	893.0
441.20	11171.0
442.20	63408.0
443.20	13072.0
444.20	1738.0

Instrument :
BNA_G
ClientSampleId :
DFTPP151