

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG071123\  
 Data File : BG058165.D  
 Acq On : 11 Jul 2023 19:04  
 Operator : CG/JU  
 Sample : PB153960BL  
 Misc :  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SBLK960

Quant Time: Jul 11 22:57:52 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\SFAM-EPA-BG070123.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Mon Jul 10 16:52:40 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.975  | 152  | 48088    | 20.000 | ng/ul | 0.00     |
| 20) Naphthalene-d8            | 10.778 | 136  | 225239   | 20.000 | ng/ul | 0.00     |
| 38) Acenaphthene-d10          | 14.603 | 164  | 157916   | 20.000 | ng/ul | 0.00     |
| 64) Phenanthrene-d10          | 17.341 | 188  | 368953   | 20.000 | ng/ul | 0.00     |
| 79) Chrysene-d12              | 21.601 | 240  | 328117   | 20.000 | ng/ul | 0.00     |
| 88) Perylene-d12              | 24.791 | 264  | 384703   | 20.000 | ng/ul | 0.00     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 3) 1,4-Dioxane-d8             | 3.404  | 96   | 6549     | 4.769  | ng/uL | 0.00     |
| 4) Pyridine-d5                | 3.816  | 84   | 111404   | 26.721 | ng/ul | 0.00     |
| 7) Phenol-d5                  | 7.135  | 99   | 145695   | 29.667 | ng/ul | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.294  | 67   | 85660    | 28.944 | ng/ul | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.505  | 132  | 101568   | 29.987 | ng/ul | 0.00     |
| 15) 4-Methylphenol-d8         | 8.675  | 113  | 112285   | 30.301 | ng/ul | 0.00     |
| 21) Nitrobenzene-d5           | 9.133  | 128  | 54940    | 29.927 | ng/ul | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.856  | 143  | 62608    | 32.449 | ng/ul | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.396 | 165  | 109813   | 29.669 | ng/ul | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.907 | 131  | 158498   | 28.244 | ng/ul | 0.00     |
| 46) Dimethylphthalate-d6      | 14.004 | 166  | 356327   | 28.598 | ng/ul | 0.00     |
| 49) Acenaphthylene-d8         | 14.297 | 160  | 409112   | 30.089 | ng/ul | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.797 | 143  | 52421    | 23.813 | ng/ul | 0.00     |
| 60) Fluorene-d10              | 15.590 | 176  | 318242   | 30.075 | ng/ul | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.708 | 200  | 45110    | 22.516 | ng/ul | 0.00     |
| 73) Anthracene-d10            | 17.441 | 188  | 510055   | 29.792 | ng/ul | 0.00     |
| 81) Pyrene-d10                | 19.726 | 212  | 593830   | 28.736 | ng/ul | 0.00     |
| 92) Benzo(a)pyrene-d12        | 24.574 | 264  | 626348   | 31.957 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG071123\  
Data File : BG058165.D  
Acq On : 11 Jul 2023 19:04  
Operator : CG/JU  
Sample : PB153960BL  
Misc :  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
SBLK960

Quant Time: Jul 11 22:57:52 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\SFAM-EPA-BG070123.MA.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Mon Jul 10 16:52:40 2023  
Response via : Initial Calibration

